



Bid Notice Abstract

Request for Quotation (RFQ)

Reference Number 11234022

Procuring Entity VALENZUELA MEDICAL CENTER

Title SUPPLY AND DELIVERY OF 1 UNIT OF ELECTROCAUTERY MACHINE

Area of Delivery Metro Manila

Solicitation Number:	VMC-2024-164	Status	Active
Trade Agreement:	Implementing Rules and Regulations	Associated Components	1
Procurement Mode:	Negotiated Procurement - Small Value Procurement (Sec. 53.9)	Bid Supplements	0
Classification:	Goods	Document Request List	3
Category:	Medical and Dental Equipment	Date Published	11/09/2024
Approved Budget for the Contract:	PHP 150,000.00	Last Updated / Time	11/09/2024 00:00 AM
Delivery Period:	30 Day/s	Closing Date / Time	16/09/2024 10:00 AM
Client Agency:			
Contact Person:	KRISTINE MANUEL BAC Secretariat Padrigal St., Karuhatan, Valenzuela City Metro Manila Philippines 63-2-(02) 294-4625 kristine.manuel905@gmail.com		

Description				
PR# 24-08-821				
ITEM NO. QTY. UNIT ITEM DESCRIPTION UNIT COST TOTAL COST				
1	1	unit	ELECTROCAUTERY MACHINE	150,000.00 150,000.00
Bipolar electrocautery machine				
1 pc Autoclavable, reusable handpiece with up and down power control with 10 ft. power cord				
Depth: 4 inches (102mm)				
Depth: 4 inches (102mm)				
Width: 8.75 inches (222mm)				
Height: 7.5 inches (190mm)				
Weight: 6 pounds (2.7kg)				
Output Power: Low 0-20 watts (1/10 watt increments up to 10 watts)				
High: 0-35 watts				
Bipolar: 0-35 watts				
220 volts, 50-60 Hz				
50 pc Disposable blunt tips, sterile				
50 pc Disposable sharp tips, sterile				
4 pc Sterile sheaths				
Three (3) years warranty				
Certificate of availability of spare parts within ten (10) years				
2 set Manual (for End-User and EFMS Biomedical Unit)				
1 pc Automatic Voltage Regulator (AVR)				
Training for biomedical technicians: for set up and machine checking, and trouble shooting				
Quarterly Preventive Maintenance Service (PMS) during warranty period, supplier must have proper equipment for PMS and must submit service report				
Must submit calibration and preventive maintenance schedule				
Calibration of equipment during warranty period, supplier must have medical equipment analyzers for				

verification/calibration and must submit calibration report
Electrical Safety Test Certificate
Loaner unit within 72 hours of non-operation
Inclusion of green/eco-procurement (preferably box or container is made of recyclable materials)

Total 150,000.00

A copy of your legal requirements (Certificate of Registration from BIR, SEC/DTI, Business/Mayor’s Permit, Platinum Philgeps Certificate, Tax Clearance/ITR 2023, Brochure, License to Operate (LTO) and Notarized Hard Copy of Omnibus Sworn Statement) is also required to be submitted along with your quotation/proposal.
Other Information
MS. RUBY S. GURREA, RN, MAN
Chairperson, BAC II
Bid and Awards Committee
Valenzuela Medical Center
Padrigal St., Karuhatan, Valenzuela City

Created by

KRISTINE MANUEL

Date Created

10/09/2024

The PhilGEPS team is not responsible for any typographical errors or misinformation presented in the system. PhilGEPS only displays information provided for by its clients, and any queries regarding the postings should be directed to the contact person/s of the concerned party.