



# PhilGEPS

Philippine Government Electronic Procurement System

Central Portal for  
Philippine Government  
Procurement Opportunities

Bid Notice Abstract

Request for Quotation (RFQ)

Reference Number12093364

Procuring EntityVALENZUELA MEDICAL CENTER

TitlePROCUREMENT OF ORTHOPEDIC EQUIPMENT - 2 ITEMS

Area of DeliveryMetro Manila

<b>Solicitation Number:</b>	VMC-2025-177	<b>Status</b>	Active
<b>Trade Agreement:</b>	Implementing Rules and Regulations	<b>Associated Components</b>	2
<b>Procurement Mode:</b>	Negotiated Procurement - Small Value Procurement (Sec. 53.9)	<b>Bid Supplements</b>	0
<b>Classification:</b>	Goods	<b>Document Request List</b>	4
<b>Category:</b>	Medical and Dental Equipment	<b>Date Published</b>	03/06/2025
<b>Approved Budget for the Contract:</b>	PHP 453,333.33	<b>Last Updated / Time</b>	03/06/2025 00:00 AM
<b>Delivery Period:</b>	60 Day/s	<b>Closing Date / Time</b>	09/06/2025 10:00 AM
<b>Client Agency:</b>			
<b>Contact Person:</b>	KRISTINE MANUEL BAC Secretariat Padrigal St., Karuhatan, Valenzuela City Metro Manila Philippines 63-2-(02) 294-4625  kristine.manuel905@gmail.com		

<b>Description</b>
PR# 25-05-677 & 25-05-679
ITEM NO. QTY. UNIT ITEM DESCRIPTION UNIT COST TOTAL COST
1 1 unit OSCILLATING SAW 193,333.33 193,333.33
Proposed location: Operating Room
Intended Use: For Orthopedic Surgery
I. Equipment Specifications
Operating speed: 0 - 14,000 cpm
Blade oscillating arc: 4.5"
Battery Pack: 4 cell
Inclusions:
- Battery Shield
- Battery Housing
- Battery Charger (3-way)
- Battery Housing Tray
II. Technical Specifications:
-FDA Certificate of Medical Device Notification (CMDN)
-The equipment must conform to the IEC 60601 Standards and any of the following standards:
a. USFDA (Food and Drug Administration, US standards)
b. CE (European Conformity) Standards
c. ISO standards (9000, 9001, 9002)
d. UL (Underwriter's Laboratories)
e. The proposed unit shall also conform to all relevant International, National and local standards and requirements of medical device regulatory agencies. Supplier shall list in the various international standards met by the supplier.
- Manufacturers Certificate Brand must be in the local market for at least 5 years
- Certification that there is established Service Center in Metro Manila or Philippines
- Certificate of Comprehensive Warranty which states the following clause:
a. The period of warranty shall be deemed to be fully comprehensive
b. All inclusive warranty, labor, spare parts, accessories, service consumables, manufacturer's Preventive Maintenance
- Warranty: One (1) year
- Certificate of availability of spare parts within five (5) years
-Certificate of Uptime of 95% and Downtime of 5%
-Preventive Maintenance and Calibration Services schedule
- Preventive Maintenance Service (Semi-Annual) during warranty period
- Calibration report from The Manufacturer or Verification report from The supplier
- Loaner Unit within 72 hours of non-operation



III. Acceptance, Testing & Commissioning of Equipment

- Training of End-User (Operation) with Certificate of Training should be provided and should contain the following details:
  - a. Name of Trainee
  - b. Modality, Brand, Model of Equipment
  - c. Type of Training Conducted
  - d. Inclusive dates of the Training
  - e. Name of Trainer, Date and Venue
- The unit to be delivered must show proof that its manufacturing date is not later than CY 2023 onwards
- Electrical Safety Test
- Provide two (2) sets of colored Technical Manual, Hardcopy and Softcopy (Flashdrive) in English Manual
  - a. User's Operational Manual
  - b. Quality Maintenance Manual
  - c. Service and Technical Manual
- Hands-on Training for Biomedical Unit, Supplier's must perform an actual:
  - a. Operations
  - b. Disassembly and Assembly
  - c. Troubleshooting
  - d. Recommended maintenance as per manufacturer
- IV. Others
  - Preferably packaging is made of recyclable materials
  - Preferably with green eco-products specification (non-toxic materials, energy efficient, sustainable, etc.)

2 1 unit PNEUMATIC TOURNIQUET 260,000.00 260,000.00

Proposed location: Operating Room

Intended Use: For Orthopedic Surgery

I. Equipment Specifications:

Pressure: 3 kPA - 100 kPA

Time: 1min - 600mins

Cuff: 4pcs

Air Hose: 1pc

II. Technical Specifications:

- FDA Certificate of Medical Device Notification (CMDN)
- The equipment must conform to the IEC 60601 Standards and any of the following standards:
  - a. USFDA (Food and Drug Administration, US standards)
  - b. CE (European Conformity) Standards
  - c. ISO standards (9000, 9001, 9002)
  - d. UL (Underwriter's Laboratories)
  - e. The proposed unit shall also conform to all relevant International, National and local standards and requirements of medical device regulatory agencies. Supplier shall list in the various international standards met by the supplier.
- Manufacturers Certificate Brand must be in the local market for at least 5 years
- Certification that there is established Service Center in Metro Manila or Philippines
- Standard nominal Volatge and Frequency: 220V / 60Hz
- Certificate of Comprehensive Warranty which states the following clause:
  - a. The period of warranty shall be deemed to be fully comprehensive
  - b. All inclusive warranty, labor, spare parts, accessories, service consumables, manufacturer's Preventive Maintenance
- Warranty: One (1) year
- Certificate of availability of spare parts within five (5) years
- Certificate of Uptime of 95% and Downtime of 5%
- Submission of Preventive Maintenance and Calibration Services schedule
- Preventive Maintenance Service (Semi-Annual) during warranty period
- Calibration report from The Manufacturer or Verification report from The supplier
- Loaner Unit within 72 hours of non-operation

III. Acceptance, Testing & Commissioning of Equipment

- Training of End-User (Operation) with Certificate of Training should be provided and should contain the following details:
  - a. Name of Trainee
  - b. Modality, Brand, Model of Equipment
  - c. Type of Training Conducted
  - d. Inclusive dates of the Training
  - e. Name of Trainer, Date and Venue
- The unit to be delivered must show proof that its manufacturing date is not later than CY 2023 onwards
- Electrical Safety Test
- Provide two (2) sets of colored Technical Manual, Hardcopy and Softcopy (Flashdrive) in English Manual
  - a. User's Operational Manual
  - b. Quality Maintenance Manual
  - c. Service and Technical Manual
- Hands-on Training for Biomedical Unit, Supplier's must perform an actual:
  - a. Operations
  - b. Dissassembly and Assembly
  - c. Troubleshooting
  - d. Recommended maintenance as per manufacturer
- IV. Others
  - Preferably packaging is made of recyclable materials
  - Preferably with green eco-products specification (non-toxic materials, energy efficient, sustainable, etc.)

Total 453,333.33

NOTE: ITEMS WILL BE SUBJECT FOR EVALUATION (DEMO UNIT)

A copy of your legal requirements (Certificate of Registration from BIR, SEC/DTI, Business/Mayor's Permit, PLATINUM Philgeps Certificate, Tax Clearance, Brochure/Data Sheet, License To Operate form FDA and Notarized Hard copy of Omnibus Sworn Statement) are also required to be submitted along with your quotation/proposal.

**Other Information**

MS. RUBY S. GURREA, RN, MAN  
Chairperson, BAC II  
VALENZUELA MEDICAL CENTER  
Padrigal St., Karuhatan, Valenzuela City

<b>Created by</b>	KRISTINE MANUEL
<b>Date Created</b>	02/06/2025

The PhilGEPS team is not responsible for any typographical errors or misinformation presented in the system. PhilGEPS only displays information provided for by its clients, and any queries regarding the postings should be directed to the contact person/s of the concerned party.