



Republic of the Philippines
Department of Health
Metro Manila Center for Health Development
VALENZUELA MEDICAL CENTER



BID BULLETIN NO. 1
22 AUGUST 2025

SUPPLY AND DELIVERY OF VARIOUS MEDICAL EQUIPMENT – BATCH 6

PUBLIC BIDDING NO. VMC-2025-070

Issued pursuant to Sec. 51.5 of the IRR of RA 12009 to clarify and/or amend certain provisions in the Bidding Documents issued for this project, considering the issues raised and clarifications made by the prospective bidders during the **Pre-bid** Conference held on August 20, 2025, likewise, respond to bidder's written queries received within the prescriptive period for filing.

Posted in Bidding Documents	Amendment/Clarifications
1. FETAL MONITOR WITH BATTERY TWIN ANTEPARTUM	
PRINTER	PRINTER
High speed: Review and print catch-up at up to 20cm/minute	High speed: Review and print catch-up at up to 10-20cm/minute
2. PEDIATRIC CRIB	
Headrests: Stainless Steel	Headrests: Stainless Steel (Adjustable)
4" Heavy Duty Caster wheels, double ball bearing (2 fixed, 2 with locking break)	4-6" Heavy Duty Caster wheels, double ball bearing (2 fixed, 2-4 with locking break)
3. PATIENT TRANSPORT STRETCHER	
I. Equipment Specification:	I. Equipment Specification:
Bed Frame: Carbon Steel with electrostatic powder coat finish and oxygen cylinder holder	Bed Frame: Carbon Steel with electrostatic powder coat finish and oxygen cylinder holder Size: W-80cm - 90cm (including bed rails) x L-210cm
IV pole with two (2) pcs sockets on corners with height adjustable IV pole with two (2) to four (4) hooks	Steel IV pole with two (2) pcs sockets on corners with height adjustable IV pole with two (2) to four (4) plastic or steel hooks
4. PATIENT TRANSPORT STRETCHER	
I. Equipment Specification:	I. Equipment Specification:
Bed Frame: Carbon Steel with electrostatic powder coat finish and oxygen cylinder holder	Bed Frame: Carbon Steel with electrostatic powder coat finish and oxygen cylinder holder Size: W-80cm - 90cm (including bed rails) x L-210cm
IV pole with two (2) pcs sockets on corners with height adjustable IV pole with two (2) to four (4) hooks	Steel IV pole with two (2) pcs sockets on corners with height adjustable IV pole with two (2) to four (4) plastic or steel hooks

***ITEMS** will be subject for product demonstration as part of Post Qualification Evaluation.

"PHIC Accredited Healthcare Provider"

"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <http://vmc.doh.gov.ph/>


BID OPENING: SEPTEMBER 2, 2025 at 10:00 A.M.

All other provisions of the bidding documents not affected by revisions included in this Bid Bulletin shall remain in effect. Amendments/inclusions/clarifications made herein shall be considered an integral part of the Bidding Documents.

Prepared by:


MS. KRISTINE JOY MANUEL
BAC Secretariat

Noted by:


MS. RUBY S. GURREA, RN, MAN
Chairperson, BAC II

"PHIC Accredited Healthcare Provider"
"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <https://vmc.doh.gov.ph/>