



Republic of the Philippines
Department of Health
Metro Manila Center for Health Development
VALENZUELA MEDICAL CENTER



BID BULLETIN NO. 1
29 OCTOBER 2025

**SUPPLY AND DELIVERY OF THREE (3) UNITS FETAL MONITOR WITH
BATTERY TWIN ANTEPARTUM**

PUBLIC BIDDING NO. VMC-2025-076

Issued pursuant to Sec. 51.5 of the IRR of RA 12009 to clarify and/or amend certain provisions in the Bidding Documents issued for this project, considering the issues raised and clarifications made by the prospective bidders during the **Pre-bid** Conference held on October 28, 2025.

Posted in Bidding Documents	Amendment/Clarifications
<u>FETAL MONITOR WITH BATTERY TWIN ANTEPARTUM</u>	
PRINTER	PRINTER
High speed: Review and print catch-up at up to 20cm/minute	High speed: Review and print catch-up at up to 10cm-20cm/minute
Warranty: One (1) year warranty on parts and service under normal use and conition except on expendable parts.	REMOVED
II. Electrical Specification:	II. Electrical Specification:
Equipment shall compatible with hospital power supply of 220 to 240VAC, single phase, 60Hz with grounding	Equipment shall compatible with hospital power supply of 220 to 240VAC, single phase, 60Hz with Hospital Grade Power Cord with grounding
III. Technical Specification:	III. Technical Specification:
NONE	Certificates for submission during Post Qualification
a. General Requirements	a. General Requirements
FDA Certificate of Medical Device Notification (CMDN)	With valid FDA Certificate of Medical Device Notification (CMDN)/Certificate of Medical Device Registration (CMDR) or pending renewal at least one (1) month before expiration
'The equipment or devices must conform to theIEC 60601 Standards and any of the following Standards:	'The equipment or devices must conform to any of the following Standards:
'a. USFDA (Food and Drug administration,US Standards b. CE(European conformity) Standards c. ISO Standards (9000, 9001, 9002) d. UL (Underwriters Laboratories) Standards e. Theproposed unit shall also conform to all relevant International, National and local standards and requirements of medical device regulatory agencies. Supplier shall list in the	'a. USFDA (Food and Drug administration,US Standards b. CE(European conformity) Standards c. ISO Standards (9000, 9001, 9002) d. UL (Underwriters Laboratories) Standards e. Theproposed unit shall also conform to all relevant International, National and local standards and requirements of medical device regulatory agencies. Supplier shall list in the

“PHIC Accredited Healthcare Provider”

“Valenzuela Medical Center...Where your health matters most”



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Web site: <https://vmc.doh.gov.ph/>

various international standards met by the supplier	various international standards met by the supplier
NONE	Certificates for submission during delivery:
One (1) years comprehensive warranty on parts and services for equipment	Two (2) years comprehensive warranty on parts and services for equipment
Preventive Maintenance and Calibration Services Schedule	Submit Preventive Maintenance and Calibration Services Schedule
Semi-Annual Preventive Maintenance and Annual Calibration Service during warranty period	Conduct Semi-Annual Preventive Maintenance and Annual Calibration Service during warranty period
Any corrective action requiring replacement of part/component(s) shall be conducted and completed within at most 3 days, during the warranty period.	REMOVED
Service support shall cover 24 hours/day, 7 days/week during the warranty period. Remote service should be provided within 12 hours and on-site service within 24 hours must be provided during the warranty period.	REMOVED
IV. Installation, Acceptance, Testing and Commissioning	IV. Acceptance, Testing and Commissioning
Installation, Acceptance, Testing & Commissioning of Equipment	Acceptance, Testing & Commissioning of Equipment
Certification that the bidder shall be responsible for notification, transportation, delivery, installation and commissioning at no cost to the government.	REMOVED
V. Other Terms & Condition:	
Bidder must undertake a site inspection. A Site inspection certificate shall be issued as proof of compliance. This certificate shall be submitted with the bid.	REMOVED
Preferably packaging is made of recyclable materials	REMOVED
Preferably with green eco-products specification (non-toxic materials, energy efficient, sustainability, etc.)	Preferably with green eco-products specification (non-toxic materials, energy efficient, sustainability, made of recyclable materials, etc.)
Certification that the Bid offer is in Philippine Peso to include taxes and duties, transportation to site, delivery, installation, and testing expenses on site by the bidder	REMOVED


BID OPENING: NOVEMBER 13, 2025 at 10:00 A.M.

All other provisions of the bidding documents not affected by revisions included in this Bid Bulletin shall remain in effect. Amendments/inclusions/clarifications made herein shall be considered an integral part of the Bidding Documents.

Prepared by:


MS. KRISTINE JOY MANUEL
BAC Secretariat

Noted by:


MS. RUBY S. GURREA, RN, MAN
Chairperson, BAC II