



BID BULLETIN NO. 2
17 January 2024

SUPPLY AND DELIVERY OF VARIOUS MEDICAL EQUIPMENT- BATCH 3

PUBLIC BIDDING NO. VMC-2023-041

Issued pursuant to Sec. 22.5 of the IRR of RA 9184 to clarify and/or amend certain provisions in the Bidding Documents issued for this project, considering the issues raised and clarifications made by the prospective bidders during the Pre-bid Conference held on January 11, 2024, likewise, respond to bidder's written queries received within the prescriptive period for filing.

Posted in Bidding Documents	Amendment/Clarifications
4. ELECTROSURGICAL UNIT	
Schedule of Requirements (Page 31)	Schedule of Requirements (Page 31)
- Trained personnel to use the equipment	REMOVED
Technical Specifications (Page 57)	Technical Specifications (Page 57)
None	- With protection for voltage surge
None	- Wattage/Ampere rating of the equipment
None	- With uptime and downtime certificate (95% to 98%)
OTHER REQUIREMENTS: (Schedule of Requirements and Technical Specifications)	OTHER REQUIREMENTS: (Schedule of Requirements and Technical Specifications)
None	- with Certificate of availability of spare parts within five (5) years
None	- with Certification of Calibration and Quarterly Preventive Maintenance
None	- with three (3) years warranty on parts and services
None	- with training on operation, maintenance and repair for Medical Equipment Technician (MET) and end-user


Bid opening: January 23, 2024 @ 10am

All other provisions of the bidding documents not affected by revisions included in this Bid Bulletin shall remain in effect. Amendments/inclusions/clarifications made herein shall be considered an integral part of the Bidding Documents.

Prepared by:


LIGAYA E. UBALDE
BAC Secretariat

Noted by:


RUBY S. GURREA, RN, MAN
Chairperson, BAC

"PHIC Accredited Healthcare Provider"

"Valenzuela Medical Center...Where your health matters most"