



Republic of the Philippines
Department of Health
Metro Manila Center for Health Development
VALENZUELA MEDICAL CENTER



BIDS AND AWARDS COMMITTEE

September 19, 2025

NOTICE OF AWARD

Supply and Delivery of one (1) unit Intradermal Injector
BAC RESO NO. 2025-306 dated September 12, 2025

MS. MARIANNE LORRAINE F. NAZARENO
Authorized Representative
SPECTRUMED INC.
2308 Chino Roces Ave., Extension,
Makati City
Contact No.: 0917-8655933
Email add.: lorraine@spectrumed.com.ph

Dear **Ms. Nazareno**:

Please be notified that the result of **Direct Contracting** for **SUPPLY AND DELIVERY OF ONE (1) UNIT INTRADERMAL INJECTOR** under P.R. No. 25-08-1193, is hereby awarded to you as the lowest quotation with a contract amount equivalent to **SIX HUNDRED EIGHTY-NINE THOUSAND FOUR HUNDRED SIXTEEN PESOS ONLY (₱ 689,416.00)**.

ITEM NO.	QTY	UNIT	ITEM / DESCRIPTION	UNIT COST	TOTAL AMOUNT
1	1	unit	INTRADERMAL INJECTOR	689,416.00	689,416.00
			OFFER: U225 INTRADERMAL INJECTOR / NEEDLE CONCEPT		
			Equipment Specification		
			Purpose: delivery of substances like platelet-rich plasma and hair biostimulators directly into the dermis		
			I. Equipment Specifications		
			Drive: Mechanical solenoid valve		
			Operating Modes: Manual or automatic (16 speeds)		
			Injection Depth Selection: Adjustable from 1-10 mm (±0.5)		
			Speed Range: Automatic Mode from 0.47 Hz to 7.1 Hz (28 inj. / min. to 426 injections/min)		

"PHIC Accredited Healthcare Provider"
"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <https://vmc.doh.gov.ph/>



Republic of the Philippines
Department of Health
Metro Manila Center for Health Development
VALENZUELA MEDICAL CENTER



BAGONG PILIPINAS

			Casing: Stainless steel structure and parts, polymethyl methacrylate covers, other parts are polyacetal		
			Power Supply Manufacturer XP Power, reference ACM06 US12 compliant with the IEC 60601-1:2005+A1 standard (supplied with the injector)		
			Input voltage: 100-240V AC / 50-60 Hz		
			Output voltage: 12 V DC – 650 mA		
			Power consumption: < 0.5 A - < 5 watts		
			Cable length: 1.80 Metres		
			Operating Temperature: 10° C to 40° C		
			Operating atmospheric pressure: 800 to 1600 hPa		
			Storage temperature: Room temperature		
			Relative air humidity: 5% to 95% (without condensation)		
			Classification: BF type applied part		
			Dimensions: Approximately 275 mm (L) x 25 mm (W) x 155 mm (H) (injector only)		
			Weight: About 650 g (injector only)		
			Injection Pressure: Adjustable from 0.5 bar to 5 bar (± 0.5)		
			Consumables:		
			One (1) box of 27G mesotherapy needles (60 pcs/box) – MESBIO Kits		
			One (1) box of 30G mesotherapy needles (60 pcs/box) – MESBIO Kits		
			One (1) box of 31G mesotherapy needles (60 pcs/box) – MESBIO Kits		
			II. Electrical Specification:		
			Equipment shall compatible with hospital power supply of 220 to 240VAC, single phase, 60Hz with grounding		
			With dedicated compatible automatic voltage regulator (AVR)		
			III. Technical Specification:		
			Warranty:		
			One (1) year comprehensive warranty on parts and services for equipment		

"PHIC Accredited Healthcare Provider"

"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karubatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <https://vmc.doh.gov.ph/>



Republic of the Philippines
Department of Health
Metro Manila Center for Health Development
VALENZUELA MEDICAL CENTER



BAGONG PILIPINAS

			One (1) year warranty on parts and services for AVR		
			Semi-Annual Functionality Tests and Annual Verification during warranty period		
			Certification that the bidder will provide a Service Unit that the end-user can use in case the equipment or any system component will be pulled-out for repair or maintenance within the warranty period.		
			IV. Acceptance, Testing and Commissioning		
			Acceptance, Testing & Commissioning of Equipment		
			Training of End-User (Operation) and Biomedical Unit (Operation & Troubleshooting) with Certificate of Training should be provided and should contain the following details:		
			a. Name of Trainee b. Modality, Brand, Model of Equipment c. Type of Training Conducted d. Inclusive Dates of the Training e. Name of Trainer, Date and Venue.		
			The unit to be delivered must show proof that its manufacturing date is not later than CY 2024 onwards		
			Conduct Verification Test during Delivery		
			Conduct Electrical Safety Test		
			Provide two (2) sets of colored Technical Manual, Hardcopy and Softcopy (Flashdrive) in English Manual for all equipment including peripherals, etc.		
			a. User's Operational Manual b. Quality and Maintenance Manual c. Service and Technical Manual		
			V. Other Terms & Condition:		
			Preferably with green eco-products specification (made of recyclable materials, non-toxic materials, energy efficient, sustainable, etc.)		
				TOTAL:	689,416.00

"PHIC Accredited Healthcare Provider"
"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <https://vmc.doh.gov.ph/>




Republic of the Philippines
Department of Health
Metro Manila Center for Health Development
VALENZUELA MEDICAL CENTER



The Valenzuela Medical Center reserves the right to withdraw / terminate this Notice of Award in view of non-delivery on the time specified, delivery of items of poor quality / substandard and any other violation of the terms and conditions provided in the Notice of Award.

Approved by:


EDILBERTO V. CAVANEYRO, MD, MHA, FPCS, FPSGS
OIC - Medical Center Chief II


SEP 25 2025

Conformed:

Authorized Representative
SPECTRUMED INC.
Company

Date: _____

mailed
9/25/25

VALENZUELA MEDICAL CENTER	
BAC RECEIVED	
Sign	
Date	9/25/25

9:20pm

"PHIC Accredited Healthcare Provider"
"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karubatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <https://vmc.doh.gov.ph/>