



BID BULLETIN NO. 1
11 January 2024

SUPPLY AND DELIVERY OF VARIOUS MEDICAL EQUIPMENT- BATCH 3

PUBLIC BIDDING NO. VMC-2023-041

Issued pursuant to Sec. 22.5 of the IRR of RA 9184 to clarify and/or amend certain provisions in the Bidding Documents issued for this project, considering the issues raised and clarifications made by the prospective bidders during the Pre-bid Conference held on January 11, 2024, likewise, respond to bidder's written queries received within the prescriptive period for filing.

Posted in ITB/Bidding Documents	Amendment/Clarifications
Total Amount to be bid: P 33,409,192.34	Total Amount to be bid: P 32,909,192.34
1. 3-CHANNEL ECG MACHINE	
- 5-inch diagonal, 16:9 Aspect Ratio, TFT LCD with LED Graphics Backlit	- 5-inch diagonal, 15-16:9 Aspect Ratio, TFT LCD with LED Graphics Backlit
- Weight: 1.1kg (without AC adapter, battery and recording paper)	- Weight: 1.1-1.1kgs (without AC adapter, battery and recording paper)
- Up to 300 seconds of rhythm analysis for easier in arrhythmia locating	- Up to 300 seconds of rhythm analysis for easier in arrhythmia locating or its equivalent
- Massive local storage capacity up to 1500 files	- Massive local storage capacity up to 1500 files or with copying/saving data thru flash drive
2. INCUBATOR (FOR MICROBIOLOGY)	
*Temperature Stability:(+/-) 0.1°C @ 37°C	*Temperature Stability:(+/-) 0.05-0.1°C @ 37°C
*Temperature Uniformity:(+/-) 0.5°C @ 37°C	*Temperature Uniformity:(+/-) 0.5-1°C @ 37°C
*Touch Screen Controller Features: Large, easy-to-read display screen • Data export via USB • Stores up to 5 years of log data • Password locking capabilities • Data logged once per minute • Export in .CSV file • Smart alarm: accounts for door openings and temperature set point changes before alarm • Temperature deviation alarm * with UPS/AVR	*Touch Screen Controller Features: Large, easy-to-read display screen • Data export via USB • Stores up to 5 years of log data • Password locking capabilities • Data logged once per minute • Export in .CSV file • Smart alarm: accounts for door openings and temperature set point changes before alarm • Temperature deviation alarm * with UPS/AVR or equivalent functions
4. ELECTROSURGICAL UNIT	
-1 pc reusable grounding pad	-1 pc reusable grounding pad/plate
-With at least 15.7 cm LCD touchscreen	-With at least 15.7 cm LCD touchscreen or LED numerical display
-must have the followubg settings: cut, coag, bipolar and shared coag	-must have the following settings: cut, coag, bipolar and shared coag or its equivalent
-Monopolar cutting modes: pure cut, blended cut	-Monopolar cutting modes: pure cut, blended cut or its equivalent
5. PORTABLE TRANSPORT MECHANICAL VENTILATOR	
>portable ventilator can be used in hospital, aeromedical and ground transport	>portable ventilator can be used in hospital, aeromedical (optional) and ground transport
>Able to ventilate adult and pediatric patients (at least 5kg of body weight)	>Able to ventilate adult and pediatric patients (at least 0.5kgs of body weight)
> MRI conditional/compatibility	DELETED

"PHIC Accredited Healthcare Provider"

"Valenzuela Medical Center...Where your health matters most"

• Padrigal St., Karuhatan, Valenzuela City 1441 • Trunk Line: 294-67-11 loc. 134; • Fax: 291-42-59


• URL: <http://vmc.doh.gov.ph> • Email: valgen_hosp@yahoo.com

6. INFANT INCUBATOR	
- Dual LCD screen, with 12.1-inch LCD color touch screen and 5.6-inch LCD color screen can be achieved by simple click	- Dual LCD screen, with 7-12.1-inch LCD color touch screen and 5.6-inch LCD color screen can be achieved by simple click
- Preferably Masimo Blood oxygen technology to effectively eliminate motion disturbance and improve the measurement performance of weak perfusion	- Preferably Masimo Blood oxygen technology and/or melcor technology to effectively eliminate motion disturbance and improve the measurement performance of weak perfusion
- Accuracy of skin temperature sensor: ± 0.2 degrees celcius	- Accuracy of skin temperature sensor: $\pm 0.2 - 0.3$ degrees Celsius
- Precision of tilt Angle display for crib equipped with electric adjustment	- Precision of tilt Angle display for crib equipped with electric and/or mechanical adjustment
- Mattress size: at least 635mm x 375mm x 30mm	- Mattress size: at least 630mm x 340mm x 30mm
7. STATIONARY CEILING MOUNTED DIGITAL X-RAY MACHINE	
STATIC WIFI FLAT PANEL DETECTOR	STATIC WIFI FLAT PANEL DETECTOR
Quantity: 4 pieces (17x17) and 4 pcs (14x17) with 2 batteries and 1 extra battery	Quantity: 2 pieces (17x17) and 2 pcs (14x17) with 2 batteries and 1 extra battery
Pixel Matrix: 2350x3072	Pixel Matrix: Minimum 2350x3072
Pixel Pitch 125 – 150 μm	Pixel Pitch 100 – 150 μm
Image Acquisition Time: 2s	Image Acquisition Time: maximum 2s
Image Processing Time: 3s	Image Processing Time: maximum 3s
Recommended cycle time:10s	Recommended cycle time: maximum 10s
Battery: Lithium ion, build in, rechargeable,3200mAh 2 pieces	Battery: Lithium ion, build in, rechargeable, at least 3200mAh 2 pieces
Housing Material: Carbon, Alloy	Housing Material: Carbon, Alloy or its equivalent
	Other Concerns:
	➤ Document Request List (DRL)
	➤ Notarized Certificate of Stocks Availability


Bid opening: January 23, 2024 @ 10am

All other provisions of the bidding documents not affected by revisions included in this Bid Bulletin shall remain in effect. Amendments/inclusions/clarifications made herein shall be considered an integral part of the Bidding Documents.

Prepared by:


LIGAYA E. UBALDE
BAC Secretariat

Noted by:


RUBY S. GURREA, RN, MAN
Chairperson, BAC