



Republic of the Philippines
Department of Health
Metro Manila Center for Health Development
VALENZUELA MEDICAL CENTER



BAGONG PILIPINAS

BID BULLETIN NO. 2
18 JULY 2024

**SUPPLY AND DELIVERY OF VARIOUS MEDICAL EQUIPMENT FOR CY 2024 –
BATCH 1**


PUBLIC BIDDING NO. VMC-2024-037

Issued pursuant to Sec. 22.5 of the IRR of RA 9184 to clarify and/or amend certain provisions in the Bidding Documents issued for this project, considering the issues raised and clarifications made by the prospective bidders during the **Pre-bid** Conference held on July 12, 2024, likewise, respond to bidder's written queries received within the prescriptive period for filing.


Posted in Bid Bulletin No. 1	Amendment/Clarifications
<u>2. INFANT INCUBATOR</u>	
Warm up time: (<35 MINUTES) Silent Mode: 60min, <45dB, Normal Mode: 45min, <55dB, Fast Mode: 30min, <66dB	Warm up time: at least 35 MINUTES Silent Mode: 60min, <45dB, Normal Mode: 45min, <55dB, Fast Mode: 30min, <66dB
Functions: mechanical tilting at least 10 degrees, Lifting Height: 651mm (lowest position) – 851mm (highest position)	Functions: mechanical tilting at least 10 degrees, Lifting Height: at least 651mm (lowest position) – at least 851mm (highest position)
<u>3. OPERATING ROOM TABLE, C-ARM COMPATIBLE</u>	
• Table must be maneuverable, with a maximum total table weight of at most 280 kg.	• Table must be maneuverable, with a maximum total table weight of at most 315 kgs.

All other provisions of the bidding documents not affected by revisions included in this Bid Bulletin shall remain in effect. Amendments/inclusions/clarifications made herein shall be considered an integral part of the Bidding Documents.

Prepared by:


LIGAYA E. UBALDE
BAC Secretariat

Noted by:


RUBY S. GURREA, RN, MAN
Chairperson, BAC