



Republic of the Philippines
Department of Health
Metro Manila Center for Health Development
VALENZUELA MEDICAL CENTER



BAGONG PILIPINAS

BIDS AND AWARDS COMMITTEE

June 23, 2025

NOTICE OF AWARD

Supply and Delivery of Various Ophtha Equipment – Batch 4
Public Bidding No. VMC-2025-055 dated May 14, 2025

MR. CESAR E. ROXAS

Authorized Representative

MTC OPTO-MEDIC INC.

631 Ronquillo St., Quaipo,

Manila

Contact No.: 0942-0694640

Email Add.: mtc.jhara.cabrera@gmail.com

mtcoptomedicsales@gmail.com

Dear **Mr. Roxas**:

We are happy to notify your Bid dated May 14, 2025 for execution of the Valenzuela Medical Center's IAEB Ref. No. 11984392 for contract price equivalent to **THREE MILLION TWO HUNDRED TWENTY-TWO THOUSAND ONE HUNDRED PESOS ONLY (P 3,222,100.00)** for the Supply and Delivery of Various Ophtha Equipment – Batch 4 as corrected and defined in accordance with the Instruction to Bidders is hereby accepted.

ITEM NO.	QTY	UOM	ITEM DESCRIPTION	BID PRICE	TOTAL BID PRICE
2	1	unit	SLIT LAMP WITH IMAGE CAPTURE	1,700,000.00	1,700,000.00
			BRAND: HUVITZ / KOREA		
			MODEL: HS 5000		
			Location: Ophthalmology Clinic		
			Personnel to use the equipment: Medical Specialist		
			I. Equipment Specification:		
			Capability to capture Color and B/W images		
			Auto Exposure Function		
			Capability to capture rapid series of images with one shot		
			Real time live image will be seen on the monitor		
			Video Capability with LAN connection interface included		
			Slit Length (mm) 0.3~12 continuous		
			Slit Width 0.12~ continuous		

"PHIC Accredited Healthcare Provider"

"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karubatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <https://vmc.doh.gov.ph/>



Republic of the Philippines
Department of Health
Metro Manila Center for Health Development
VALENZUELA MEDICAL CENTER



			Slit Projection 1x		
			Aperture Diaphragms 12.5x		
			Filter: Cobalt Blue, Red Free, Grey, Yellow and Heat Absorption , UV Cut Filter, ND Filter		
			Slit Rotation 0° ~ 180° continuous		
			Angle Incidence 0°, 5°, 10°, 15°, 20°		
			Patient's Eye/Prism 80mm		
			Surface Working Distance 80mm		
			Type 5 Position Rotating Drum		
			Eyepieces 12.5x		
			Total Magnifications 5x-6x, 10x, 16x, 25x, 40x-50x		
			Vertical Movement 28mm		
			Longitudinal Movement 78mm		
			Instrument Voltage 12V DC		
			Halogen Bulb 12V 30W or LED Lamp (10V 4.4W)		
			Fixation Point Bulb 3.4V 20mA		
			Digital Camera		
			Image Sensor 1/2" Interline CMOS		
			Image Size Up to 3,840 x 2,748 pixels		
			Cell size 1.67 µm X 1.67 µm		
			Resolution Depth 8bit or 2bit Raw RGB, yuv 4 : 2 : 2 (Offer: 32-bit)		
			With USB 2.0		
			Power Consumption less than 2.5W (5V DC, from USB Cable)		
			Accessories:		
			Motorized Stand, Heavy Duty		
			Chairs (2 pieces)		
			With Applanation Tonometer, Hang Type		
			Stand Alone		
			II. Electrical Specification:		

"PHIC Accredited Healthcare Provider"
"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <https://vmc.doh.gov.ph/>



Republic of the Philippines
Department of Health
Metro Manila Center for Health Development
VALENZUELA MEDICAL CENTER



BAGONG PILIPINAS

			Equipment shall compatible with hospital power supply of 220 to 240VAC, single phase, 60Hz with grounding		
			With compatible Uninterruptible Power Supply (UPS) with voltage regulation function that can provide back-up power at least 30 minutes		
			III. Technical Specification:		
			a. General Requirements		
			Certificate for the Cost of Preventive Maintenance and Calibration Program for equipment after warranty period up to 7 years. Price validity of 7 years upon delivery of equipment		
			Warranty and Preventive Maintenance Services:		
			Five (5) years comprehensive warranty on parts and services for equipment		
			Three (3) years warranty on parts and services for UPS		
			Certificate of availability of spare parts within ten (10) years		
			Certification that the 95% uptime of the product is within the warranty period and that any accumulated downtime in excess of 5% shall be added to the warranty period.		
			Preventive Maintenance and Calibration Services Schedule		
			Semi-Annual Preventive Maintenance Service during warranty period		
			Annual Calibration Service during warranty period and should conducted by manufacturers qualified Service Engineers/Technicians during warranty period		
			Calibration Report from the Manufacturer or Verification Report from the Supplier		
			Certification that the bidder will provide a Service Unit that the end-user can use in case the equipment or any system component will be pulled-out for repair or maintenance within the warranty period.		
			Any corrective action requiring replacement of part/component(s) shall be conducted and completed within at most 3 days, during the warranty period.		

"PHIC Accredited Healthcare Provider"

"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <https://vmc.doh.gov.ph/>



Republic of the Philippines
Department of Health
Metro Manila Center for Health Development
VALENZUELA MEDICAL CENTER



BAGONG PILIPINAS

			Service support shall cover 24 hours/day, 7 days/week during the warranty period. Remote service should be provided within 12 hours and on-site service within 24 hours must be provided during the warranty period.		
			Pricelist for major spare parts, accessories and consumables that is valid for 5 years.		
			Submit complete parts, accessories and consumables list with identification number or codes		
			Free software upgrades and updates (security, revision updates) for all Clinical and Technical applications during the period of warranty.		
			IV. Installation, Acceptance, Testing and Commissioning		
			Installation, Acceptance, Testing & Commissioning of Equipment		
			Certification that the bidder shall be responsible for notification, transportation, delivery, installation and commissioning at no cost to the government.		
			Comprehensive on-site operations and applications training for at least three (3) days for the end-users, to be conducted by the supplier's clinical applications specialist on the installed unit after passing the performance testing. A follow-up/refresh session after at least 3 months shall be conducted. Training should cover operation and use of hardware and software (all applications) of all equipment and their accessories and peripherals.		
			Training of End-User (Operation) with Certificate of Training should be provided and should contain the following details:		
			a. Name of Trainee b. Modality, Brand, Model of Equipment c. Type of Training Conducted d. Inclusive Dates of the Training e. Name of Trainer, Date and Venue		
			The unit to be delivered must show proof that its manufacturing date is not later than CY 2023 onwards		
			Verification test during delivery		
			Electrical Safety Test		
			Provide two (2) sets of colored Technical Manual, Hardcopy and Softcopy (Flashdrive) in English Manual for all equipment including peripherals, UPS, etc.		

"PHIC Accredited Healthcare Provider"

"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karubatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <https://vmc.doh.gov.ph/>



Republic of the Philippines
 Department of Health
Metro Manila Center for Health Development
VALENZUELA MEDICAL CENTER



BAGONG PILIPINAS

			a. User's Operational Manual b. Quality and Maintenance Manual c. Service and Technical Manual		
			Comprehensive technical training for hospital biomedical engineers and unit head on maintenance and troubleshooting of all equipment and their accessories and peripherals		
			Hands-on Training for Biomedical Unit, suppliers must perform an actual:		
			a. Operations b. Disassembly and Assembly c. Troubleshooting d. Recommended Maintenance as per manufacturer		
			V. Other Terms & Condition:		
			Preferably packaging is made of recyclable materials		
			Preferably with green eco-products specification (non-toxic materials, energy efficient, sustainability, etc.)		
			Certification that the Bid offer is in Philippine Peso to include taxes and duties, transportation to site, delivery, installation, and testing expenses on site by the bidder		
5	1	unit	PHOROPTER HEAD WITH REFRACTIVE CHAIR	1,360,000.00	1,360,000.00
			BRAND: CHAROPS / CHINA		
			MODEL: VT – 5B		
			Location: Ophthalmology Clinic		
			Personnel to use the equipment: Medical Specialist		
			I. Equipment Specification:		
			Phoropter head		
			Sphere power range: +16.75D~ -19.00D		
			Step: 0.25D		
			Cylinder power range : OD ~ -6.00D		
			Step: 0.25D (When applied with auxiliary cylinder lens +0.12D, Step: 0.12D		
			Cylinder axis range : 0°~180° Step: 5°		
			Cross cylinder : ± 0.25D		

"PHIC Accredited Healthcare Provider"

"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
 Telephone Nos: 8294-6711 to 17
 Director's Office Direct Line: 8291-4259
 Email Address: valgen_hosp@yahoo.com
 Website: <https://vmc.doh.gov.ph/>



Republic of the Philippines
Department of Health
Metro Manila Center for Health Development
VALENZUELA MEDICAL CENTER



BAGONG PILIPINAS

			Rotary prism range : 0Δ ~20 Δ Step : 1 Δ		
			P.D. Adjustment: 50-80mm, Step:1mm		
			Convergence adjustment: 40cm (When P.D. 64mm)		
			Forehead rest adjustment: 16mm Vertex distance : 13.75mm (From corneal point to standardn lens surface)		
			Refracting Unit and Chair		
			Up/Down angle of swing arm: ±30°		
			Table size: 670-900mm x 405-450mm (Offer: 920mm x 420mm)		
			Table rotational angle: 90°		
			Chair height adjustment range: 410~750mm		
			Loading for chair: 160-200kgs		
			Stand Alone		
			II. Electrical Specification:		
			Equipment shall compatible with hospital power supply of 220 to 240VAC, single phase, 60Hz with grounding		
			With compatible Uninterruptible Power Supply (UPS) with voltage regulation function that can provide back-up power at least 30 minutes		
			III. Technical Specification:		
			a. General Requirements		
			Certificate for the Cost of Preventive Maintenance and Calibration Program for equipment after warranty period up to 7 years. Price validity of 7 years upon delivery of equipment		
			Warranty and Preventive Maintenance Services:		
			Five (5) years comprehensive warranty on parts and services for equipment		
			Three (3) years warranty on parts and services for UPS		
			Certificate of availability of spare parts within ten (10) years		
			Certification that the 95% uptime of the product is within the warranty period and that any accumulated downtime in excess of 5% shall be added to the warranty period.		

"PHIC Accredited Healthcare Provider"

"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <https://vmc.doh.gov.ph/>



Republic of the Philippines
Department of Health
Metro Manila Center for Health Development
VALENZUELA MEDICAL CENTER



BAGONG PILIPINAS

			Preventive Maintenance and Calibration Services Schedule		
			Semi-Annual Preventive Maintenance Service during warranty period		
			Annual Calibration Service during warranty period and should be conducted by manufacturers qualified Service Engineers/Technicians during warranty period		
			Calibration Report from the Manufacturer or Verification Report from the Supplier		
			Certification that the bidder will provide a Service Unit that the end-user can use in case the equipment or any system component will be pulled-out for repair or maintenance within the warranty period.		
			Any corrective action requiring replacement of part/component(s) shall be conducted and completed within at most 3 days, during the warranty period.		
			Service support shall cover 24 hours/day, 7 days/week during the warranty period. Remote service should be provided within 12 hours and on-site service within 24 hours must be provided during the warranty period.		
			Pricelist for major spare parts, accessories and consumables that is valid for 5 years.		
			Submit complete parts, accessories and consumables list with identification number or codes		
			Free software upgrades and updates (security, revision updates) for all Clinical and Technical applications during the period of warranty.		
			IV. Installation, Acceptance, Testing and Commissioning		
			Installation, Acceptance, Testing & Commissioning of Equipment		
			Certification that the bidder shall be responsible for notification, transportation, delivery, installation and commissioning at no cost to the government.		
			Comprehensive on-site operations and applications training for at least three (3) days for the end-users, to be conducted by the supplier's clinical applications specialist on the installed unit after passing the performance testing. A follow-up/refresher session after at least 3 months shall be conducted. Training should cover operation and use of hardware and software (all applications) of all equipment and their accessories and peripherals.		

"PHIC Accredited Healthcare Provider"

"Valenzuela Medical Center...Where your health matters most"



9001:2015



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <https://vmc.doh.gov.ph/>



Republic of the Philippines
Department of Health
Metro Manila Center for Health Development
VALENZUELA MEDICAL CENTER



BAGONG PILIPINAS

			Training of End-User (Operation) with Certificate of Training should be provided and should contain the following details:		
			a. Name of Trainee b. Modality, Brand, Model of Equipment c. Type of Training Conducted d. Inclusive Dates of the Training e. Name of Trainer, Date and Venue		
			The unit to be delivered must show proof that its manufacturing date is not later than CY 2023 onwards		
			Verification Test during Delivery		
			Electrical Safety Test		
			Provide two (2) sets of colored Technical Manual, Hardcopy and Softcopy (Flashdrive) in English Manual for all equipment including peripherals, UPS, etc.		
			a. User's Operational Manual b. Quality and Maintenance Manual c. Service and Technical Manual		
			Comprehensive technical training for hospital biomedical engineers and unit head on maintenance and troubleshooting of all equipment and their accessories and peripherals		
			Hands-on Training for Biomedical Unit, suppliers must perform an actual:		
			a. Operations b. Disassembly and Assembly c. Troubleshooting d. Recommended Maintenance as per manufacturer		
			V. Other Terms & Condition:		
			Preferably packaging is made of recyclable materials		
			Preferably with green eco-products specification (non-toxic materials, energy efficient, sustainability, etc.)		
			Certification that the Bid offer is in Philippine Peso to include taxes and duties, transportation to site, delivery, installation, and testing expenses on site by the bidder		
7	1	unit	GONIOSCOPY LENS	56,800.00	56,800.00
			BRAND: VOLK / U.S.A.		
			Location: Ophthalmology Clinic		

"PHIC Accredited Healthcare Provider"

"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <https://vmc.doh.gov.ph/>



Republic of the Philippines
Department of Health
Metro Manila Center for Health Development
VALENZUELA MEDICAL CENTER



BAGONG PILIPINAS

			Personnel to use the equipment: Medical Specialist, Resident Doctors		
			4 x 64° mirror angles		
			1.0x image magnification		
			1.0x laser spot size		
			15 mm (flange) / 8.4 mm(no flange) contact diameter		
			One (1) year warranty		
8	1	unit	28D LENS	31,300.00	31,300.00
			BRAND: VOLK / U.S.A.		
			Location: Ophthalmology Clinic		
			Personnel to use the Instrument: Medical Specialist, Resident Doctors		
			53° / 69° field of view		
			2.27x image magnification		
			0.44x laser spot magnification		
			33 mm working distance		
			Available in autoclave sterilizable design or single use design		
			One (1) year warranty		
9	1	unit	78D LENS	37,000.00	37,000.00
			BRAND: VOLK / U.S.A.		
			Location: Ophthalmology Clinic		
			Personnel to use the instrument: Medical Specialist, Resident Doctors		
			81° / 97° field of view		
			0.93x image magnification		
			1.08x laser spot magnification		
			8 mm working distance		
			Provides great balance of magnification and field of view		
			Optimized for use within range of motion of all slit lamps		
			One (1) year warranty		

"PHIC Accredited Healthcare Provider"

"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <https://vmc.doh.gov.ph/>



Republic of the Philippines
Department of Health
Metro Manila Center for Health Development
VALENZUELA MEDICAL CENTER



BAGONG PILIPINAS

10	1	unit	90D LENS	37,000.00	37,000.00
			BRAND: VOLK / U.S.A.		
			Location: Ophthalmology Clinic		
			Personnel to use the equipment: Medical Specialist, Resident Doctors		
			74° / 89 ° field of view		
			0.76x image magnification		
			1.32x laser spot magnification		
			7 mm working distance		
			Small diameter ring in excellent for dynamic funduscopy and easy manipulation within the orbit		
			One (1) year warranty		
			TOTAL AMOUNT		3,222,100.00

You are required to submit performance security in the form and amount stipulated in the Instruction to Bidders. Failure to provide the performance security shall constitute sufficient grounds for cancellation of the award and failure of the bid security.

The Valenzuela Medical Center reserves the right to withdraw / terminate this Notice of Award in view of non-delivery on the time specified, delivery of items of poor quality / substandard and any other violation of the terms and conditions provided in the Instruction to Bidders.

Form of Performance Security	Amount of Performance Security (Equal to Percentage of the Total Contract Price)
a) Cash or Cashier's/manager (payable to BUREAU OF TREASURY), check issued by a Universal or Commercial Bank	Goods and Consulting Services - Five percent (5%); Infrastructure Projects - Ten percent (10%)
b) Bank draft/guarantee or irrevocable letter of credit issued by a Universal or Commercial Bank: Provided, however, that it shall be confirmed or authenticated by a Universal or Commercial Bank, if issued by a foreign bank.	
c) Surety bond callable upon demand issued by a surety or insurance company duly certified by the Insurance Commission as authorized to issue such security.	Thirty percent (30%)

"PHIC Accredited Healthcare Provider"

"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <https://vmc.doh.gov.ph/>



Republic of the Philippines
Department of Health
Metro Manila Center for Health Development
VALENZUELA MEDICAL CENTER




BAGONG PILIPINAS


INSTRUCTIONS:

If you have no corrections to the contents of this NOTICE OF AWARD (NOA), please submit Performance Security. The original copy of the NOA should be signed and claimed at the BAC Office, Valenzuela Medical Center. Further, please submit a certificate stating your Bank Account details.

Approved by:


JUN 26 2025
EDILBERTO V. CAVANEYRO, MD, MHA, FPCS, FPSGS
OIC-Medical Center Chief II

Conformed:


Erwin Naboya
Authorized Representative
MTC OPTO-MEDIC INC.
Company

Date: 6/26/25

emailed
6/26/25

"PHIC Accredited Healthcare Provider"

"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <https://vmc.doh.gov.ph/>