



Republic of the Philippines
Department of Health
Metro Manila Center for Health Development
VALENZUELA MEDICAL CENTER



BID BULLETIN NO. 1
5 SEPTEMBER 2025

SUPPLY AND DELIVERY OF OPHTHA EQUIPMENT – REBID

PUBLIC BIDDING NO. VMC-2025-073

Issued pursuant to Sec. 51.5 of the IRR of RA 12009 to clarify and/or amend certain provisions in the Bidding Documents issued for this project, considering the issues raised and clarifications made by the prospective bidders during the Pre-bid Conference held on September 5, 2025.

Posted in Bidding Documents	Amendment/Clarifications
<u>2. AUTOMATED PERIMETRY</u>	
Stimulus size : 50 um to 500 um	or manufacturer's standard
Single stimulus 26, 54 and 86 point tests (incorporating both 24-2 and 10-2 test pattern points)	10-2 or its equivalent
Multiple stimulus 26, 54 and 86 point tests (incorporating both 24-2 and 10-2 test pattern points)	10-2 or its equivalent
Standard :Threshold central 10-2: / 30/24-2 (extendable in-test)	10-2 or its equivalent
Fast :Threshold central 10-2: / 30/24-2 (extendable in-test)	10-2 or its equivalent
Single stimulus ~1 min (26 points); ~3.5 minutes for fully extended 24-2 plus 10-2 test (86 points)	10-2 or its equivalent
Both eyes camera monitoring for binocular tests	binocular test or its equivalent
BID OPENING – SEPTEMBER 17, 2025 @10:00AM	BID OPENING – SEPTEMBER 18, 2025 @11:00AM

BID OPENING: SEPTEMBER 18, 2025 at 11:00 A.M.

All other provisions of the bidding documents not affected by revisions included in this Bid Bulletin shall remain in effect. Amendments/inclusions/clarifications made herein shall be considered an integral part of the Bidding Documents.

Prepared by:


MS. KRISTINE JOY MANUEL
BAC Secretariat

Noted by:


MS. RUBY S. GURREA, RN, MAN
Chairperson, BAC II

"PHIC Accredited Healthcare Provider"

"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <http://vmc.doh.gov.ph/>