



Republic of the Philippines
Department of Health
Metro Manila Center for Health Development
VALENZUELA MEDICAL CENTER



BID BULLETIN NO. 1
26 MAY 2025

SUPPLY AND DELIVERY OF VARIOUS MEDICAL EQUIPMENT – BATCH 5


PUBLIC BIDDING NO. VMC-2025-056

Issued pursuant to Sec. 51.5 of the IRR of RA 12009 to clarify and/or amend certain provisions in the Bidding Documents issued for this project, considering the issues raised and clarifications made by the prospective bidders during the **Pre-bid** Conference held on May 23, 2025, likewise, respond to bidder's written queries received within the prescriptive period for filing.

Posted in Bidding Documents	Amendment/Clarifications
<u>1. PATIENT TRANSPORT STRETCHER</u>	
FDA Certificate of Medical Device Notification (CMDN)	With valid FDA Certificate of Medical Device Notification (CMDN)/Certificate of Medical Device Registration (CMDR) or pending renewal
<u>2. AUTOCLAVE STERILIZER (SQUARE OR CYLINDER)</u>	
-Sterilization Temperature: 120-130°C	-Sterilization Temperature: 120-134°C
-Steam Flush Pressure Pulse (SFPP) Control System	-Steam Flush Pressure Pulse (SFPP) Control System or Controlled by single-chip microcomputer (circuit board)
With consumable accessories: Door Gasket - 2 pcs	With consumable accessories: Door Gasket - 2 pcs (upon delivery)
-Delivery: 60 calendar days upon receipt of the Notice to Proceed	-Delivery: 90 calendar days upon receipt of the Notice to Proceed
<u>3. MINOR INSTRUMENT SET (SURGICAL)</u>	
NONE	Kelly Curved 6-7" (2 pcs)
NONE	Kelly Straight 6-7" (2 pcs)
NONE	Allis Grasper 6-7" (2 pcs)
Needle Holder 8"	Needle Holder 6-7"
Iris Scissor Curved	Iris Scissor 4-5" Curved
Iris Scissor Straight	Iris Scissor 4-5" Straight
Pair of skin retractor	Pair of senn (1 blunt and 1 sharp) retractor 6-7"
Scalpel blade holder	Scalpel blade holder no. 3

"PHIC Accredited Healthcare Provider"

"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <https://vmc.doh.gov.ph/>

Tissue forcep 8"	Tissue forcep 6-7"
Thumb Forcep 8"	Thumb Forcep 6-7"
Adson forcep toothed	Adson tissue forcep 4-5"
Adson Forcep w/o teeth	Adson thumb Forcep 4-5"
Tray with cover	Stainless steel Tray with cover (able to fit all instruments inside) autoclavable and plasma compatible
NONE	With valid FDA Certificate of Medical Device Notification (CMDN)/Certificate of Medical Device Registration (CMDR) or pending renewal
Warranty: 2 years warranty on parts and services	Warranty: 2 years warranty
NONE	Delivery: 90 calendar days upon receipt of the Notice to Proceed
NONE	NOTE: Sizes may vary, but minor differences that do not compromise its intended function are acceptable.
<u>4. FETAL MONITOR WITH BATTERY TWIN ANTEPARTUM</u>	
Colour 8.4" touchscreen display	Colour 8.4"-13.3" touchscreen display
CARE Analysis -Dawes-Redman (100,000 normal records data base)	100,000 normal records data base
DISPLAY	DISPLAY
Technology: full colour TFT Liquid Crystal display	Technology: full colour TFT Liquid Crystal display or Multicolor LED Touchscreen
Size: 8.4" diagonal 4:3 aspect ration	Size: 8.4 - 13.3" diagonal 4:3 aspect ration
PRINTER	PRINTER
Print head: 128 mm thick film	Print head: Manufacturer's standard
PAPER	PAPER
Plain paper: Thermal paper, z-fold 45m length (75 hours @ 1cm/minute) - 50 ROLLS PER UNIT	Plain paper: Thermal paper, z-fold 45m length (75 hours @ 1cm/minute) - 50 ROLLS PER UNIT or Manufacturer's standard
Pre-printed paper: HP/Philips/Agilent Corometrics Spacelabs	REMOVED
"Rear panel: *IEC-320 C14-mains power * Fetal Event Marker socket: 1/4 inch (6.35mm) jack plug connection *Equipotential Earth Point: provides common earthing point for connected equipment *RS232: CRS interface via 9 way D-Type connector Auxiliary: for wireless telemetry system USB Ports: External keyboard, barcode reader, upgrader memory stick Ethernet Port: Future CRS"	"Rear panel: *IEC-320 C14-mains power * Fetal Event Marker socket: 1/4 inch (6.35mm) jack plug connection or Manufacturer's standard *Equipotential Earth Point: provides common earthing point for connected equipment *RS232: CRS interface via 9 way D-Type connector Auxiliary: for wireless telemetry system USB Ports: External keyboard, barcode reader, upgrade memory stick Ethernet Port: Future CRS"

"PHIC Accredited Healthcare Provider"

"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <http://vmc.doh.gov.ph/>

Battery *Capacity: 4400mAh *Type: Lithium Ion *Use: Up to 4 hours depending on operating mode *Charging : approx. 4 hours	Battery *Capacity: 4400-5100mAh *Type: Lithium Ion *Use: Up to 4-7.2 hours depending on operating mode *Charging : approx. 4-5.5 hours
Physical Height:23.4cm(9.2 in) Width: 32.0 com (12.6) Length: 23.0 cm (9.0) Weight: 6 kg (13.2 lbs)	REMOVED
ENVIRONMENTAL	ENVIRONMENTAL
Operating * Temperature range: +10 C to +40 C * Relative Humidity: 10% to 90% (non condensing) *Pressure: 860mb to 1060 mb	Operating * Temperature range: +10 C to +40 C * Relative Humidity: 10% to 95% (non condensing) *Pressure: 860mb to 1060 mb
NONE	ADDITIONAL: Trolley
One (1) years comprehensive warranty on parts and services for equipment	Two (2) years comprehensive warranty on parts and services for equipment
5. LAPAROSCOPIC INSTRUMENTS	
1 set- Needle Holder, Curved Right	
Modular needle holder, curved right with carbide insert 5mm WL 310 w/ axial handle	Modular needle holder, curved right with carbide insert 5mm WL 310-330mm w/ axial handle
1 set- Needle Holder, Curved Left	
1 set- Modular needle holder, curved left with carbide insert 5mm WL 310	Modular needle holder, curved left with carbide insert 5mm WL 310-330 w/ axial handle
1 set-Puncture Cannula	1 set-Puncture/Injection Cannula
1 set- Puncture cannula with suction control and 4 mm Luer connector WL315mm	Puncture/Injection cannula with suction control and 3-4 mm Luer connector WL315mm
1 set-Injection Cannula	
1 set- Injection cannula with 3 mm Luer connector, 1.8 (15 G) WL 345mm	Injection cannula with 3 mm Luer connector, 1.4 (17 gauge) - 1.8 (15 gauge) WL 340-350mm
3 set- Trocar Sleeves 5.5 mm	
"3 sets- Metal sleeve, standard, oblique distal tip, 60 mm capacity, WL, 5.5 mm. Blunt conical tip, conical tip and pyramidal tip"	"Metal sleeve, standard, oblique distal tip, 60 mm capacity, WL, 5.5 mm. 1 Blunt conical tip, 1 conical tip and 1 pyramidal tip per set "
2 set- Trocar Sleeves 10 mm with insufflation tap	
3 sets - Metal sleeve, standard, oblique, distal tip, WL 100mm, 10 mm capacity, blunt conical tip, cinical tip, pyramidal tip	Metal sleeve, standard, oblique, distal tip, WL 100mm, 10 mm capacity, 1 blunt conical tip, 1 conical tip, 1 pyramidal tip per set
2 sets- Scissors Straight, double action, slim, monopolar 0.5 mm	2 sets- Scissors Straight, double action, slim, monopolar 5 mm
1 set- Maryland-Dissector	
Grasping and dissecting forceps Maryland-Dissector, JL= 23 mm, WL 330 mm, Handle, rotatable	Grasping and dissecting forceps Maryland-Dissector, JL= 23 mm, WL 320-330 mm, Handle, rotatable

"PHIC Accredited Healthcare Provider"
"Valenzuela Medical Center...Where your health matters most"




Address: Padrigal St., Karuhatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <http://vmc.doh.gov.ph/>

1 set- Metzenbaum scissors	
Metzenbaum scissors, JL= 22 mm WL330 Handle, rotatable	Metzenbaum scissors, JL= 22 mm WL 320-330 Handle, rotatable
1 set- Grasping forceps, fenestrated	
Grasping forceps, fenestrated, JL= 23 mm WL 330 Handle, rotatable	Grasping forceps, fenestrated, JL= 23 mm WL 320-330 Handle, rotatable
ADDITIONAL REQUIRED DOCUMENTS (to be submitted during Post-Qualification)	
5. The Supplier's audited financial statements, showing, among others, the Supplier's total and current assets and liabilities, stamped "received" by the BIR with 2023 ITR or its duly accredited and authorized institutions, for the preceding calendar year which should not be earlier than two (2) years from the date of bid submission	5. The Supplier's audited financial statements, showing, among others, the Supplier's total and current assets and liabilities, stamped "received" by the BIR with 2024 ITR or its duly accredited and authorized institutions, for online submission, an email confirmation from BIR for the preceding calendar year which should not be earlier than two (2) years from the date of bid submission
8. Proof of evidence for Single Largest Completed Contract (SLCC) should be fifty percent (50%) of the ABC to be bid – Purchase Order or Notice of Award or Contract Agreement	8. Proof of evidence for Single Largest Completed Contract (SLCC) should be fifty percent (50%) of the ABC – Purchase Order or Notice of Award or Contract Agreement.
11. Certificate of Stocks Availability from Bidder (Notarized)	REMOVED
NONE	11. Other requirements stated in the Bidding Documents/Technical Specifications.



BID OPENING: JUNE 4, 2025 at 10:00 A.M.

All other provisions of the bidding documents not affected by revisions included in this Bid Bulletin shall remain in effect. Amendments/inclusions/clarifications made herein shall be considered an integral part of the Bidding Documents.

Prepared by:


MS. KRISTINE JOY MANUEL
BAC Secretariat

Noted by:


MS. RUBY S. GURREA, RN, MAN
Chairperson, BAC II 



Address: Padrigal St., Karubatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <https://vmc.doh.gov.ph/>

"PHIC Accredited Healthcare Provider"
"Valenzuela Medical Center...Where your health matters most"