



Republic of the Philippines  
Department of Health  
Metro Manila Center for Health Development  
**VALENZUELA MEDICAL CENTER**



**BIDS AND AWARDS COMMITTEE**

March 4, 2025

**NOTICE OF AWARD**

Supply and Delivery of Arterial Blood Gas Reagents including 1CC Heparinized Syringe with Free Use of Two (2) Units ABG Machine (Table Top) for CY 2025  
Public Bidding No. VMC-2025 – 021 dated December 17, 2024

**MS. CLAIR B. ERNESTO**

Authorized Representative

**HEALTHRUSH ENTERPRISES**

Unit 704 West Trade Center,  
132 West Avenue, Quezon City

Email Add: ernestoclaire@yahoo.com

Dear **Ms. Ernesto**:

We are happy to notify your Bid dated December 17, 2025 for execution of Valenzuela Medical Center's IAEB Ref. No. 11529599 for contract price equivalent to **Five Million Eight Hundred Ninety Thousand Five Hundred Pesos Only (P P5,890,500.00)** for Supply and Delivery of Arterial Blood Gas Reagents Including 1cc Heparinized Syringe with Free Use of Two (2) Units ABG Machine (Table Top) for CY 2025 as corrected and defined in accordance with the Instruction to Bidders is hereby accepted.

ITEM NO	QTY	UNIT	ITEM DESCRIPTION	UNIT AMOUNT	TOTAL AMOUNT
1	9,900	pc	Arterial Blood Gas Reagents including 1cc Heparinized Syringe	595.00	5,890,500.00
			With free use of 2 units ABG Machine (Table Top) <b>OFFERED:</b> Brand/Model: CORNLEY BG-800 Origin: CHINA <b>SPECIFICATIONS:</b> <ul style="list-style-type: none"><li>- Arterial Blood Gas Reagents should be Cartridge Type or its equivalent</li><li>- To provide sample cartridge to onsite demonstration</li><li>- All wasted ABG Analysis Cartridge due to malfunctioning must be</li></ul>		

*"PHIC Accredited Healthcare Provider"*

*"Valenzuela Medical Center...Where your health matters most"*



Address: Padrigal St., Karuhatan, Valenzuela City, 1441  
Telephone Nos: 8294-6711 to 17  
Director's Office Direct Line: 8291-4259  
Email Address: valgen\_hosp@yahoo.com  
Website: <https://vmc.doh.gov.ph/>

			<ul style="list-style-type: none"> <li>- replaced immediately and with proper documentation of parties</li> </ul> <p><b><u>SPECIFICATIONS OF ABG MACHINE (TABLE TOP)</u></b></p> <ul style="list-style-type: none"> <li>- Model or Manufacturing date at least 2019</li> <li>- Touch screen menus and easy to use</li> <li>- Easy, quick and convenient zero maintenance auto sampling</li> <li>- Parameters ;ph,pCO2,pO2,Be,HCO3,SaO2</li> <li>- Allows continuous loading of samples</li> <li>- Provision of two (2) Chiller for storage of consumables</li> <li>- All accessories and consumables necessary running the machine free of charge from the hospital</li> <li>- Power requirements: <ul style="list-style-type: none"> <li>A. Provision of two (2) units UPS (Uninterrupted Power Supply Unit) that can run at least 30 minutes to 1 hour <ul style="list-style-type: none"> <li>a. All power cords should be three (3) prong</li> </ul> </li> </ul> </li> <li>- Availability of operational and Service manual</li> <li>- Any defect in the functionality of ABG machine will be repaired or replaced within 24 hours. The said repair will be free of charge to the hospital</li> <li>- Auto calibration before every sample test</li> <li>- With rechargeable internal battery that can run at least 30 minutes - 1 hour</li> <li>- With web software ready, with data management system or its equivalent</li> <li>- Desktop and printer provided by winning bidder with toner or ink</li> </ul> <p>Preventive Maintenance:</p> <ul style="list-style-type: none"> <li>- Quarterly PM by trained personnel without cost to the hospital during the entire usage</li> <li>- Certificate of calibration valid for 1year from the date of installation up to time decommissioning</li> <li>- Training for the end user conducted by winning bidder</li> </ul>		
			<b>TOTAL...</b>	<b>P 5,890,500.00</b>	

"PHIC Accredited Healthcare Provider"  
"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441  
Telephone Nos: 8294-6711 to 17  
Director's Office Direct Line: 8291-4259  
Email Address: valgen\_hosp@yahoo.com  
Website: <https://vmc.doh.gov.ph/>

You are required to submit the performance security in the form and amount stipulated in the Instruction to Bidders. Failure to provide the performance security shall constitute sufficient grounds for cancellation of the award and failure of the bid security.


The Valenzuela Medical Center reserves the right to withdraw / terminate this Notice of Award in view of non-delivery on the time specified, delivery of items of poor quality / substandard and any other violation of the terms and conditions provided in the Instruction to Bidders.

Form of Performance Security	Amount of Performance Security (Equal to Percentage of the Total Contract Price)
a) Cash or Cashier's/manager (payable to <b>BUREAU OF TREASURY</b> ), check issued by a Universal or Commercial Bank	Goods and Consulting Services - Five percent (5%); Infrastructure Projects - Ten percent (10%)
b) Bank draft/guarantee or irrevocable letter of credit issued by a Universal or Commercial Bank: Provided, however, that it shall be confirmed or authenticated by a Universal or Commercial Bank, if issued by a foreign bank.	
c) Surety bond callable upon demand issued by a surety or insurance company duly certified by the Insurance Commission as authorized to issue such security.	Thirty percent (30%)


#### INSTRUCTIONS:

If you have no corrections to the contents of this NOTICE OF AWARD (NOA), please submit Performance Security. The original copy of the NOA should be signed and claimed at the BAC Office, Valenzuela Medical Center. Further, please submit a certificate stating your Bank Account details.

Approved by:

 MAR 05 2025  
**EDILBERTO V. CAVANEYRO, MD, MHA, FPCS, FPSGS**  
 OIC- Medical Center Chief II

Conformed:

  
**Clair B. Ernesto**  
~~Medical Sales Representative~~  
 Authorized Representative  
**HEALTHRUSH ENTERPRISES**  
 Company

Date:

3/6/25

*emailed  
3/6/25*

"PHIC Accredited Healthcare Provider"

"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441  
 Telephone Nos: 8294-6711 to 17  
 Director's Office Direct Line: 8291-4259  
 Email Address: valgen\_hosp@yahoo.com  
 Website: <https://vmc.doh.gov.ph/>