



**BID BULLETIN NO. 1
18 JULY 2024**

**SUPPLY AND DELIVERY OF VARIOUS MEDICAL EQUIPMENT FOR CY 2024 –
BATCH 1**

PUBLIC BIDDING NO. VMC-2024-037

Issued pursuant to Sec. 22.5 of the IRR of RA 9184 to clarify and/or amend certain provisions in the Bidding Documents issued for this project, considering the issues raised and clarifications made by the prospective bidders during the Pre-bid Conference held on July 12, 2024, likewise, respond to bidder's written queries received within the prescriptive period for filing.

Posted in Bidding Documents	Amendment/Clarifications
1. <u>HUMAN MILK PASTEURIZER</u>	
Manufacturer's Certificate Periodic quarterly preventive maintenance and calibration for (5) years to provide schedule, service report with checklist	Semi annual preventive maintenance and annual calibration for three (3) years to provide schedule, service report with checklist
Manufacturer's Certificate of field service engineers performing preventive and corrective maintenance and calibration	Training certificate from Manufacturer's of field service engineers performing preventive and corrective maintenance and calibration
95% uptime and 98% downtime	98% uptime and 2% downtime
NONE	Two (2) years warranty on parts and services
NONE	Delivery Schedule: 90CD
2. <u>INFANT INCUBATOR</u>	
Has Help function (Quick guide) which provides the simple information about the equipment such as Part of full set, Main control panel and main display, how to use and change mode and parameters	Quick Guide attached on the Machine
System Alarms: Motor failed, Heater system failed, Humidity heater failed, Stuck key, Humidity sensor failure, Air sensor failure, Fan power failure, Air circulation failure, Skin probe fail, remove skin 2 probe, Sensor module disconnected, Sensor module not in position	System Alarms: Motor failed, Heater system failed, Humidity heater failed, Stuck key, Humidity sensor failure, Air sensor failure, Fan power failure, Air circulation failure, Skin probe fail, remove skin 2 probe, Sensor module disconnected, Sensor module not in position (Manufacturer's Specification)
Temperature Alarms Low: air temperature, High air temperature, Low skin temperature, High skin temperature, High skin 1 temperature, High skin 2 temperature, High temperature cutout, Over skin temperature	Temperature Alarms Low: air temperature, High air temperature, Low skin temperature, High skin temperature, High skin 1 temperature, High skin 2 temperature, High temperature cutout, Over skin temperature (Manufacturer's Specification)
Humidity Alarms Low humidity, Water empty	Humidity Alarms Low humidity, Water empty (Manufacturer's Specification)
Oxygen Alarms: O2 module power failure, O2 module detached, O2 sensor fault, Low oxygen, High oxygen	Oxygen Alarms: O2 module power failure, O2 module detached, O2 sensor fault, Low oxygen, High oxygen (Manufacturer's Specification)
Weighing Scale Alarms	Weighing Scale Alarms (Manufacturer's Specification)

"PHIC Accredited Healthcare Provider"

"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <https://vmc.doh.gov.ph/>

Skin Temperature Range: Normal: 35- 37.5C, Override: 37.6 - 39C, Skin measurement range: 25 - 45C $\pm 0.3C$	Skin Temperature Range: Normal: 34-38C, Override: 37 - 38C, Skin measurement range: 25 - 45C $\pm 0.3C$
Warm up time: Silent Mode: 60min, <45dB, Normal Mode: 45min, <55dB, Fast Mode: 30min, <66dB	Warm up time: (<35 MINUTES) Silent Mode: 60min, <45dB, Normal Mode: 45min, <55dB, Fast Mode: 30min, <66dB
Functions: mechanical tilting at least 12 degrees, Lifting Height: 651mm (lowest position) - 851mm (highest position)	Functions: mechanical tilting at least 10 degrees, Lifting Height: 651mm (lowest position) - 851mm (highest position)
Overall dimension must not higher than 1,024W x 690D x 1,354H mm	Overall dimension must not higher than 1,040-1500L x 675-750 x 1,354H - 1650mm
Noise level <45dB	Noise level <45-47dB
Power input AC 220V, 60Hz	Power input AC 220-240V, 60Hz
Built-in Phototherapy:	Built-in Phototherapy: EXTERNAL/PORTABLE
LED Lifespan 100,000 hours operating time	AT LEAST 50,000 HOURS OPERATING TIME
Wavelength peak between 450 - 475nm, effective surface Area 40x20cm	Wavelength peak between 400 - 550nm, effective surface Area 40x20-30cm or Manufacturer's specification
Intensity (at 40cm) Low: 25 - 35 $\mu W/cm^2/nm$, High: 35 - 55 $\mu W/cm^2/nm$	Intensity (at 40cm) Low: 25 - 75 $\mu W/cm^2/nm$, High: 35 - 55 $\mu W/cm^2/nm$
Overall dimension at least 340(W) x 210(D) x 75(H)mm	Overall dimension at least L 590MM X W 400MM X H 1360MM-1750MM
Manufacturer's Certificate Periodic quarterly preventive maintenance and calibration for (5) years to provide schedule, service report with checklist	Three (3) years quarterly Preventive Maintenance and annual calibration during warranty period
Manufacturer's Certificate of field service engineers performing preventive and corrective maintenance and calibration	Training certificate from Manufacturer's of field service engineers performing preventive and corrective maintenance and calibration
95% uptime and 98% downtime	98% uptime and 2% downtime
NONE	Three (3) years warranty parts and services
NONE	Delivery Schedule: 60CD
3. OPERATING ROOM TABLE, C-ARM COMPATIBLE	
• Column and base covers of the table must be made of pure electro - polished stainless steel	• Column and base covers of the table must be made of pure electro - polished stainless steel or ABS plastic
• Attachments are hook and unhook, screw and gear systems	• Attachments are hook/clamp and unhook/unclamp, screw and gear systems
• With integrated control panel that indicates the status of the correct functioning of the OR table	• With integrated control panel or its equivalent that indicates the status of the correct functioning of the OR table
• Height variation (without cushions) Lowest must be capable of 68 cm or lower	• Height variation (without cushions) Lowest must be capable of 68 cm to 69cm or lower
• Back section inclination: must be capable to achieve at least +80° up /-500 down	• Back section inclination: must be capable to achieve at least +80° up /-40°-50° down
• Removable and Interchangeable Head and Leg Sections	• Removable and/or Interchangeable Head and Leg Sections
• Head Section can be adjusted to at least +90o up/-50o down	• Head Section can be adjusted to at least +45° -90° up/30°-90° down
• Leg Section can be adjusted to at least +30o up/-90o down, and 90o to the side	• Leg Section can be adjusted to at least +15°-30° up/-90° down, and 90° to the side
• Table must be manoeuvrable, with a maximum total table weight of at most 210 kg.	• Table must be maneuverable, with a maximum total table weight of at most 280 kg.
Stirrups with Lift Assist: allows easy adjustment of Abduction & Lithotomy in patients up to 500 lbs (227 kg) while monitoring the sterile field	Stirrups with Lift Assist: allows easy adjustment of Abduction & Lithotomy in patients up to 600 lbs (272 kg) while monitoring the sterile field
-self adjusting boot	-self adjusting/rotating boot
-500 lbs patients (227kg) weight capacity	500-600 lbs patients (272kg) weight capacity
-Lithotomy range: +84° to -33°	-Lithotomy range: +84°-85° to -33°

"PHIC Accredited Healthcare Provider"
"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <https://vmc.doh.gov.ph/>

Technical Measurements	Technical Measurements
o Net Weight of Table: at most 210 kg (Table only)	o Net Weight of Table: at most 280 kg (Table only)
Basic Accessories	Basic Accessories
• 1 Pc Hand Pendant	• 1 Pc Hand Pendant or hand control
• 1 Pc AVR	• 1 Pc AVR with trolley
Battery: One week runtime, with built-in battery system in case of Power failure	Battery: One week runtime/minimum of 30 full movement, with built-in battery system in case of Power failure
Warranty: Five (5) years on parts and services includes onsite and offsite support	Warranty: two (2) years on parts and two (2) years on services includes onsite and offsite support
NONE	provision with yellofin stirrup
NONE	98% uptime, 2% downtime
NONE	Training certificate from Manufacturer's of field service engineers performing preventive and corrective maintenance and calibration
NONE	Delivery Schedule: 60CD
4. <u>PORTABLE ULTRASOUND/ECHOCARDIOGRAPHY MACHINE</u>	
PARAMETERS:	PARAMETERS:
Width - at least 390 mm	Width - at least 375 mm
Weight (no peripherals): not more than 5 kg with battery	Weight (no peripherals): not more than 7 kgs with battery
Electrical Power	Electrical Power
Power consumption maximum of not more than 200VA with peripherals	Power consumption maximum of 500VA with peripherals
Console Design	Console Type or Laptop Design
Easily removable air filters	Easily removable air filters OPTIONAL
User Interface Operator keyboard	User Interface Operator keyboard
Full alphanumeric keypad covered with washable protection film	Full alphanumeric keypad (QWERTY) covered with washable protection film
at least 8 TGC pods	at least 6 TGC pods
Monitor	Monitor
at least 15.6" high-resolution LCD Monitor: 1920 x 1080 resolution	at least 15.6" high-resolution LCD Monitor: 1280 x 800 resolution
Cart Dimension	Cart Dimension
Length - at least 555mm	Length - at least 525mm
System mounting with easy lock/unlock push button feature	System mounting feature
Power cord hooks and clips for neat arrangement of cords	Cable Manager
Probe holders, removable for cleaning and washing	Probe holders, removable/not, transfix/transducer with removable silicon cover for cleaning and washing
Probe cord management holder	Cable Manager
Gel Holder, removable for cleaning and washing	Gel Holder, removable/not for cleaning and washing
Front and rear handles	Front handle only
Cart tray storage up to 1kg load (Option)	Cart tray storage up to 4kgs load (Optional)
Trolley Case	OPTIONAL
3 protective compartments for the probes	
3 additional compartments for power adapters, cord, and manuals.	
Length - at least 495mm	

"PHIC Accredited Healthcare Provider"
"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <https://vmc.doh.gov.ph/>

Depth - at least 275mm	
Height - at least 460mm	
Weight - at least 4kg	
Operating Modes	Operating Modes
Coded Harmonic Imaging	Coded Harmonic Imaging OR ITS EQUIVALENT
Color Flow Mode (CFM)	Color Flow Mode (CFM) (optional)
PW Doppler with high PRF	PW Doppler with high PRF (optional)
B-Flow (B-Flow Color) (Option)	B-Flow (B-Flow Color) OR ITS EQUIVALENT (Optional)
Cine memory/ image memory	Cine memory/ image memory
384 MB of Cine Memory	384 MB of Cine Memory (optional)
Cine review speed: at least 11 steps	Cine review speed: at least 11 steps (optional)
System Scanning Parameters	System Scanning Parameters
Max frame rate: at least 1790 f/s (depends on probes and modes)	Max frame rate: at least 409 f/s (depends on probes and modes)
Transmission focus: 1 - 8 focal points selectable	Transmission focus: 1 - 8 focal points selectable (optional)
Continuous dynamic receive focus/aperture	Continuous dynamic receive focus/aperture (optional)
Multifrequency/wideband technology	Multifrequency/wideband technology (optional)
Frequency range: 1.7 to 18Mhz (depends on probe)	Frequency range: 1 to 18Mhz (depends on probe)
Systematic dynamic range: at least 269 dB	Systematic dynamic range: at least 10 db
System Standard Features	System Standard Features
Provides 3,5,7,9 frames of spatial compounding	Provides 3,5,7,9 frames of spatial compounding (on and off spatial compounding)
Coded Phase Inversion Harmonic Imaging	Coded Phase Inversion Harmonic Imaging 9 (OPTIONAL)
Image archive on integrated HDD	Image archive on integrated SSD
Raw Data Analysis	Raw Data Analysis OR ITS EQUIVALENT
Real-time automatic Doppler Calculations	Real-time automatic Doppler Calculations OR ITS EQUIVALENT
MPEGvue	MPEGvue or viewer, OR ITS EQUIVALENT
Abstracted from basic user manual, it lists out FAQs from customers and instructs customer how to solve problems by themselves timely.	Abstracted from basic user manual, it lists out FAQs from customers and instructs customer how to solve problems by themselves timely. (OPTIONAL)
Provides an automated exam script that moves you through an exam step by step.	Provides an automated exam script that moves you through an exam step by step. (OPTIONAL)
Idle mode	Idle mode or standby mode (OPTIONAL)
Transducer Types	Transducer Types
1. 8C-RS	1. 8C-RS OR MICROCONVEX
FOV: at least 131 degrees	FOV: at least 110 degrees or MANUFACTURER'S SPECS
Footprint: at least 22.0 x 12.0 mm	Footprint: at least 12.8 x 29 mm (OPTIONAL)
Frequency: at least 4.2 and 10.0 MHz	Frequency: at least 3.5 or MANUFACTURER'S SPECS
2. Phased Array Sector probe (Pedia)	2. Phased Array Sector probe (Pedia)
FOV: atleast 120 degrees	FOV: at least 90 degrees or MANUFACTURER'S SPECS
Footprint: 23.5 x 16.8 mm	Footprint: at least 30.1 x at least 24.1 mm (OPTIONAL)
Frequency: at least 2.5 and 7 MHz	Frequency: at least 1 MHz or MANUFACTURER'S SPECS
3. Phased Array Sector Probe (Neonatal)	3. Phased Array Sector Probe (Neonatal)

"PHIC Accredited Healthcare Provider"

"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
 Telephone Nos: 8294-6711 to 17
 Director's Office Direct Line: 8291-4259
 Email Address: valgen_hosp@yahoo.com
 Website: <https://vmc.doh.gov.ph/>

FOV: atleast 120 degrees	FOV: at least 90 degrees or MANUFACTURER'S SPECS
Footprint: 17.6 x 13.2 mm	Footprint: at least 15.2 x 14.1 mm (OPTIONAL)
Frequency: at least 4.1 and 12.0 MHz	Frequency: at least 3 MHz or MANUFACTURER'S SPECS
Peripheral/Add-Ons	Peripheral/Add-Ons
USB ECG kits (Option)	USB ECG kits or CLIP TYPE (Optional)
95% uptime and 98% downtime	98% uptime and 2% downtime
NONE	The brand of the unit to be delivered must have been available in the Philippines for at least 15 years.
NONE	WARRANTY 3 YEARS ON PARTS AND SERVICES
NONE	DELIVERY SCHED: 60 CD
NONE	Preventive Maintenance (QUARTERLY) and Calibration (annual) during warranty period
NONE	Training certificate from Manufacturer's of field service engineers performing preventive and corrective maintenance and calibration
5. <u>RADIANT WARMER</u>	
Product Dimensions:	Product Dimensions:
Device weight: not more than 20 kg	Device weight: not more than 200 kg.
95% uptime and 98% downtime	98% uptime and 2% downtime
NONE	The brand of the unit to be delivered must have been available in the Philippines for at least 15 years.
NONE	WARRANTY 3 YEARS ON PARTS AND SERVICES
NONE	Preventive Maintenance (SEMI ANNUAL) and Calibration (annual) during warranty period
NONE	Training certificate from Manufacturer's of field service engineers performing preventive and corrective maintenance and calibration
NONE	DELIVERY SCHED: 60 CD
6. <u>AUTOClave STERILIZER (SQUARE)</u>	
Specifications:	Specifications:
-Buzzer alarm in case of door opening during sterilization process	-Buzzer alarm or with screen indicator in case of door opening during sterilization process
-Automatic cut-off power if water is insufficient	-Automatic cut-off power if water is insufficient
95% uptime and 98% downtime	98% uptime and 2% downtime
NONE	The brand of the unit to be delivered must have been available in the Philippines for at least 15 years.
NONE	Training certificate from Manufacturer's of field service engineers performing preventive and corrective maintenance and calibration
NONE	DELIVERY SCHED: 60 CD
7. <u>VERTICAL LAMINAR FLOW</u>	
> External Dimensions (W x D x H): 1340 x 784 x 1270 mm	> External Dimensions (W x D x H): 1340 x 780 - 784 x 1270 mm
> Internal Work Area, Dimensions (W x D x H): 1270 x 739 x 689 mm	> Internal Work Area, Dimensions (W x D x H): 1270 x 695 - 739 x 689 mm
Loaner unit (to all equipment)	(REMOVE!)
95% uptime and 98% downtime	98% uptime and 2% downtime
NONE	Two (2) years warranty on parts and (3) YEARS ON services
NONE	Preventive Maintenance (SEMI ANNUAL) and Calibration (annual) during warranty period
NONE	The brand of the unit to be delivered must have been available in the Philippines for at least 15 years.

"PHIC Accredited Healthcare Provider"
"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <https://vmc.doh.gov.ph/>

NONE	Training certificate from Manufacturer's of field service engineers performing preventive and corrective maintenance and calibration
NONE	Delivery Schedule: 90CD
8. <u>LABORATORY FREEZER (MILKBANK FREEZER)</u>	
loaner unit (to all equipment)	(REMOVE!)
95% uptime and 98% downtime	98% uptime and 2% downtime
NONE	Two (2) years warranty on parts and three (3) YEARS ON services
NONE	Preventive Maintenance (SEMI ANNUAL) and Calibration (annual)
NONE	The brand of the unit to be delivered must have been available in the Philippines for at least 15 years.
NONE	Training certificate from Manufacturer's of field service engineers performing preventive and corrective maintenance and calibration
NONE	Delivery Schedule: 90CD
9. <u>LABORATORY REFRIGERATOR (MILKBANK REFRIGERATOR)</u>	
NONE	Two (2) years warranty on parts and three (3) YEARS ON services
NONE	Preventive Maintenance (SEMI ANNUAL) and Calibration (annual)
NONE	The brand of the unit to be delivered must have been available in the Philippines for at least 15 years.
NONE	Training certificate from Manufacturer's of field service engineers performing preventive and corrective maintenance and calibration
NONE	Delivery Schedule: 90CD
10. <u>TRACTION SET</u>	
With five (5) years warranty on parts and services	With three (3) years warranty on parts and services
NONE	The brand of the unit to be delivered must have been available in the Philippines for at least 15 years.
NONE	Training certificate from Manufacturer's of field service engineers performing preventive and corrective maintenance and calibration
11. <u>PATIENT MONITOR</u>	
At least 12-inch TFT color LCD, maximum 6 waveforms, wall mounted and cart	At least 12-inch TFT color LCD, maximum 8 waveforms, wall mounted and cart
Electrode offset potential tolerance $\geq +500$ mV	Electrode offset potential tolerance $\pm 400 - 800$ mV
Onscreen guide with helpful illustrations makes it easy to operate the monitor, also includes an explanation of abbreviations for parameters, arrhythmias and icons	Onscreen guide with helpful illustrations makes it easy to operate the monitor, also includes an explanation of abbreviations for parameters, arrhythmias and icons (Optional)
Up to 120 hours of recall data can be saved for arrhythmias	Up to 120 hours or 200 alarm events data can be saved for arrhythmias
Synchronization mark: Heart rate sync mark, pulse rate sync mark, respiratory sync mark	Synchronization mark: Heart rate sync mark, pulse rate sync mark, respiratory sync mark (Optional)
ECG recovery time after defibrillation: 10 seconds	ECG recovery time after defibrillation: 5 - 10 seconds
Systolic range: Child: 40 - 280 mmHg, Neonate: 30 - 140 mmHg	Systolic range: Child: 25 - 240 mmHg, Neonate: 30 - 140 mmHg
Diastolic range: Child: 10 - 235 mmHg, Neonate: 10 - 110 mmHg	Diastolic range: Child: 10 - 220 mmHg, Neonate: 10 - 110 mmHg
Mean range: Child: 20 - 255 mmHg, Neonate: 15 - 125 mmHg	Mean range: Child: 15 - 225 mmHg, Neonate: 15 - 125 mmHg
Initial pressurization value: Child: 140 mmHg, Neonate: 100 mmHg	Initial pressurization value: Child: 100-140 mmHg, Neonate: 100 mmHg
Maximum pressurization value: Child: 300 mmHg, Neonate: 150 mmHg	Maximum pressurization value: Child: 240-300 mmHg, Neonate: 150 mmHg

"PHIC Accredited Healthcare Provider"
"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <https://vmc.doh.gov.ph/>

Safety - Maximum pressurization value cuff inflation limiter : Child: 300 to 330 mmHg; Neonate: 150 to 165 mmHg	Safety - Maximum pressurization value cuff inflation limiter : Child: 250 to 330 mmHg; Neonate: 150 to 165 mmHg
QRS detection (at × 1 sensitivity): Child and neonate: Width: 40 to 120 ms, Amplitude: 0.5 to 5 mV, rate: 30 to 250 beats/min	QRS detection (at × 1 sensitivity): Child and neonate: Width: 40 to 120 ms, Amplitude: 0.5 to 5 mV, rate: 30 to 250 beats/min (Optional)
Counting range: 0, 15 to 300 beats/min (±2 beats/min)	Counting range: 0, 15 to 300 beats/min (±2 beats/min) (Optional)
Power discontinuity: Deflate immediately after power down	Power discontinuity: Deflate immediately after power down (Optional)
SPO2 Sync tone modulation: Changes tone depending on SpO2 value	SPO2 Sync tone modulation: Changes tone depending on SpO2 value (Optional)
SPO2 Waveform sensitivity: ×1/8, ×1/4, ×1/2, ×1, ×2, ×4, ×8 or AUTO	SPO2 Waveform sensitivity: ×1/8, ×1/4, ×1/2, ×1, ×2, ×4, ×8 or 2,3,4,6,7,8
Up to 120 hours of ECG full disclosure, wave duration 60 s, up to 4 traces	Up to 48 -120 hours of ECG full disclosure, wave duration 60 s, up to 4 traces
At least 6hours battery operation	At least 4 hours battery operation
NONE	Two (2) years warranty on parts and two (2) years ON services
NONE	Preventive Maintenance (QUARTERLY) and Calibration (annual) during warranty period
NONE	The brand of the unit to be delivered must have been available in the Philippines for at least 15 years.
NONE	Training certificate from Manufacturer's of field service engineers performing preventive and corrective maintenance and calibration
NONE	Delivery Schedule: 90CD
12. T-PIECE RESUSCITATOR	
NONE	Two (2) years warranty on parts and two (2) YEARS ON services
NONE	Preventive Maintenance (QUARTERLY) and Calibration (annual) during warranty period
NONE	The brand of the unit to be delivered must have been available in the Philippines for at least 15 years.
NONE	Training certificate from Manufacturer's of field service engineers performing preventive and corrective maintenance and calibration
NONE	Delivery Schedule: 90CD
13. INFANT PHOTOTHERAPY	
NONE	Two (2) years warranty on parts and two (2) YEARS ON services
NONE	Preventive Maintenance (Semi annual) and Calibration (annual) during warranty period
NONE	The brand of the unit to be delivered must have been available in the Philippines for at least 15 years.
NONE	Training certificate from Manufacturer's of field service engineers performing preventive and corrective maintenance and calibration
NONE	Delivery Schedule: 90CD
14. BUBBLE CPAP WITH OXYGEN BLENDER	
Flow rate adjustable range: 0 L/min -18L/min and 0 L/min-3.5L/min	Flow rate adjustable range: 0 L/min -18L/min and 0.1 L/min-1L/min
Gas bleed valve: Yes	Gas bleed valve: Yes (MANUFACTURER'S SPECS)
Main part dimension: 123mm x 68mm x 71mm	Main part dimension: 123mm x 68mm x 71mm (MANUFACTURER'S SPECS)
Weight: 0.84kg	Weight: 0.84kg (MANUFACTURER'S SPECS)
Packing dimension: 315mm x 300mm x 265mm	Packing dimension: 315mm x 300mm x 265mm (MANUFACTURER'S SPECS)
Gross Weight: 3.2kg	Gross Weight: 3.2kg (MANUFACTURER'S SPECS)
Babi Plus Bubble CPAP Kit	Plus Bubble CPAP Kit

"PHIC Accredited Healthcare Provider"
"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <https://vmc.doh.gov.ph/>

Babi Plus Bubble PAP Valve	Plus Bubble PAP Valve OR SAFETY VALVE
Babi Plus Bubble PAP Valve Pole Clamp	Plus Bubble PAP Valve Pole Clamp (OPTIONAL)
NONE	THREE (3) years warranty on parts and two (2) YEARS ON services
NONE	Preventive Maintenance (SEMI ANNUAL) and Calibration (annual) during warranty period
NONE	The brand of the unit to be delivered must have been available in the Philippines for at least 15 years.
NONE	Training certificate from Manufacturer's of field service engineers performing preventive and corrective maintenance and calibration
NONE	Delivery Schedule: 90CD
15. MILKBANK LABELLING SYSTEM	
> MBLS Pasteurization freezing labels, 24 roll	> MBLS Pasteurization freezing labels, 2 ROLL
> MBLS Resin print ribbon, 12 roll	> MBLS Resin print ribbon, 1 ROLL
loaner unit during warranty (to all equipment) with response time of 72 hours	(REMOVE)
95% uptime and 98% downtime	(REMOVE)
NONE	TWO (2) years warranty on parts and two (2) YEARS ON services
NONE	The brand of the unit to be delivered must have been available in the Philippines for at least 15 years.
NONE	Delivery Schedule: 90CD
16. THERAPEUTIC ULTRASOUND	
With five (5) years warranty on parts and services	With three (3) years warranty on parts and services
NONE	The brand of the unit to be delivered must have been available in the Philippines for at least 15 years
NONE	Training certificate from Manufacturer's of field service engineers performing preventive and corrective maintenance and calibration
NONE	Delivery Schedule: 90CD
NONE	Preventive Maintenance (SEMI ANNUAL) and Calibration (annual)
17. TABLE TOP PULSE OXIMETER	
Nellcor Compatible probes	Compatible probes
4 screen modes	3 - 4 screen modes
15 days trend memory /10seconds	10 - 15 days trend memory /10seconds
Alarm Paused Tone - 30, 60, 90, 120 seconds	Alarm Paused Tone - 1, 2, 3 MINS
Trends Memory - Save continuously for 15 days (for 10 seconds saving period)	Trends Memory - Save continuously for 10 - 15 days (for 10 seconds saving period)
NONE	THREE (3) years warranty on parts and two (2) YEASR ON services
NONE	The brand of the unit to be delivered must have been available in the Philippines for at least 15 years
NONE	Training certificate from Manufacturer's of field service engineers performing preventive and corrective maintenance and calibration
NONE	Delivery Schedule: 90CD
NONE	Preventive Maintenance (SEMI ANNUAL) and Calibration (annual)
18. PORTABLE SUCTION MACHINE with CART, HEAVY DUTY	
Max. vacuum: ≥ 680 mmHg	Max. vacuum: ≥ 600 - 700 mmHg
Flow rate: ≥ 32 L/min	Flow rate: ≥ 15 - 30 L/min
Input power: 150 VA	Input power: 100 - 150 VA
Gross Weight: at least 6kg	Gross Weight: at least 5 kg

"PHIC Accredited Healthcare Provider"

"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
 Telephone Nos: 8294-6711 to 17
 Director's Office Direct Line: 8291-4259
 Email Address: valgen_hosp@yahoo.com
 Website: <https://vmc.doh.gov.ph/>

NONE	THREE (3) years warranty on parts and two (2) YEARS ON services
NONE	The brand of the unit to be delivered must have been available in the Philippines for at least 15 years.
NONE	Training certificate from Manufacturer's of field service engineers performing preventive and corrective maintenance and calibration
NONE	Delivery Schedule: 90CD
NONE	Preventive Maintenance (SEMI ANNUAL) and Calibration (annual)
19. PHARMACEUTICAL REFRIGERATOR	
95% uptime and 98% downtime	95% uptime and 2% downtime
NONE	The brand of the unit to be delivered must have been available in the Philippines for at least 15 years.
NONE	Delivery Schedule: 60CD
20. LARYNGOSCOPE SET	
Fiber Optic Miller Blades with standard 2.5V LED	Fiber Optic Miller Blades with standard 2.5-3.5V LED
Miller 00 (Overall length) 76mm , (Blade length) 51mm ,	Miller 00 (Overall length) 65 - 76mm , (Blade length) 42 - 51 mm
(Distal width) 10mm	(Distal width) 10 - 11.3 mm
Miller 0 (Overall length) 80mm , (Blade length) 55mm,	Miller 0 (Overall length) 77 - 80 mm , (Blade length) 54 - 55 mm
(Distal width) 11mm	(Distal width) 11 - 11.3 mm
Miller 1 (Overall length) 100mm, (Blade length) 78mm (Distal width) 12mm	Miller 1 (Overall length) 100 mm, (Blade length) 78 - 79 (distal width) 11.3 - 12 mm
With up to 6,500 individual micro-fibers for improved light transmission and longer life.	With up to 6,500 (OR ITS EQUIVALENT) individual micro-fibers for improved light transmission and longer life.
Fade-out feature: brightness reduces slowly even with low residual capacity	Fade-out feature: brightness reduces slowly even with low residual capacity (OPTIONAL)
Reusable and autoclavable blades up to 4000 cycles	Reusable and autoclavable blades up to 4000 cycles (OR ITS EQUIVALENT)
loaner unit during warranty (to all equipment) with response time of 72 hours	loaner unit during warranty (to all equipment) with response time of 72 hours (REMOVE)
95% uptime and 98% downtime	95% uptime and 98% downtime (REMOVE)
Calibration certificate from manufacturer	Calibration certificate from manufacturer (REMOVE)
NONE	ONE (1) years warranty on parts and ONE(1) YEAR ON services
NONE	The brand of the unit to be delivered must have been available in the Philippines for at least 15 years
NONE	Delivery Schedule: 90CD
21. ELECTROCAUTERY MACHINE	
Quarterly Preventive Maintenance Service (PMS) during warranty period, supplier must have proper equipment for PMS and must submit report	semi annual preventive maintenance and annual calibration
NONE	The brand of the unit to be delivered must have been available in the Philippines for at least 15 years.
NONE	Training certificate from Manufacturer's of field service engineers performing preventive and corrective maintenance and calibration
NONE	Delivery Schedule: 60CD
22. ELECTRICAL STIMULATOR UNIT	
With five (5) years warranty on parts and services	With three (3) years warranty on parts and services
95% uptime and 98% downtime	95% uptime and 2% downtime
NONE	The brand of the unit to be delivered must have been available in the Philippines for at least 15 years.
NONE	Delivery Schedule: 60CD

"PHIC Accredited Healthcare Provider"
"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <https://vmc.doh.gov.ph/>


23. <u>BASSINET WITH STAND AND MATTRESS</u>	
loaner unit during warranty (to all equipment) with response time of 72 hours	loaner unit during warranty (to all equipment) with response time of 72 hours (REMOVE)
95% uptime and 98% downtime	95% uptime and 98% downtime (REMOVE)
Calibration certificate from manufacturer	Calibration certificate from manufacturer (REMOVE)
NONE	TWO (2) years warranty on parts and TWO (2) YEARS ON services
NONE	The brand of the unit to be delivered must have been available in the Philippines for at least 15 years
NONE	Delivery Schedule: 90CD

All other provisions of the bidding documents not affected by revisions included in this Bid Bulletin shall remain in effect. Amendments/inclusions/clarifications made herein shall be considered an integral part of the Bidding Documents.

Prepared by:


LIGAYA E. UBALDE
BAC Secretariat

Noted by:


RUBY S. GURREA, RN, MAN
Chairperson, BAC

“PHIC Accredited Healthcare Provider”
“Valenzuela Medical Center...Where your health matters most”



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director’s Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <https://vmc.doh.gov.ph/>