



Republic of the Philippines
Department of Health
Metro Manila Center for Health Development
VALENZUELA MEDICAL CENTER



**BID BULLETIN NO. 1
27 FEBRUARY 2025**

**SUPPLY AND DELIVERY OF FIFTY (50) SETS OF MECHANICAL HOSPITAL BED
PUBLIC BIDDING NO. VMC-2025-037**

Issued pursuant to Sec. 22.5 of the IRR of RA 9184 to clarify and/or amend certain provisions in the Bidding Documents issued for this project, considering the issues raised and clarifications made by the prospective bidders during the Pre-bid Conference held on February 27, 2025.

Posted in Bidding Documents	Amendment/Clarifications
1. MECHANICAL HOSPITAL BED	
NONE	Size: W-900mm to 1000mm x L-1800mm to 2300mm
Three (3) cranks, Heavy Duty	At least three (3) cranks, Heavy Duty
Standard-type channel frame	Standard-type channel/tubular frame (Manufacturer's standard)
5" heavy-duty rubber caster wheels	4"- 5" heavy-duty rubber caster wheel lock
Double row thrust ball bearing mechanism	Double row thrust ball bearing mechanism (Manufacturer's standard)
Solid metal (epoxy powder coated frame) bed board with ventilating holes	Solid metal (epoxy powder coated frame) bed board with ventilating holes (Solid metal according to manufacturer's standard)
IV pole sockets on corners with two (2) pcs IV pole with 4 hooks	IV pole sockets on corners with two (2) pcs IV pole, stainless with 4 hooks, adjustable height
Stainless type sliding railing at least 18" in height	Stainless type or metal coated sliding railing/collapsible or split type at least 16" in height from mattress platform
4" thick high density and heavy-duty foam mattress (with lifetime warranty) with leatherette cover, color black or gray	4" thick high density and heavy-duty foam mattress (with two (2) years warranty) with leatherette cover, color: manufacturers standard
Load capacity: 300kg.	Load capacity: able to bear 220kgs.
Bedside Table:	Bedside Table:
Powder coated steel metal body	Powder coated steel metal body or ABS material
Certificates to be submitted during Post-Qualification:	Certificates to be submitted during Post-Qualification:
1. FDA CMDN	1. FDA CMDN or Proof of Application at least six (6) months from the date of opening.

"PHIC Accredited Healthcare Provider"
"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karubatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <https://vmc.doh.gov.ph/>

Certificates to be submitted upon delivery:	Certificates to be submitted upon delivery:
Warranty: Certificate of two (2) years comprehensive warranty which states the following clauses: The period of warranty shall be deemed to be fully comprehensive, all inclusive warranty, inclusive of labor, all spare parts, all accessories, all service consumables, manufacturers preventive maintenance and unlimited number of breakdown calls.	Warranty: Notarized Certificate of two (2) years comprehensive warranty which states the following clauses: The period of warranty shall be deemed to be fully comprehensive, all inclusive warranty, inclusive of labor, all spare parts, all accessories, all service consumables, manufacturers preventive maintenance and unlimited number of breakdown calls.
Certificate of Semi-annual Preventive Maintenance Service (PMS) within warranty period	Notarized Certificate of Semi-annual Preventive Maintenance Service (PMS) within warranty period


DELIVERY SCHEDULE: 60CD
BID OPENING: MARCH 11, 2025 at 11:00 A.M.

All other provisions of the bidding documents not affected by revisions included in this Bid Bulletin shall remain in effect. Amendments/inclusions/clarifications made herein shall be considered an integral part of the Bidding Documents.

Prepared by:


MS. KRISTINE JOY MANUEL
BAC Secretariat

Noted by:


MS. RUBY S. GURREA, RN, MAN
Chairperson, BAC II

“PHIC Accredited Healthcare Provider”
“Valenzuela Medical Center...Where your health matters most”