




BID BULLETIN NO. 1
11 April 2025


Public Bidding No. VMC-2025-053
SUPPLY AND DELIVERY OF VARIOUS PM & R MEDICAL EQUIPMENT – BATCH 2

Issued pursuant to Sec. 22.5 of the IRR of RA 9184 to clarify and/or amend certain provisions in the Bidding Documents issued for this project, considering the issues raised and clarifications made by the prospective bidders during the **Pre-Bid Conference** held on April 11, 2025.

Posted in Bidding Documents	Amendment
TECAR MACHINE Frequencies between: 460-540 MHz	Frequencies between: 460-540 KHz
CRYOTHERAPY MACHINE Dry airflow, least of -40 and below	Dry airflow, at least -40 degree c and below
CRYOTHERAPY MACHINE Flexible Treatment Hose Holder	Flexible Treatment Hose Holder/ arm

All other provisions of the bidding documents not affected by revisions included in this Bid Bulletin shall remain in effect. Amendments/inclusions/clarifications made herein shall be considered an integral part of the Bidding Documents.

Prepared by:

CHRISTALLYNE S. CASTRO
BAC Secretariat

Noted by:

MS. RUBY S. GURREA, RN, MAN
Chairperson, BAC

"PHIC Accredited Healthcare Provider"
"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <https://vmc.doh.gov.ph/>