



Bid Notice Abstract

Request for Quotation (RFQ)

Reference Number 11410283  
Procuring Entity VALENZUELA MEDICAL CENTER  
Title SUPPLY, DELIVERY & TESTING OF DRY BATH MACHINE  
Area of Delivery Metro Manila

<b>Solicitation Number:</b>	VMC-2024-206	<b>Status</b>	Active
<b>Trade Agreement:</b>	Implementing Rules and Regulations		
<b>Procurement Mode:</b>	Negotiated Procurement - Small Value Procurement (Sec. 53.9)	<b>Associated Components</b>	1
<b>Classification:</b>	Goods	<b>Bid Supplements</b>	0
<b>Category:</b>	General Repair and Maintenance Services		
<b>Approved Budget for the Contract:</b>	PHP 56,000.00	<b>Document Request List</b>	0
<b>Delivery Period:</b>	30 Day/s		
<b>Client Agency:</b>		<b>Date Published</b>	29/10/2024
<b>Contact Person:</b>	LESTER JOHN JAKE R DIVINO BAC SECRETARIAT Padrigal St., Karuhatan, Valenzuela City Metro Manila Philippines 63-2-2920604  vmc_bac@yahoo.com	<b>Last Updated / Time</b>	29/10/2024 00:00 AM
		<b>Closing Date / Time</b>	04/11/2024 10:00 AM

<b>Description</b>
STOCK/ PROPERTY NO. UNIT ITEM DESCRIPTION QTY. UNIT COST TOTAL COST
SUPPLY, DELIVERY & TESTING OF unit DRY BATH MACHINE 1 56,000.00 56,000.00 Block Capacity: 2 Temperature Readout: Thermometer Temperature Setting: Analog Temperature Range: Ambient ±5 °C to 125 °C Temperature Control: Analog IC Temperature Sensor: NTC Thermistors Uniformity: ±5 °C at 37 °C Stability: ±5 °C Power Supply: 220V/60Hz Test tubes insert to aluminum blocks which maintain temperature stability with ISO 901 Quality Certification Issuance of Calibration Certificate Warranty: 6 months for parts and services Delivery: 30 calendar days upon receipt of Purchase Order **** nothing follows **** 56,000.00
A copy of your UPDATED legal requirements (Certificate of Registration from BIR, SEC/DTI, Business/Mayor's Permit with Official Receipt, PhilGEPS Certificate, Tax Clearance, Audited FS with 2023 ITR, and Omnibus Sworn Statement) is also required to be submitted along with your quotation/proposal.

**Other Information**

Quotations may be submitted manually at the office of the BAC Secretariat, BAC Office, Valenzuela Medical Center, Padrigal St., Karuhatan, Valenzuela City., through email to [vmc\\_bac@yahoo.com](mailto:vmc_bac@yahoo.com). For any clarification, you may call Lester John Jake R. Divino at telephone no. 294-46-25 or via email at [vmc\\_bac@yahoo.com](mailto:vmc_bac@yahoo.com)

**Created by**                   LESTER JOHN JAKE R DIVINO  
**Date Created**            28/10/2024

The PhilGEPS team is not responsible for any typographical errors or misinformation presented in the system. PhilGEPS only displays information provided for by its clients, and any queries regarding the postings should be directed to the contact person/s of the concerned party.