



**BID BULLETIN NO. 1
11 OCTOBER 2024**

**SUPPLY AND DELIVERY OF VARIOUS MEDICAL EQUIPMENT – BATCH 2
(20 ITEMS)**

PUBLIC BIDDING NO. VMC-2024-042

Issued pursuant to Sec. 22.5 of the IRR of RA 9184 to clarify and/or amend certain provisions in the Bidding Documents issued for this project, considering the issues raised and clarifications made by the prospective bidders during the Pre-bid Conference held on October 9, 2024, likewise, respond to bidder’s written queries received within the prescriptive period for filing.

Posted in Bidding Documents	Amendment/Clarifications
1. BRAND NEW FULL DIGITAL RADIOGRAPHY WITH FLUOROSCOPY (CEILING MOUNTED)	
FLAT PANEL DETECTOR	FLAT PANEL DETECTOR
a. 17"x17" or 43 x 43 cm (portable detachable type)	a. 17"x17" or 43 x 43 cm
H. Standard accessories	H. Standard accessories
g. Two (2) double fluoroscopy pedals	g. One (1) or Two (2) double fluoroscopy pedals
L. MONITOR	L. MONITOR
b. One high quality color monitor for acquisition station at least 23"	b. One high quality color monitor for acquisition station at least 21-23"
O. DICOM FEATURES	O. DICOM FEATURES
d. Compatible image (AVI/BMP/JPG/DICOM)	d. Compatible image to DICOM convertible to AVI/BMP/JPG or compatible images with AVI/BMP/JPG/DICOM
P. PRINTING	P. PRINTING
d. Compatible image (AVI/BMP/JPG/DICOM)	d. Compatible image to DICOM convertible to AVI/BMP/JPG or compatible images with AVI/BMP/JPG/DICOM
Q. 4 FLAT PANEL DETECTORS (FPD): 17" X 17"	Q. 3 FLAT PANEL DETECTORS (FPD): 17" X 17"
a. 2 FPD for the table detector-for both fluoroscopy and radiology exams	a. 1 FPD for the table detector-for both fluoroscopy and radiology exams

"PHIC Accredited Healthcare Provider"
"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <https://vmc.doh.gov.ph/>

W. MANUFACTURING COMPANY/BRAND	W. MANUFACTURING COMPANY/BRAND
f. Manufacturing company must have service engineers within the philippines who will service the machine	f. Manufacturing/DISTRIBUTOR company must have service engineers within the philippines who will service the machine
X. DOCUMENTS	X. DOCUMENTS
c. Manufacturer's Certificate must be in the local market for at least 15 years	c. Manufacturer's BRAND must be in the local market for at least 15 years
Y. DOCUMENTS, TRAINING AND MANUALS	Y. DOCUMENTS, TRAINING AND MANUALS
a. Submit Technical Specifications, product brochure, proof of compliance, and other relevant documents required by the bidding documents in English language, both hard copy and soft copy (in CD or USB)	a. Submit Technical Specifications, product brochure, proof of compliance, and other relevant documents required by the bidding documents in English language, both hard copy and soft copy (in USB)
h. Technical Manual in English Language for all equipment including peripherals, Uninterruptible Power Supply (UPS), Automatic Voltage Regulator (AVR), etc. User's Operations Manual - Two (2) sets original manuals upon delivery and two (2) duplicate copies prior to processing of payment - One (1) copy of operation manuals in DVD/USB upon delivery Quality Control & Maintenance Manual - Two (2) sets original manuals upon delivery and two (2) duplicate copies prior to processing of payment - One (1) copy of QC and Maintenance manual in DVD/USB upon delivery Service and Technical Manuals - Two (2) sets original manuals upon delivery and two (2) duplicate copies prior to processing of payment - One (1) copy of service and technical manuals in DVD/USB upon delivery	h. Technical Manual in English Language for all equipment including peripherals, Uninterruptible Power Supply (UPS), Automatic Voltage Regulator (AVR), etc. User's Operations Manual - Two (2) sets original manuals upon delivery and two (2) duplicate copies prior to processing of payment - One (1) copy of operation manuals in USB upon delivery Quality Control & Maintenance Manual - Two (2) sets original manuals upon delivery and two (2) duplicate copies prior to processing of payment - One (1) copy of QC and Maintenance manual in USB upon delivery Service and Technical Manuals - Two (2) sets original manuals upon delivery and two (2) duplicate copies prior to processing of payment - One (1) copy of service and technical manuals in USB upon delivery
NONE	DELIVERY SCHEDULE: 120CD
2. OPHTHALMIC TAB LASER	2. OPHTHALMIC YAG LASER
Accessories: 1 pc. Irodotomy (Iridectomy) and 1pc Capsulotomy Lenses, 1pc. Wooden and stackable elbow rest, 1 pc. SLT Latina Lenses	Accessories: 1 pc. Irodotomy (Iridectomy) and 1pc Capsulotomy Lenses, 1pc. Wooden and stackable elbow rest
NONE	DELIVERY SCHEDULE: 90CD
4. SURGICAL HEADLIGHT WITH LOUPE/LENS	
Lens magnification – 1.5x / 2.0x / 2.5x (if loupes are unavailable)	Lens magnification – 2.5x - 3.5x (if loupes are unavailable)
Loupes magnification – 2.0/2.5x	Loupes magnification – 2.5x - 3.5x
Mount type: Preferably Ergonomic C-band type	Mount type: Preferably Ergonomic round-head mounted type (adjustable)
Weight: 100g to 205g	Weight: at least 70g
5. HUMAN MILK PASTEURIZER	
g. Treatment of water - Bathwater filtration above 0.2 microns to avoid any contamination. (3 Stages of water filtration)	g. Treatment of water - Bathwater filtration above 0.2 microns to avoid any contamination. (3 Stages of water filtration WITH 2 BUILT-IN WATER FILTERATION IN MACHINE)

"PHIC Accredited Healthcare Provider"

"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
 Telephone Nos: 8294-6711 to 17
 Director's Office Direct Line: 8291-4259
 Email Address: valgen_hosp@yahoo.com
 Website: <https://vmc.doh.gov.ph/>

1 unit Laptop (applicable for machine)	1 unit Laptop (applicable for machine) FOR SIMULTANEOUS DISPLAY BOF PARAMETERS TO THE SCREEN AND LAPTOP
The equipment or device shall conform to the following 1. IEC 60601 Standard 2. Any of the following: USFDA (food and drugs administration, u.s) Standard CE (European Conformity) Standard EN (European Standards) ISO 9000, 9001, 9002 Standards UL (Underwriters laboratories) Standards 3. The proposed unit shall also conform to all national and local standards and requirements of medical device regulatory agencies. Supplier shall list standards met by the supplier.	The equipment or device shall conform to the following 1. Any of the following: USFDA (food and drugs administration, u.s) Standard CE (European Conformity) Standard EN (European Standards) ISO 9000, 9001, 9002 Standards UL (Underwriters laboratories) Standards 2. The proposed unit shall also conform to all national and local standards and requirements of medical device regulatory agencies. Supplier shall list standards met by the supplier.
NONE	WARRANTY: 2 YEARS FOR PARTS AND SERVICES
DELIVERY SCHEDULE: 60CD	DELIVERY SCHEDULE: 120CD
6. VERTICAL LAMINAR FLOW	
Preventive Maintenance (QUARTERLY) and Calibration (annual) during warranty period Two (2) years warranty on parts and (3) YEARS ON services	Preventive Maintenance (SEMI-ANNUAL) and Calibration (annual) during warranty period Two (2) years warranty on parts and services
The equipment or device shall conform to the following 1. IEC 60601 Standard 2. Any of the following: USFDA (food and drugs administration, u.s) Standard CE (European Conformity) Standard EN (European Standards) ISO 9000, 9001, 9002 Standards UL (Underwriters laboratories) Standards 3. The proposed unit shall also conform to all national and local standards and requirements of medical device regulatory agencies. Supplier shall list standards met by the supplier.	The equipment or device shall conform to the following 1. Any of the following: USFDA (food and drugs administration, u.s) Standard CE (European Conformity) Standard EN (European Standards) ISO 9000, 9001, 9002 Standards UL (Underwriters laboratories) Standards 2. The proposed unit shall also conform to all national and local standards and requirements of medical device regulatory agencies. Supplier shall list standards met by the supplier.
DELIVERY SCHEDULE: 60CD	DELIVERY SCHEDULE: 120CD
7. LABORATORY FREEZER (MILKBANK FREEZER)	
*Design and materials	*Design and materials
Material of adjustable shelves in the freezer compartment: Glass	Material of adjustable shelves in the freezer compartment: Glass or plastic coated grids
Manufacturer's Certificate Periodic quarterly preventive maintenance and calibration for (3) years to provide schedule, service report with checklist, 3 YEARS WARRANTY ON PARTS AND SERVICE	Manufacturer's Certificate Periodic semi-annual preventive maintenance and calibration for (2) years to provide schedule, service report with checklist, 2 YEARS WARRANTY ON PARTS AND SERVICE
The equipment or device shall conform to the following 1. IEC 60601 Standard 2. Any of the following: USFDA (food and drugs administration, u.s) Standard	The equipment or device shall conform to the following 1. Any of the following: USFDA (food and drugs administration, u.s) Standard CE (European Conformity) Standard

CE (European Conformity) Standard EN (European Standards) ISO 9000, 9001, 9002 Standards UL (Underwriters laboratories) Standards 3. The proposed unit shall also conform to all national and local standards and requirements of medical device regulatory agencies. Supplier shall list standards met by the supplier.	EN (European Standards) ISO 9000, 9001, 9002 Standards UL (Underwriters laboratories) Standards 2. The proposed unit shall also conform to all national and local standards and requirements of medical device regulatory agencies. Supplier shall list standards met by the supplier.
DELIVERY SCHEDULE: 60CD	DELIVERY SCHEDULE: 120CD
8. LABORATORY REFRIGERATOR (MILKBANK REFRIGERATOR)	
Manufacturer's Certificate Periodic quarterly preventive maintenance and calibration for (3) years to provide schedule, service report with checklist, 3 YEARS WARRANTY ON PARTS AND SERVICE	Manufacturer's Certificate Periodic semi-annual preventive maintenance and calibration for (2) years to provide schedule, service report with checklist, 2 YEARS WARRANTY ON PARTS AND SERVICE
The equipment or device shall conform to the following 1. IEC 60601 Standard 2. Any of the following: USFDA (food and drugs administration, u.s) Standard CE (European Conformity) Standard EN (European Standards) ISO 9000, 9001, 9002 Standards UL (Underwriters laboratories) Standards 3. The proposed unit shall also conform to all national and local standards and requirements of medical device regulatory agencies. Supplier shall list standards met by the supplier.	The equipment or device shall conform to the following 1. Any of the following: USFDA (food and drugs administration, u.s) Standard CE (European Conformity) Standard EN (European Standards) ISO 9000, 9001, 9002 Standards UL (Underwriters laboratories) Standards 2. The proposed unit shall also conform to all national and local standards and requirements of medical device regulatory agencies. Supplier shall list standards met by the supplier.
DELIVERY SCHEDULE: 60CD	DELIVERY SCHEDULE: 120CD
9. MILKBANK LABELLING SYSTEM	
a. TT Printer 300dpi, APAC cord bundle, USB, LAN	a. TT Printer 200-300dpi, APAC cord bundle, USB, LAN
Manufacturer's Certificate Periodic quarterly preventive maintenance and calibration for (3) years to provide schedule, service report with checklist, 3 YEARS WARRANTY ON PARTS AND SERVICE	Manufacturer's Certificate Periodic semi-annual preventive maintenance and calibration for (2) years to provide schedule, service report with checklist, 2 YEARS WARRANTY ON PARTS AND SERVICE
DELIVERY SCHEDULE: 60CD	DELIVERY SCHEDULE: 120CD
10. TABLE TOP PULSE OXIMETER	
IV pole mounting support	IV pole or trolley mounting support
IV pole clamp (2 pcs)	IV pole clamp or trolley (3 pcs)
loaner unit during warranty (to all equipment) with response time of 72 hours	REMOVE
11. PORTABLE ULTRASOUND/ECHOCARDIOGRAPHY MACHINE	
Electrical Power	Electrical Power
Power consumption maximum of 500VA with peripherals	Power consumption maximum of 160-500VA with peripherals
CONSOLE TYPE OR LAPTOP DESIGN	LAPTOP DESIGN
User Interface Operator keyboard	User Interface Operator keyboard
Full alphanumeric keypad (QWERTY) covered with washable protection film	Full alphanumeric keypad (QWERTY)

"PHIC Accredited Healthcare Provider"
"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <https://vmc.doh.gov.ph/>

Monitor	Monitor
at least 15.6" high-resolution LCD Monitor: 1280 x 800 resolution	at least 15.6" high-resolution LCD OR LED Monitor: 1280 x 800 resolution
Cart Dimension	Cart Dimension
Length - at least 555mm (at least 525mm)	Length - Manufacturer's specs
Depth - at least 515mm	Depth - Manufacturer's specs
Height - at least 890mm	Height - Manufacturer's specs
Weight - at least 21kg	Weight - Manufacturer's specs
Trolley Case	Trolley Case
3 protective compartments for the probes	Protective compartments for the probes - Manufacturer's specs
3 additional compartments for power adapters, cord, and manuals.	Additional compartments for power adapters, cord, and manuals - Manufacturer's specs
Length - at least 495mm	Length - Manufacturer's specs
Depth - at least 275mm	Depth - Manufacturer's specs
Height - at least 460mm	Height - Manufacturer's specs
Weight - at least 4kg	Weight - Manufacturer's specs
Cine memory/ image memory	Cine memory/ image memory
384 MB of Cine Memory	Cine Memory - Manufacturer's specs
Selectable Cine Sequence for Cine review	Selectable Cine Sequence for Cine review - Manufacturer's specs
Prospective Cine Mark	Prospective Cine Mark - Manufacturer's specs
Measurements/calculations and annotations on Cine Playback	Measurements/calculations and annotations on Cine Playback - Manufacturer's specs
Scrolling timeline memory	Scrolling timeline memory - Manufacturer's specs
Dual Image Cine Display	Dual Image Cine Display - Manufacturer's specs
Quad Image Cine Display	Quad Image Cine Display - Manufacturer's specs
Cine gauge and Cine image number display	Cine gauge and Cine image number display - Manufacturer's specs
Cine review loop	Cine review loop - Manufacturer's specs
Cine review speed: at least 11 steps	Cine review speed - Manufacturer's specs
System Standard Features	System Standard Features
NONE	Provides 3,5,7,9 frames of spatial compounding (on and off spatial compounding)
Abstracted from basic user manual	Abstracted from basic user manual OR ITS EQUIVALENT
Transducer Types	Transducer Types
1. 8C-RS OR MICROCONVEX	1. 8C-RS OR MICROCONVEX
Footprint: at least 20 X 15 mm	Footprint: Manufacturer's specs
2. Phased Array Sector probe (Pedia)	2. Phased Array Sector probe (Pedia)
Footprint: at least 23 X 18 mm	Footprint: Manufacturer's specs
3. Phased Array Sector Probe (Neonatal)	3. Phased Array Sector Probe (Neonatal)
Footprint:at least 15.2 x 14.1 mm	Footprint: Manufacturer's specs
12. T-PIECE RESUSCITATOR	
Vacuum Suction Function	Vacuum Suction Function
Vacuum setting knob setting range: 0~ 18.67 #1.33kPa 0~(14010mmHg)	Vacuum setting knob setting range: 0~ 18.67 #1.33kPa 0~(140+-10mmHg)

Vacuum response time: When the input gas source pressure is 500kPa, vacuum in 10 seconds should be at least 17.34kPa (130mmHg)	Vacuum response time: When the input gas source pressure is 500kPa, vacuum in 10 seconds should be at least 17.34kPa (130mmHg)
14. INFANT PHOTOTHERAPY	
LED Lifespan 100,000 hours operating time	LED Lifespan at least 50,000 hours operating time
Wavelength peak between 450 – 475nm, effective surface Area 40x20cm	Wavelength peak between at least 400nm, effective surface Area 40x20cm or its equivalent
Variation in Intensity ± 10	Variation in Intensity ± 10 or its equivalent
Intensity (at 40cm) Low: 25 – 35 μ W/cm ² /nm, High: 35 – 55 μ W/cm ² /nm	Manufacturer's specs
With treatment timer: 30 min - 999hrs/30min	With treatment timer: at least 30 - 1 hr min. - maximum of at least 99hrs - 999hrs.
Input power AC 100-240V (50/60Hz), consumption should be not higher than 70VA	Input power AC 100-240V (50/60Hz), consumption should be not higher than 75VA
Light head dimension & weight must not higher than 340(W) x 210(D) x 75(H)mm, 3.6kg	Light head dimension & weight: Manufacturer's specs
Cart dimension & weight must not higher than 326(W) x 276(D) x 96(H)mm, 8.4kg	Cart dimension & weight: Manufacturer's specs
NONE	Training certificate from Manufacturer's of field service engineers performing preventive and corrective maintenance and calibration
15. PORTABLE SUCTION MACHINE	15. PORTABLE SUCTION MACHINE with CART, HEAVY DUTY
Flow rate: $\geq 15 - 30$ L/min	Flow rate: $\geq 15 - 32$ L/min
Noise: ≤ 60 dB (A)	Noise: ≤ 60 dB (A) or its equivalent
Power Supply: ~ 220 V, 60 Hz	Power Supply: ~ 220 V - 230V, 60 Hz
Input power: 100 - 150 VA	Input power: 78 - 150 VA
Gross Weight: at least 5 kg	Gross Weight: at least 4 kg
Manufacturer's Certificate Brand must be in the local market for at least 15 years	Manufacturer's Certificate Brand must be in the local market for at least 5 years
16. BASSINET WITH STAND AND MATTRESS	
The bassinet basket can hold 13kg weight	The bassinet basket can hold at least 10kg weight
It has an easy-to-group handles for easy transportation. The basket holder grip is made of Polypropylene.	Manufacturer's specs
The mattress is made of urethane foam and Polyester cover, it is water-proof, flame retardant and with MRSA anti-bacterial treatment.	Manufacturer's specs
The bassinet height can be adjusted from 810-812 (Lowest) to 1095-1097mm (Highest) for easy transportation and changing of diapers. The medical personnel can adjust the bassinet to a comfortable height.	The bassinet height can be adjusted from 650-812 (Lowest) to 900-1097mm (Highest) for easy transportation and changing of diapers. The medical personnel can adjust the bassinet to a comfortable height.
Total Length: at least 862mm	Total Length: at least 860mm
Total Width: at least 530mm	Total Width: at least 510mm
Total Height: 810-812mm (Lowest) to 1095-1097mm (Highest)	Total Height: 650-855mm (Lowest) to 900-1097mm (Highest)
Safe working load at least 13kg	Safe working load at least 10kg

"PHIC Accredited Healthcare Provider"

"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
 Telephone Nos: 8294-6711 to 17
 Director's Office Direct Line: 8291-4259
 Email Address: valgen_hosp@yahoo.com
 Website: <https://vmc.doh.gov.ph/>

Mattress dimension at least 660 x 325 x 40mm	Mattress dimension: Manufacturer's specs
17. SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING OF AUTOCLAVE STERILIZER (SQUARE)	
-Quarterly Preventive Maintenance	-Semi-Annual Preventive Maintenance
20. MAJOR INSTRUMENT SET	
1 pc - Scalpel #3	1 pc - Scalpel #3 (4" - 6")
1 pc - Scalpel #4	1 pc - Scalpel #4 (4" - 6")
1 pc - Mayo Scissor Straight 7-8"	1 pc - Mayo Scissor Straight 7-8" (blunt)
1 pc - Metzenbaum Straight 7-8"	1 pc - Metzenbaum Straight 7-8" (blunt)
1 pc - Mayo Scissors Curve 7-8"	1 pc - Mayo Scissors Curve 7-8" (blunt)
1 pc - Metzenbaum Curve 7-8"	1 pc - Metzenbaum Curve 7-8" (blunt)
1 pc - Iris Scissors Straight	1 pc - Iris Scissors Straight (4" - 4.5")
1 pc - Iris Scissors Curve	1 pc - Iris Scissors Curve (4" - 4.5")
2 sets - Self retaining retractor (medium, large) (balfur w/ bladder retractor)	2 sets - Self retaining retractor (max spread of: 200 - 250mm) (balfur w/ bladder retractor)
5 pcs - Mosquito Straight	5 pcs - Mosquito Straight (5 - 5.5")
5 pcs- Mosquito Curve	5 pcs- Mosquito Curve (5 - 5.5")
4 pcs - Babcock 7-8"	4 pcs - Babcock 7-8" (standard)
1 pair - Richardson (double-ended)	1 pair - Richardson (double-ended) (10 - 11")
1 pair - Baby Richardson	1 pair - Baby Richardson (double-ended) (8 - 9")
2 pcs - Mixters (blunt) 7-8"	2 pcs - Mixters clamp (blunt) 7-8"
2 pcs - Mixters (fine tip) 7-8"	2 pcs - Mixters clamp (fine tip) 7-8"
6 pcs - Allis	6 pcs - Allis (7-8") (grasper), with teeth
2 pcs - Kidney Basin 10-11"	2 pcs - Kidney Basin 9-11"
1 pc - Deaver Retractor, narrow	1 pc - Deaver Retractor, narrow (at least 1"W), handgripped
1 pc - Deaver Retractor, medium	1 pc - Deaver Retractor, medium (at least 1.5"W), handgripped
1 pc - Deaver Retractor, wide	1 pc - Deaver Retractor, wide (3 - 3.5"W), handgripped
2 pcs - Ovum Forcep 12"	2 pcs - Ovum Forcep 12.5" (straight)
1 Sterilization Container System	Manufacturer's specs
- Stainless steel 316	Manufacturer's specs
- Hardness (Scissors): 50 to 58 rockwell	Manufacturer's specs
- Have unique device identification code per set; with laser printing of OR - VMC per instrument	Have unique device identification code per set; with laser printing of OR - VMC per instrument (with year acquired)
- With at least 5 years comprehensive warranty on all parts and services	With at least 3 years comprehensive warranty on all parts and services
NONE	Matte finish for all instruments
NONE	CMDM Requirement
NONE	Delivery Schedule: 60CD

"PHIC Accredited Healthcare Provider"

"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
 Telephone Nos: 8294-6711 to 17
 Director's Office Direct Line: 8291-4259
 Email Address: valgen_hosp@yahoo.com
 Website: <https://vmc.doh.gov.ph/>


BID OPENING: OCTOBER 21, 2024 at 10:00 A.M.

All other provisions of the bidding documents not affected by revisions included in this Bid Bulletin shall remain in effect. Amendments/inclusions/clarifications made herein shall be considered an integral part of the Bidding Documents.

Prepared by:


LIGAYA E. UBALDE
BAC Secretariat

Noted by:


RUBY S. CURREA, RN, MAN
Chairperson, BAC

“PHIC Accredited Healthcare Provider”
“Valenzuela Medical Center...Where your health matters most”