



Document Type Quality Manual	Revision No: 0	Date Adopted: 13 Aug 2019	Page No: Page 1 of 2
Document Code: QM 3-05	METROLOGICAL TRACEABILITY		

Introduction

This document provides the principles of measurement traceability of the laboratories of BFAR-RFLD XII according to the requirements of the sub clause 6.5 of the PNS ISO/IEC 17025:2017 standard.

1. Metrological Traceability

All measurements conducted by the Testing Laboratories of BFAR-RFLD XII have traceability either to international standards, if any, SI, or to other national standards. In this respect, depending upon the level of the national standard of the respective laboratory of BFAR-RFLD XII, traceability to other national standards is achieved either through their calibration or through measurement inter-comparisons or through proficiency testing.

The measurement traceability of each quantity is described by the respective traceability chart according to the Laboratory Procedures.

2. Calibration of Laboratory Equipment

All measuring equipment, regardless of the level of its accuracy is calibrated, according to specific procedures and instructions of the quality system, which describe the methods and provide the responsibilities. Calibration intervals of the equipment are specified according to international standards and practices, while the results of calibrations are on continuous monitoring and control.

The program for calibration of equipment demands that calibrations and measurements made by the laboratory are traceable to the International System of Units. Contracted external calibration providing services to the laboratory are to provide documentation demonstrating measurement capability; competence to perform the calibration services and are accredited to ISO/IEC 17025. Calibration certificates issued are to include the measurement results, including the measurement uncertainty and a statement of conformance with an identified metrological specification.

When metrological traceability to the SI units is not technically possible, the laboratory demonstrates metrological traceability to an appropriate reference, e.g.:

- a. certified values of certified reference materials provided by a competent producer;
- b. results of reference measurement procedures, specified methods or consensus standards that are clearly described and accepted as providing measurement results fit for their intended use and ensured by suitable comparison.

Prepared by: MA. MAURICE A. BANQUILING Quality Assurance Manager	Reviewed and Approved by: EUGENE GAY B. JAMORA Laboratory Manager
--	---



Republic of the Philippines
Department of Agriculture
BUREAU OF FISHERIES AND AQUATIC RESOURCES
REGIONAL FISHERIES LABORATORY DIVISION XII
J. Catolico St., Lagao, General Santos City

Document Type Quality Manual	Revision No: 0	Date Adopted: 13 Aug 2019	Page No: Page 2 of 2
Document Code: QM 3-05	METROLOGICAL TRACEABILITY		

4. Documents

SOP-04, Metrological Traceability

5. Document History

Revision No.	Issue Date	Original Entry	Revised as	Recall Date

Prepared by:

MA. MAURICE A. BANQUILING
Quality Assurance Manager

Reviewed and Approved by:

EUGENE GAY B. JAMORA
Laboratory Manager