



Republic of the Philippines
Department of Agriculture
BUREAU OF FISHERIES AND AQUATIC RESOURCES
REGIONAL FISHERIES LABORATORY XII
J. Catolico St., Lagao, General Santos City

Document Type Standard Operating Procedures	Revision No: 0	Date Adopted: 13 Aug 2019	Page No: Page 1 of 2
Document Code: SOP-04	METROLOGICAL TRACEABILITY		

Introduction

This procedure outlines the principles of the programme and procedures used by the Laboratory for the calibration of equipment, with the purpose of ensuring traceability to the fundamental units of SI.

1. Responsibilities

1. Laboratory Manager
2. Quality Assurance Manager
3. Technical Manager
4. Laboratory Staff

2. Principle of the Analyses

2.1. The RFL XII ensures that all equipment for testing that affect the reliability of the test result is calibrated before use. The laboratory follows the procedure for selecting, using, calibrating, checking, controlling and maintaining equipment used to perform test.

3. Calibration

3.1. Calibrations performed on Laboratory's measuring equipment are documented through:

- Calibration certificates issued either by the external laboratory
- Calibration labels, stickers or other appropriate labelling, indicating calibration status including date of calibration

Note: Criteria for selecting external laboratories suitable for providing traceability, are outlined in SOP-05 and accredited by Philippine Accreditation Bureau.

3.2. Every Measuring Instrument, which is considered to be critical for carrying out analyses, is individually registered in Form LF 03-01, List of Laboratory Equipment, as described in WI-01, Selection, Verification and Validation of Methods with information related to its calibration specifications and requirements.

3.3. The laboratory strictly follows the calibration program found in LF 03-03, Calibration Schedule.

Prepared by: MA. MAURICE A. BANQUILING Quality Assurance Manager	Reviewed and Approved by: EUGENE GAY B. JAMORA Laboratory Manager
--	---



Republic of the Philippines
Department of Agriculture
BUREAU OF FISHERIES AND AQUATIC RESOURCES
REGIONAL FISHERIES LABORATORY XII
J. Catolico St., Lagao, General Santos City

Document Type Standard Operating Procedures	Revision No: 0	Date Adopted: 13 Aug 2019	Page No: Page 2 of 2
Document Code: SOP-04	METROLOGICAL TRACEABILITY		

3.4. The calibration of equipment and glassware is done annually.

4. Eligible suppliers of Standards and Calibration Services

4.1. Approved external suppliers of services are the following:

- DOST-ITDI-NML
- Accredited Laboratories by Philippine Accreditation Bureau (PAB)
- In the absence of the above, laboratories which have been assessed successfully and approved by the Laboratory
- Suppliers of Standards and Reference Materials with traceability to national/international standards and SI.

5. Documents

LF 03-01, List of Laboratory Equipment

LF 03-02, Equipment History

LF 03-03, Calibration Schedule

WI-01, Selection, Verification and Validation of Methods

6. Document History

Revision No.	Issue Date	Original Entry	Revised as	Recall Date

Prepared by:

MA. MAURICE A. BANQUILING
Quality Assurance Manager

Reviewed and Approved by:

EUGENE GAY B. JAMORA
Laboratory Manager