Periodic Safety Update Report

Endotracheal tubes — extract: PMS information (b) to (d), MDCG 2022-21 Annex I

Cover page item	Entry
Manufacturer	Tianjin Medis Medical Device Co., Ltd. (SRN CN-MF-000020126)
Medical device(s) covered	Endotracheal tubes — Basic UDI-DI 69206873ETT002PF; 236 UDI-DIs; Class IIa; EMDN R01030201 ENDOTRACHEAL TUBES, CUFFED, NOT REINFORCED
Notified body	NB 0197 — certificate HZ 2309178-1
PSUR reference number / version	[manufacturer]
Data collection period	2024-10-07 to 2026-10-08

Post-Market Surveillance: information including general PMCF information

b) Scientific literature review of relevant specialist or technical literature

A literature search was performed on 2026-10-08 in PubMed/MEDLINE, Europe PMC and Crossref for records first available between 2024-10-07 and 2026-10-08, using a structured search for the device type, its synonyms and subject headings (full search strings in the search documentation). 1514 records were screened on title and abstract; 227 met the inclusion criteria (the device type is a subject of the publication and safety, performance or clinical-use findings are reported), of which 122 report safety-related findings. Publications concerning the device(s) covered by this PSUR: 0.

#	Publication	Finding type
1	Mester P et al. (2026). An unusual cause of sudden complete endotracheal tube occlusion. International journal of infectious diseases.	Safety
2	Chen R et al. (2026). Postoperative pulmonary complications after totally thoracoscopic cardiac surgery: a retrospective cohort study comparing single-lumen and double-lumen endotracheal tubes. Frontiers in medicine.	Safety
3	Kim MJ et al. (2026). Effect of Intracuff Pressure Adjustment on Postoperative Sore Throat, Hoarseness, and Oral Comfort in Robotic Gynecological Surgery: A Quasi-experimental Study. Journal of perianesthesia nursing.	Safety
4	Onda T et al. (2026). [Tracheal Membranous Wall Injury Secondary to Endotracheal Intubation:Report of a Case]. Kyobu geka. The Japanese journal of thoracic surgery.	Safety
5	Zhao J et al. (2026). Development of a Protective Device to Prevent Damage to an Endotracheal Tube Caused by the Eruption of Deciduous Teeth. Reports (MDPI).	Safety

Table continues: 222 further publications.

Manufacturer's assessment against the risk management file and clinical evaluation: [to be completed by the manufacturer].

c) Public databases and/or registry data

Database reviewed	Type	Period	Similar devices: findings
FDA MAUDE (USA)	Adverse event reports	2024-10-07 – 2026-10-08	3,196 reports (death 18, injury 346, malfunction 2,830); product code BTR
FDA Medical Device Recalls (USA)	Field safety corrective actions	5 years to 2026-10-08	4 recall events, 2 in the data collection period
Health Canada Recalls and Safety Alerts	Field safety corrective actions	5 years	5 recalls
MHRA Field Safety Notices (UK)	Field safety notices	5 years	4 notices
EUDAMED (EU)	Device registration	at 2026-10-08	2,752 UDI-DIs registered under EMDN R01030201

Most frequent device problems, similar devices (FDA MAUDE)	Reports
Gas/Air Leak	943
Use of Device Problem	904
Material Split, Cut or Torn	206
Disconnection	189
Leak/Splash	119

Most frequent health effects, similar devices (FDA MAUDE)	Reports
Unintended Extubation	757
Low Oxygen Saturation	104
Hypoventilation	90
Cardiac Arrest	54

57% of the MAUDE reports for similar devices in the period were submitted by one manufacturer (Covidien Juarez MMJ); year-on-year changes mainly reflect that manufacturer's reporting. Reports concerning the device(s) covered by this PSUR are excluded from these figures.

Date	Firm / authority	Field action (excerpt)
2026-07-31	WELL LEAD MEDICAL CO. LTD	The 6Fr suction catheter could not pass through the samples smoothly.
2025-02-13	Smiths Medical ASD, Inc.	Affected devices have a smaller diameter than expected and may potentially result in inadequate ventilation to the patient. Patient may experience hypoxia, …
2024-02-26	MEDLINE INDUSTRIES, LP - Northfield	Complaints have been received that the device inflation tube detached and/or tore from the main tube, resulting in potential moisture buildup, loss of …
2026-08-14	Health Canada	Endotracheal Tube, Cuffed
2025-02-26	Health Canada	Smith Medical Intubation Oral/Nasal Endotracheal Tubes
2025-02-25	Smiths Medical	Intubation ORAL/NASAL Endotracheal Tube
2023-12-12	Medtronic	NIM Trivantage EMG Endotracheal tube

Registries: no device registry with public device-level data was identified for this device type.

Manufacturer's assessment against the risk management file and clinical evaluation: [to be completed by the manufacturer].

d) Publicly available information about similar medical devices

Similar devices are those registered under EMDN R01030201 (ENDOTRACHEAL TUBES, CUFFED, NOT REINFORCED) and, for third-country data, FDA product code BTR. In EUDAMED, 12 other manufacturers were identified in a sample of 300 registrations. In the USA, 9 new similar devices received 510(k) clearance in the period; the 510(k) summaries describe indications for use and performance testing.

Decision	510(k)	Applicant	Device
2026-09-03	K251794	Bryan Medical, Inc.	Tenax® Laser Resistant Endotracheal Tube
2026-06-12	K253437	Spiro Robotics	Spiro-VISTA
2025-10-06	K250173	Covidien, LLC	Shiley™ Oral RAE Tracheal Tube Cuffless, Murphy Eye (113-XX) Shiley™ Nasal RAE Tracheal Tube Cuffless, Murphy
2025-09-25	K243785	Covidien, LLC	Shiley™ Oral/Nasal Tracheal Tube with TaperGuard™ Cuff Reinforced, Murphy Eye Shiley™ Tracheal Tube TaperGuar

Manufacturer of similar devices (EU market)	Country
Henan Tuoren Medical Device Co., Ltd	China
Well Lead Medical Co., Ltd.	China
Hangzhou Tappa Medical Technology Co , Ltd.	China
Hangzhou Shanyou Medical Equipment Co., Ltd.	China

Summary of safety and clinical performance (SS(C)P): not applicable — the device is Class IIa and non-implantable (MDR Art. 32).

Manufacturer's assessment against the risk management file and clinical evaluation: [to be completed by the manufacturer].