Periodic Safety Update Report

LightWalker/TimeWalker — extract: PMS information (b) to (d), MDCG 2022-21 Annex I

Cover page item	Entry
Manufacturer	FOTONA, proizvodnja optoelektronskih naprav d.o.o. (SRN SI-MF-000000126)
Medical device(s) covered	LightWalker/TimeWalker — Basic UDI-DI 3830054230004S7; 11 UDI-DIs; Class IIb; EMDN Z12011019 MULTI–PLATFORM LASER
Notified body	NB 0123 — certificate G10 025155 0029
PSUR reference number / version	[manufacturer]
Data collection period	2025-10-08 to 2026-10-08

Post-Market Surveillance: information including general PMCF information

b) Scientific literature review of relevant specialist or technical literature

A literature search was performed on 2026-10-08 in PubMed/MEDLINE, Europe PMC and Crossref for records first available between 2025-10-08 and 2026-10-08, using a structured search for the device type, its synonyms and subject headings (full search strings in the search documentation). 2286 records were screened on title and abstract; 613 met the inclusion criteria (the device type is a subject of the publication and safety, performance or clinical-use findings are reported), of which 61 report safety-related findings. Publications concerning the device(s) covered by this PSUR: 0.

#	Publication	Finding type
1	Benosman S et al. (2026). Insulin Resistance as a Potential Predictor of Post-Inflammatory Hyperpigmentation After CO2 Laser Resurfacing: A Real-World Study of a Dual Technique Protocol (Maghreb Experience). Clinical, cosmetic and investigational dermatology.	Safety
2	Su T et al. (2026). Short-term efficacy and safety of CO2 laser vaporization for anal intraepithelial neoplasia. Frontiers in medicine.	Safety
3	Bajerski JC et al. (2026). Transdermal 1064-NM ND:YAG laser for reticular veins larger than 4 mm: ONE-YEAR outcomes and seven-year follow-up. Phlebology.	Safety
4	Bibi Z et al. (2026). Fully ablative CO2 laser excision of cutaneous lesions: A case series including patients with cardiac implantable electronic devices. Clinical and experimental dermatology.	Safety
5	Almahzoum K et al. (2026). Comparative Efficacy and Safety of Needling-Based Therapies Versus Ablative Fractional Lasers for Scar Management: A Systematic Review and an Exploratory Meta-Analysis. Photodermatology, photoimmunology & photomedicine.	Safety
6	Chua KR et al. (2026). Efficacy and Safety of 755-nm Picosecond Alexandrite Laser for Acquired Bilateral Naevus of Ota-Like Macules (Hori Naevus): A Systematic Review and Meta-Analysis. The Australasian journal of dermatology.	Safety

Table continues: 607 further publications.

Manufacturer's assessment against the risk management file and clinical evaluation: [to be completed by the manufacturer].

c) Public databases and/or registry data

Database reviewed	Type	Period	Similar devices: findings
FDA MAUDE (USA)	Adverse event reports	2025-10-08 – 2026-10-08	759 reports (death 4, injury 303, malfunction 445); product code GEX
FDA Medical Device Recalls (USA)	Field safety corrective actions	5 years to 2026-10-08	9 recall events, 4 in the data collection period
Health Canada Recalls and Safety Alerts	Field safety corrective actions	5 years	6 recalls
MHRA Field Safety Notices (UK)	Field safety notices	5 years	3 notices
EUDAMED (EU)	Device registration	at 2026-10-08	287 UDI-DIs registered under EMDN Z12011019

Most frequent device problems, similar devices (FDA MAUDE)	Reports
Adverse Event Without Identified Device or Use Problem	162
Break	81
Nonstandard Device	61
Failure to Align	58
Output below Specifications	52

Most frequent health effects, similar devices (FDA MAUDE)	Reports
Burn(s)	89
Eye Injury	49
Blister	26
Full thickness (Third Degree) Burn	21

Date	Firm / authority	Field action (excerpt)
2026-08-06	Lutronic Corporation	Utilizing higher fluences at 1064nm wavelength with current cryogen parameters may result in a spark/loud noise during treatment, potentially resulting in …
2026-05-05	IPG Medical Corporation	Software anomaly that causes a false display of error code 5018.
2026-01-27	Lutronic Corporation	Reports of devices sparking/popping and potentially burning patients.
2026-08-26	Health Canada	Cyanosure DermaV Laser System
2026-05-13	Health Canada	VersaPulse Dual Wavelength and CO2 Laser Systems
2026-03-20	Gyrus	Soltive Laser System
2024-03-07	Olympus	SOLTIVE SuperPulsed Laser System

Registries: no device registry with public device-level data was identified for this device type.

Manufacturer's assessment against the risk management file and clinical evaluation: [to be completed by the manufacturer].

d) Publicly available information about similar medical devices

Similar devices are those registered under EMDN Z12011019 (MULTI–PLATFORM LASER) and, for third-country data, FDA product code GEX. In EUDAMED, 15 other manufacturers were identified in a sample of 267 registrations. In the USA, 70 new similar devices received 510(k) clearance in the period; the 510(k) summaries describe indications for use and performance testing.

Decision	510(k)	Applicant	Device
2026-09-24	K254199	BLUECORE COMPANY Co., Ltd.	PicoreMax (PicoreMax)
2026-09-19	K241973	Lumenis Be, Ltd.	SmartIPL_NI System
2026-09-18	K261047	Guilin Kevin Peter Technology Co., Ltd.	Dental Diode Laser Device (Laser-Pen)
2026-09-17	K261817	Zhengzhou PZ Laser Slim Technology Co., Ltd.	Diode Laser Therapy System (PZ-BDT67-01, PZ-BDT67-02)

Manufacturer of similar devices (EU market)	Country
Quanta System S.p.A.	Italy
Lumenis Be LTD	Israel
Cynosure Lutronic Technology Corporation	South Korea
Lutronic Corporation	South Korea

Summary of safety and clinical performance (SS(C)P): not applicable — the device is Class IIb and non-implantable (MDR Art. 32).

Manufacturer's assessment against the risk management file and clinical evaluation: [to be completed by the manufacturer].