# Periodic Safety Update Report

A32BL-Set — extract: PMS information (b) to (d), MDCG 2022-21 Annex I

| Cover page item | Entry |
|---|---|
| Manufacturer | Anaesthetic Medical Systems Ltd (SRN GB-MF-000016544) |
| Medical device(s) covered | A32BL-Set — Basic UDI-DI 506047191ASPF; 20 UDI-DIs; Class IIa; EMDN A070103 INFUSION TUBING ADAPTERS AND CONNECTORS |
| Notified body | NB 2862 — certificate 28620126993 |
| PSUR reference number / version | [manufacturer] |
| Data collection period | 2024-10-07 to 2026-10-08 |

## Post-Market Surveillance: information including general PMCF information

### b) Scientific literature review of relevant specialist or technical literature

A literature search was performed on 2026-10-08 in PubMed/MEDLINE, Europe PMC and Crossref for records first available between 2024-10-07 and 2026-10-08, using a structured search for the device type, its synonyms and subject headings (full search strings in the search documentation). 359 records were screened on title and abstract; 38 met the inclusion criteria (the device type is a subject of the publication and safety, performance or clinical-use findings are reported), of which 17 report safety-related findings. Publications concerning the device(s) covered by this PSUR: 0.

| # | Publication | Finding type |
|---|---|---|
| 1 | Devrim İ et al. (2026). Clinical and Economic Impact of 70% Isopropyl Alcohol-Impregnated Disinfecting Caps on CLABSI Rates in Ambulatory Pediatric Hematology-Oncology Patients. Journal of pediatric hematology/oncology. | Safety |
| 2 | Çine N et al. (2026). The effect of three-way stopcock and closed vascular access systems on central line-associated bloodstream infection in the pediatric cardiac surgery intensive care unit. The journal of vascular access. | Safety |
| 3 | Marsubrin PMT et al. (2026). The efficacy of needleless connector designs (split septum vs mechanical valve) on preventing central line-associated bloodstream infection (CLABSI) in very preterm or very low-birth-weight neonates: a randomized clinical trial. Infection control and hospital epidemiology. | Safety |
| 4 | Barillo AR et al. (2026). Needleless dispensing pin with 1-way valve does not decrease bacterial contamination of communal-use intravenous fluid bags in simulated clinical scenario. Journal of the American Veterinary Medical Association. | Safety |
| 5 | Moureau N et al. (2026). Needleless connectors: A small device with big consequences. Home healthcare now. | Safety |
| 6 | Serrano Díaz L et al. (2026). The Safety and Efficacy of Disinfection Cap Use in a Real-World Environment. Medical devices (Auckland, N.Z.). | Safety |

*Table continues: 32 further publications.*

Manufacturer's assessment against the risk management file and clinical evaluation: [to be completed by the manufacturer].

### c) Public databases and/or registry data

| Database reviewed | Type | Period | Similar devices: findings |
|---|---|---|---|
| FDA MAUDE (USA) | Adverse event reports | 2024-10-07 – 2026-10-08 | 3,494 reports (death 1, injury 92, malfunction 3,397); product code FPA; brands of this device type only |
| FDA Medical Device Recalls (USA) | Field safety corrective actions | 5 years to 2026-10-08 | 1 recall event, 1 in the data collection period |
| Health Canada Recalls and Safety Alerts | Field safety corrective actions | 5 years | 11 recalls |
| MHRA Field Safety Notices (UK) | Field safety notices | 5 years | 7 notices |
| EUDAMED (EU) | Device registration | at 2026-10-08 | 425 UDI-DIs registered under EMDN A070103 |

| Most frequent device problems, similar devices (FDA MAUDE) | Reports |
|---|---|
| Leak/Splash | 775 |
| Fluid/Blood Leak | 648 |
| Complete Blockage | 378 |
| No Flow | 304 |
| Crack | 220 |

| Most frequent health effects, similar devices (FDA MAUDE) | Reports |
|---|---|
| Hemorrhage/Blood Loss/Bleeding | 81 |
| Chemical Exposure | 19 |
| Pain | 14 |
| Exposure to Body Fluids | 12 |

45% of the MAUDE reports for similar devices in the period were submitted by one manufacturer (Sistemas Medicos Alaris); year-on-year changes mainly reflect that manufacturer's reporting. Reports concerning the device(s) covered by this PSUR are excluded from these figures.

| Date | Firm / authority | Field action (excerpt) |
|---|---|---|
| 2025-12-19 | ICU Medical, Inc. | Needle free access device intended as accessory to vascular access used in hemodialysis or accessory to intravascular administration sets has a silicone seal … |
| 2026-10-01 | Health Canada | Medline Convenience Kits Containing Cannula Components with Female Luer Lock Connectors |
| 2026-06-19 | Health Canada | BD MaxPlus ™ Connector Minibore Extension Set |
| 2026-01-22 | ICU Medical | Tego |
| 2023-05-16 | B Braun | PROSET INFUSION SETs |

Registries: no device registry with public device-level data was identified for this device type.

Manufacturer's assessment against the risk management file and clinical evaluation: [to be completed by the manufacturer].

## d) Publicly available information about similar medical devices

Similar devices are those registered under EMDN A070103 (INFUSION TUBING ADAPTERS AND CONNECTORS) and, for third-country data, FDA product code FPA. In EUDAMED, 39 other manufacturers were identified in a sample of 298 registrations. In the USA, 2 new similar devices received 510(k) clearance in the period; the 510(k) summaries describe indications for use and performance testing.

| Decision | 510(k) | Applicant | Device |
|---|---|---|---|
| 2026-02-12 | K251375 | Puracath Medical, Inc. | PuraCath Firefly Needleless Connector IT (9005) |
| 2025-06-05 | K250616 | Icu Medical, Inc. | Clave™ Neutral-Displacement Needlefree Connectors |

| Manufacturer of similar devices (EU market) | Country |
|---|---|
| ERASER MEDIKAL TIBBİ CIH. SAG. URUN PAZ. SAN. VE TIC. LTD. STI. | Türkiye |
| Epic International (Thailand) Co., Ltd. | Thailand |
| Cair LGL | France |
| BQ Plus Medical Co., Ltd. | China |

Summary of safety and clinical performance (SS(C)P): not applicable — the device is Class IIa and non-implantable (MDR Art. 32).

Manufacturer's assessment against the risk management file and clinical evaluation: [to be completed by the manufacturer].