Periodic Safety Update Report

Ambulatory Blood Pressure Monitor — extract: PMS information (b) to (d), MDCG 2022-21 Annex I

Cover page item	Entry
Manufacturer	PAR Medizintechnik GmbH & Co. KG (SRN DE-MF-000017480)
Medical device(s) covered	Ambulatory Blood Pressure Monitor — Basic UDI-DI 426067137ABPMonitor3C; 1 UDI-DI; Class IIa; EMDN Z1203020501 NON–INVASIVE OSCILLOMETRIC BLOOD PRESSURE GAUGES
Notified body	NB 0482 — certificate 2210GB448240301
PSUR reference number / version	[manufacturer]
Data collection period	2024-10-07 to 2026-10-08

Post-Market Surveillance: information including general PMCF information

b) Scientific literature review of relevant specialist or technical literature

A literature search was performed on 2026-10-08 in PubMed/MEDLINE, Europe PMC and Crossref for records first available between 2024-10-07 and 2026-10-08, using a structured search for the device type, its synonyms and subject headings (full search strings in the search documentation). 3364 records were screened on title and abstract; 217 met the inclusion criteria (the device type is a subject of the publication and safety, performance or clinical-use findings are reported), of which 12 report safety-related findings. Publications concerning the device(s) covered by this PSUR: 0.

#	Publication	Finding type
1	McCarthy A et al. (2026). Postoperative Noninvasive Blood Pressure Monitoring for Patients With Osteogenesis Imperfecta Is Safe and Feasible. The Journal of the American Academy of Orthopaedic Surgeons.	Safety
2	Sriramka B et al. (2026). Rumpel-Leede phenomenon: A rare consequence of repeated non-invasive blood pressure monitoring. The National medical journal of India.	Safety
3	Florio T et al. (2026). Reduction of Wrong-Limb Blood Pressure Cuff Placement Attempts With a High-Visibility Limb Alert Sleeve: A Randomized Crossover Simulation Study. Cureus.	Safety
4	Mikkelsen LF et al. (2026). Minimal Perceived Discomfort During Oscillometric Blood Pressure Measurement in Children: A Cross-Sectional Study With Implications for Systematic Screening. Journal of clinical hypertension (Greenwich, Conn.).	Safety
5	Torino C et al. (2026). Feasibility of the extended, 48-h ambulatory blood pressure monitoring in hemodialysis patients: a multicenter study. Journal of hypertension.	Safety
6	Rahman S et al. (2026). Case Report: Blood Pressure Cuff-Associated Compartment Syndrome of the Right Upper Extremity in a Patient With Multiple Comorbidities. Journal of the American College of Emergency Physicians open.	Safety

Table continues: 211 further publications.

Manufacturer's assessment against the risk management file and clinical evaluation: [to be completed by the manufacturer].

c) Public databases and/or registry data

Database reviewed	Type	Period	Similar devices: findings
FDA MAUDE (USA)	Adverse event reports	2024-10-07 – 2026-10-08	1,337 reports (death 1, injury 40, malfunction 1,291); product code DXN
FDA Medical Device Recalls (USA)	Field safety corrective actions	5 years to 2026-10-08	3 recall events, 1 in the data collection period
Health Canada Recalls and Safety Alerts	Field safety corrective actions	5 years	2 recalls
MHRA Field Safety Notices (UK)	Field safety notices	5 years	0 notices
EUDAMED (EU)	Device registration	at 2026-10-08	1,276 UDI-DIs registered under EMDN Z1203020501

Most frequent device problems, similar devices (FDA MAUDE)	Reports
Incorrect, Inadequate or Imprecise Result or Readings	284
High Readings	159
Unable to Obtain Readings	130
Inflation Problem	126
Failure to Power Up	109

Most frequent health effects, similar devices (FDA MAUDE)	Reports
Ecchymosis	49
Hemorrhage/Blood Loss/Bleeding	14
High Blood Pressure/ Hypertension	11
Pain	8

68% of the MAUDE reports for similar devices in the period were submitted by one manufacturer (Medline Industries LP); year-on-year changes mainly reflect that manufacturer's reporting. Reports concerning the device(s) covered by this PSUR are excluded from these figures.

Date	Firm / authority	Field action (excerpt)
2025-02-24	PYRAMES INC	Single-use battery-powered blood pressure monitor, for use on neonates and infants, due to vibratory noise, may not adequately detect changes in blood pressure …
2024-08-30	MEDLINE INDUSTRIES, LP - Northfield	It has been determined that certain Blood Pressure Monitors containing a faulty microchip have failed to power on or have shown display screen issues after the …
2024-03-26	Baxter Healthcare Corporation	Baxter received reports of an issue related to the construction of the power cord not meeting the insulation rating per country-specific requirements and …
2025-04-11	Health Canada	Baxter Corporation Reusable Blood Pressure Cuffs
2024-05-23	Health Canada	BP-200 Plus Blood Pressure Monitor

Registries: no device registry with public device-level data was identified for this device type.

Manufacturer's assessment against the risk management file and clinical evaluation: [to be completed by the manufacturer].

d) Publicly available information about similar medical devices

Similar devices are those registered under EMDN Z1203020501 (NON–INVASIVE OSCILLOMETRIC BLOOD PRESSURE GAUGES) and, for third-country data, FDA product code DXN. In EUDAMED, 22 other manufacturers were identified in a sample of 280 registrations. In the USA, 60 new similar devices received 510(k) clearance in the period; the 510(k) summaries describe indications for use and performance testing.

Decision	510(k)	Applicant	Device
2026-09-10	K261929	OMRON Healthcare, Inc.	Upper Arm Blood pressure monitor (BP7155D)
2026-09-03	K254000	Happy Health, Inc.	Happy Health Blood Pressure Monitor
2026-08-31	K262658	Honsun (Nantong) Co., Ltd.	Automatic Digital Blood Pressure monitor (LD-526)
2026-08-26	K261394	JOYTECH Healthcare Co., Ltd.	Arm-type Fully Automatic Digital Blood Pressure Monitor (DBP-6186 Plus, DBP-6286B Plus, DBP-61E3 Plus, DBP-61E

Manufacturer of similar devices (EU market)	Country
JOYTECH Healthcare Co., Ltd.	China
Shenzhen AOJ Medical Technology Co., Ltd.	China
Shenzhen Urion Technology Co., Ltd.	China
Microlife Corporation	Taiwan

Summary of safety and clinical performance (SS(C)P): not applicable — the device is Class IIa and non-implantable (MDR Art. 32).

Manufacturer's assessment against the risk management file and clinical evaluation: [to be completed by the manufacturer].