Periodic Safety Update Report

AlphaScope — extract: PMS information (b) to (d), MDCG 2022-21 Annex I

Cover page item	Entry
Manufacturer	SCHÖLLY FIBEROPTIC GMBH (SRN DE-MF-000005618)
Medical device(s) covered	AlphaScope — Basic UDI-DI 42504801LAPAROSCOPESEF; 51 UDI-DIs; Class IIa; EMDN Z12029009 LAPAROSCOPES
Notified body	NB 0297 — certificate 1000285837
PSUR reference number / version	[manufacturer]
Data collection period	2024-10-07 to 2026-10-08

Post-Market Surveillance: information including general PMCF information

b) Scientific literature review of relevant specialist or technical literature

A literature search was performed on 2026-10-08 in PubMed/MEDLINE, Europe PMC and Crossref for records first available between 2024-10-07 and 2026-10-08, using a structured search for the device type, its synonyms and subject headings (full search strings in the search documentation). 1066 records were screened on title and abstract; 238 met the inclusion criteria (the device type is a subject of the publication and safety, performance or clinical-use findings are reported), of which 8 report safety-related findings. Publications concerning the device(s) covered by this PSUR: 0.

#	Publication	Finding type
1	Neel NC et al. (2026). Indocyanine Green Fluorescent Cholangiography During Laparoscopic Cholecystectomy and Bile Duct Injury. JAMA surgery.	Safety
2	Li J et al. (2025). Comparative efficacy of decontamination methods for laparoscopic equipment: a systematic review and meta-analysis. Frontiers in cellular and infection microbiology.	Safety
3	Noordmans HJ et al. (2025). Multi-Center Evaluation of Optical Degradation Patterns in Rigid Endoscopes: A Data-Driven Approach. Medical devices (Auckland, N.Z.).	Safety
4	Sun Y et al. (2025). Focusing on chylous ascites: a noteworthy complication after laparoscopic/robotic para-aortic lymphadenectomy in left-sided colorectal cancer in a high-volume single center. Techniques in coloproctology.	Safety
5	Xia S et al. (2024). The Application Effect of Fine Management Combined with Man-Machine Fixation Mode in Reducing the Attrition Rate of Laparoscopic Instruments-A Non-Randomized, Concurrent Controlled Study. Medical devices (Auckland, N.Z.).	Safety
6	Mishima K et al. (2026). Does ICG-NIRF cholangiography replace the critical view of safety in laparoscopic cholecystectomy? A prospective evaluation on biliary visualization dynamics and the impact of inflammation. Surgical endoscopy.	Performance

Table continues: 232 further publications.

Manufacturer's assessment against the risk management file and clinical evaluation: [to be completed by the manufacturer].

c) Public databases and/or registry data

Database reviewed	Type	Period	Similar devices: findings
FDA MAUDE (USA)	Adverse event reports	2024-10-07 – 2026-10-08	5,528 reports (death 0, injury 28, malfunction 5,500); product code GCJ; brands of this device type only
FDA Medical Device Recalls (USA)	Field safety corrective actions	5 years to 2026-10-08	1 recall event, 0 in the data collection period
Health Canada Recalls and Safety Alerts	Field safety corrective actions	5 years	0 recalls
MHRA Field Safety Notices (UK)	Field safety notices	5 years	3 notices
EUDAMED (EU)	Device registration	at 2026-10-08	462 UDI-DIs registered under EMDN Z12029009

Most frequent device problems, similar devices (FDA MAUDE)	Reports
Poor Quality Image	1,140
No Display/Image	995
Erratic or Intermittent Display	665
Communication or Transmission Problem	640
Loss of or Failure to Bond	633

Most frequent health effects, similar devices (FDA MAUDE)	Reports
Laceration(s)	5
Burn(s)	4
Bowel Burn	3
Hemorrhage/Blood Loss/Bleeding	2

77% of the MAUDE reports for similar devices in the period were submitted by one manufacturer (Aizu Olympus); year-on-year changes mainly reflect that manufacturer's reporting. Reports concerning the device(s) covered by this PSUR are excluded from these figures.

Date	Firm / authority	Field action (excerpt)
2022-01-14	Stryker Corporation	A software defect in the camera control unit (CCU) will cause the image on the monitor to flip upside-down into an incorrect orientation. The potential of …
2024-06-21	Olympus	HD Lightweight Pendulum Camera Head
2023-11-22	Olympus	Video telescope “ENDOEYE HD II”, 10 mm, 0° and 30°

Registries: no device registry with public device-level data was identified for this device type.

Manufacturer's assessment against the risk management file and clinical evaluation: [to be completed by the manufacturer].

d) Publicly available information about similar medical devices

Similar devices are those registered under EMDN Z12029009 (LAPAROSCOPES) and, for third-country data, FDA product code GCJ. In EUDAMED, 18 other manufacturers were identified in a sample of 297 registrations. In the USA, 10 new similar devices received 510(k) clearance in the period; the 510(k) summaries describe indications for use and performance testing.

Decision	510(k)	Applicant	Device
2026-06-04	K253994	TroCare, LLC	TroKit Laparoscope Lens Wiper
2026-04-30	K252858	Kms Medical Technology Co., Ltd.	4K UHD Endoscopy Camera System (KMS-4K-2088, KMS-4K-2188, KMS-4K-2288)
2026-04-08	K260089	Scopix , Ltd.	TL-10 Laparoscopic Scope Cleaner
2026-02-18	K260185	Stryker Endoscopy	1688 4K Camera System with Advanced Imaging Modality; 1788 4K Camera System with Advanced Imaging Modality

Manufacturer of similar devices (EU market)	Country
Gimmi GmbH	Germany
RZ-Medizintechnik GmbH	Germany
Richard Wolf GmbH	Germany
Nanjing Mindray Bio-Medical Electronics Co., Ltd.	China

Summary of safety and clinical performance (SS(C)P): not applicable — the device is Class IIa and non-implantable (MDR Art. 32).

Manufacturer's assessment against the risk management file and clinical evaluation: [to be completed by the manufacturer].