Periodic Safety Update Report

Symphony Breast Pump — extract: PMS information (b) to (d), MDCG 2022-21 Annex I

Cover page item	Entry
Manufacturer	Medela AG (SRN CH-MF-000018913)
Medical device(s) covered	Symphony Breast Pump — Basic UDI-DI 07612367072380LR; 13 UDI-DIs; Class IIa; EMDN Z12080303 BREAST PUMPS
Notified body	NB 0123 — certificate G10 011634 0235
PSUR reference number / version	[manufacturer]
Data collection period	2024-10-07 to 2026-10-08

Post-Market Surveillance: information including general PMCF information

b) Scientific literature review of relevant specialist or technical literature

A literature search was performed on 2026-10-08 in PubMed/MEDLINE, Europe PMC and Crossref for records first available between 2024-10-07 and 2026-10-08, using a structured search for the device type, its synonyms and subject headings (full search strings in the search documentation). 435 records were screened on title and abstract; 23 met the inclusion criteria (the device type is a subject of the publication and safety, performance or clinical-use findings are reported), of which 5 report safety-related findings. Publications concerning the device(s) covered by this PSUR: 0.

#	Publication	Finding type
1	Theurich MA et al. (2025). Perspectives on Breast Pump Experiences: Findings from a U.S. National Cross-Sectional Survey. Breastfeeding medicine.	Safety
2	Fuente-Garrido R et al. (2025). Raynaud's Phenomenon of the Nipple Triggered by Breast-Pump. Breastfeeding medicine.	Safety
3	Reyes SM et al. (2025). Pumping and hygiene practices are associated with bacterial load and microbial composition in human milk expressed at home. Journal of translational medicine.	Safety
4	Ginty O et al. (2025). Case Report of an Unexpected Cause of Noisy Breathing: Breast Pump Membrane as an Esophageal Foreign Body in an Infant. Breastfeeding medicine.	Safety
5	Gridneva Z et al. (2026). Exclusively Pumping Women's Milk Production and Pumping and Feeding Dynamics-An Online Survey. Healthcare (Basel, Switzerland).	Performance
6	Warden AH et al. (2026). Comparison of Milk Removal Efficiency and Maternal Perception Between a Hospital-Grade and Personal-Use Pump. Healthcare (Basel, Switzerland).	Performance

Table continues: 17 further publications.

Manufacturer's assessment against the risk management file and clinical evaluation: [to be completed by the manufacturer].

c) Public databases and/or registry data

Database reviewed	Type	Period	Similar devices: findings
FDA MAUDE (USA)	Adverse event reports	2024-10-07 – 2026-10-08	52 reports (death 0, injury 41, malfunction 11); product code HGX
FDA Medical Device Recalls (USA)	Field safety corrective actions	5 years to 2026-10-08	0 recall events, 0 in the data collection period
Health Canada Recalls and Safety Alerts	Field safety corrective actions	5 years	1 recall
MHRA Field Safety Notices (UK)	Field safety notices	5 years	0 notices
EUDAMED (EU)	Device registration	at 2026-10-08	134 UDI-DIs registered under EMDN Z12080303

Most frequent device problems, similar devices (FDA MAUDE)	Reports
Patient Device Interaction Problem	29
Decrease in Suction	3
Break	2
Defective Device	2
Device Contamination with Body Fluid	2

Most frequent health effects, similar devices (FDA MAUDE)	Reports
Breast Discomfort/Pain	32
Pain	5
Breast Feeding Problem	4
Swelling/ Edema	3

58% of the MAUDE reports for similar devices in the period were submitted by one manufacturer (Chiaro Technology); year-on-year changes mainly reflect that manufacturer's reporting. Reports concerning the device(s) covered by this PSUR are excluded from these figures.

Date	Firm / authority	Field action (excerpt)
2024-02-27	Health Canada	Wearable Breast Pumps

Registries: no device registry with public device-level data was identified for this device type.

Manufacturer's assessment against the risk management file and clinical evaluation: [to be completed by the manufacturer].

d) Publicly available information about similar medical devices

Similar devices are those registered under EMDN Z12080303 (BREAST PUMPS) and, for third-country data, FDA product code HGX. In EUDAMED, 16 other manufacturers were identified in a sample of 134 registrations. In the USA, 59 new similar devices received 510(k) clearance in the period; the 510(k) summaries describe indications for use and performance testing.

Decision	510(k)	Applicant	Device
2026-09-17	K254203	Shenzhen TPH Technology Co., Ltd.	Wearable Breast Pump (Model S36)
2026-09-03	K261622	Philips Consumer Lifestyle B.V.	Philips Avent Natural Care Hands-free Wearable Electric Breast Pump (Model: SCF593/52, SCF593/62); Philips Ave
2026-09-01	K262774	Guangdong Horigen Mother & Baby Products Co., Ltd.	Wearable breast pump (Model: XN/MD-2249BS, ZFHF)
2026-08-21	K260643	Cimilre Co., Ltd.	Cimilre Dual Max

Manufacturer of similar devices (EU market)	Country
Uzinmedicare	South Korea
Ardo medical AG	Switzerland
AOJ HEALTH Technology Co., Ltd.	China
Guangdong Horigen Mother & Baby Products Co., Ltd.	China

Summary of safety and clinical performance (SS(C)P): not applicable — the device is Class IIa and non-implantable (MDR Art. 32).

Manufacturer's assessment against the risk management file and clinical evaluation: [to be completed by the manufacturer].