Periodic Safety Update Report

Ambulatory Blood Pressure Monitor — extract: PMS information (b) to (d), MDCG 2022-21 Annex I

Cover page item	Entry
Manufacturer	PAR Medizintechnik GmbH & Co. KG (SRN DE-MF-000017480)
Medical device(s) covered	Ambulatory Blood Pressure Monitor — Basic UDI-DI 426067137ABPMonitor3C; 1 UDI-DI; Class IIa; EMDN Z1203020501 NON–INVASIVE OSCILLOMETRIC BLOOD PRESSURE GAUGES
Notified body	NB 0482 — certificate 2210GB448240301
PSUR reference number / version	[manufacturer]
Data collection period	2024-10-07 to 2026-10-08

Post-Market Surveillance: information including general PMCF information

b) Scientific literature review of relevant specialist or technical literature

A literature search was performed on 2026-10-08 in PubMed/MEDLINE, Europe PMC and Crossref for records first available between 2024-10-07 and 2026-10-08, using a structured search for the device type, its synonyms and subject headings (full search strings in the search documentation). 3,364 records were screened on title and abstract; 217 met the inclusion criteria (the device type is a subject of the publication and safety, performance or clinical-use findings are reported), of which 12 report safety-related findings. Publications concerning the device(s) covered by this PSUR: 0.

#	Publication	Finding type
1	McCarthy A et al. (2026). Postoperative Noninvasive Blood Pressure Monitoring for Patients With Osteogenesis Imperfecta Is Safe and Feasible. The Journal of the American Academy of Orthopaedic Surgeons.	Safety
2	Sriramka B et al. (2026). Rumpel-Leede phenomenon: A rare consequence of repeated non-invasive blood pressure monitoring. The National medical journal of India.	Safety
3	Florio T et al. (2026). Reduction of Wrong-Limb Blood Pressure Cuff Placement Attempts With a High-Visibility Limb Alert Sleeve: A Randomized Crossover Simulation Study. Cureus.	Safety
4	Mikkelsen LF et al. (2026). Minimal Perceived Discomfort During Oscillometric Blood Pressure Measurement in Children: A Cross-Sectional Study With Implications for Systematic Screening. Journal of clinical hypertension (Greenwich, Conn.).	Safety
5	Torino C et al. (2026). Feasibility of the extended, 48-h ambulatory blood pressure monitoring in hemodialysis patients: a multicenter study. Journal of hypertension.	Safety

Table continues: 212 further publications.

Manufacturer's assessment against the risk management file and clinical evaluation: [to be completed by the manufacturer].

c) Public databases and/or registry data

Database reviewed	Type	Period	Similar devices: findings
FDA MAUDE (USA)	Adverse event reports	2024-10-07 – 2026-10-08	5,057 reports (death 1, injury 45, malfunction 4,989, other 22); product code DXN
FDA Medical Device Recalls (USA)	Field safety corrective actions	5 years to 2026-10-08	3 recall events, 1 in the data collection period
Health Canada Recalls and Safety Alerts	Field safety corrective actions	5 years	2 recalls
MHRA Field Safety Notices (UK)	Field safety notices	5 years	0 notices
EUDAMED (EU)	Device registration	at 2026-10-13	1,280 UDI-DIs registered under EMDN Z1203020501

Most frequent device problems, similar devices (FDA MAUDE)	Reports
Computer Software Problem	1,539
Incorrect, Inadequate or Imprecise Result or Readings	1,017
High Readings	765
Unable to Obtain Readings	548

Most frequent health effects, similar devices (FDA MAUDE)	Reports
Ecchymosis	49
High Blood Pressure/ Hypertension	31
Discomfort	29
Pain	25

90% of the MAUDE reports for similar devices in the period were submitted by one manufacturer (Medline Industries LP); year-on-year changes mainly reflect that manufacturer's reporting. Reports concerning the device(s) covered by this PSUR are excluded from these figures.

Date	Firm or brand (source)	Field action (excerpt)
2025-02-24	PYRAMES INC	Single-use battery-powered blood pressure monitor, for use on neonates and infants, due to vibratory noise, may not adequately detect changes in blood pressure …
2024-08-30	MEDLINE INDUSTRIES, LP - Northfield	It has been determined that certain Blood Pressure Monitors containing a faulty microchip have failed to power on or have shown display screen issues after the …
2025-03-20	Welch Allyn, Inc. (Health Canada)	Baxter Corporation is issuing a voluntary product recall for reusable blood pressure cuffs because the product is labeled "not made with natural rubber latex," …

Figures as retrieved from FDA and EUDAMED on 2026-10-13, from Health Canada and MHRA on 2026-10-04. These databases continue to add records for past dates, so later queries can return more.

Registries: no device registry with public device-level data was identified for this device type.

Manufacturer's assessment against the risk management file and clinical evaluation: [to be completed by the manufacturer].

d) Publicly available information about similar medical devices

Similar devices are those registered under EMDN Z1203020501 (NON–INVASIVE OSCILLOMETRIC BLOOD PRESSURE GAUGES) and, for third-country data, FDA product code DXN. In EUDAMED, 22 other manufacturers were identified in a sample of 280 registrations. In the USA, 61 new similar devices received 510(k) clearance in the period; the 510(k) summaries describe indications for use and performance testing.

Decision	510(k)	Applicant	Device
2026-10-01	K260470	Shenzhen Brav Electronic Technologies Co., Ltd.	Electronic Blood Pressure Monitor (BP-204C, BP-204A, BP-215, BP-203C, BP-205A, …
2026-09-10	K261929	OMRON Healthcare, Inc.	Upper Arm Blood pressure monitor (BP7155D)
2026-09-03	K254000	Happy Health, Inc.	Happy Health Blood Pressure Monitor

Manufacturer of similar devices (EU market)	Country
JOYTECH Healthcare Co., Ltd.	China
Shenzhen AOJ Medical Technology Co., Ltd.	China
Shenzhen Urion Technology Co., Ltd.	China

Summary of safety and clinical performance (SS(C)P): not applicable — the device is Class IIa and non-implantable (MDR Art. 32).

Manufacturer's assessment against the risk management file and clinical evaluation: [to be completed by the manufacturer].