Periodic Safety Update Report

ImagingRing m — extract: PMS information (b) to (d), MDCG 2022-21 Annex I

Cover page item	Entry
Manufacturer	medPhoton GmbH (SRN AT-MF-000002648)
Medical device(s) covered	ImagingRing m — Basic UDI-DI 1912007532IRm33; 4 UDI-DIs; Class IIb; EMDN Z11039009 FLUOROSCOPY DEVICES
Notified body	NB 0483 — certificate D4005200012
PSUR reference number / version	[manufacturer]
Data collection period	2025-10-08 to 2026-10-08

Post-Market Surveillance: information including general PMCF information

b) Scientific literature review of relevant specialist or technical literature

A literature search was performed on 2026-10-08 in PubMed/MEDLINE, Europe PMC and Crossref for records first available between 2025-10-08 and 2026-10-08, using a structured search for the device type, its synonyms and subject headings (full search strings in the search documentation). 2200 records were screened on title and abstract; 415 met the inclusion criteria (the device type is a subject of the publication and safety, performance or clinical-use findings are reported), of which 107 report safety-related findings. Publications concerning the device(s) covered by this PSUR: 0.

#	Publication	Finding type
1	Rettig N et al. (2026). Query-based prediction of 3D x-ray scatter fields in an interventional suite using a transformer model. Journal of radiological protection.	Safety
2	Schneider C et al. (2026). Angio-CT does not increase patient radiation exposure compared to conventional cone-beam CT for intra-arterial liver procedures. European radiology.	Safety
3	Sari S et al. (2026). Radiation Exposure in Elective Endovascular Carotid Artery Stenting. Clinical neuroradiology.	Safety
4	Hyvärinen T et al. (2026). Tight collimation substantially increases air kerma rate and decreases image contrast during selected fluoroscopy procedures. Physica medica.	Safety
5	Ishibashi T et al. (2026). Effects of the Institutional Volume of Endovascular Treatment for Lower Extremity Artery Disease on Radiation Exposure. Journal of endovascular therapy.	Safety

Table continues: 410 further publications.

Manufacturer's assessment against the risk management file and clinical evaluation: [to be completed by the manufacturer].

c) Public databases and/or registry data

Database reviewed	Type	Period	Similar devices: findings
FDA MAUDE (USA)	Adverse event reports	2025-10-08 – 2026-10-08	674 reports (death 1, injury 3, malfunction 670); product code OWB; brands of this device type only
FDA Medical Device Recalls (USA)	Field safety corrective actions	5 years to 2026-10-08	3 recall events, 3 in the data collection period
Health Canada Recalls and Safety Alerts	Field safety corrective actions	5 years	3 recalls
MHRA Field Safety Notices (UK)	Field safety notices	5 years	37 notices
EUDAMED (EU)	Device registration	at 2026-10-13	60 UDI-DIs registered under EMDN Z11039009

Most frequent device problems, similar devices (FDA MAUDE)	Reports
Failure to Advance	320
Application Program Freezes or Fails to Launch	93
Incorrect, Inadequate or Imprecise Result or Readings	92
Mechanical Problem	91
Image Display Error/Artifact	70

Most frequent health effects, similar devices (FDA MAUDE)	Reports
Spinal Column Injury	2
Bone Fracture(s)	1
Pain	1
Perforation	1

98% of the MAUDE reports for similar devices in the period were submitted by one manufacturer (Medtronic Navigation Boxborough); year-on-year changes mainly reflect that manufacturer's reporting. Reports concerning the device(s) covered by this PSUR are excluded from these figures.

Date	Firm (source)	Field action (excerpt)
2026-08-04	Canon Medical System, USA, INC.	Due to the fixing screws of ceiling movement gear in ceiling suspended arm that are part of the X-ray interventional systems may become loose.
2026-05-19	Medtronic Navigation, Inc.-Boxborough	Potential for image artifacts caused by an anomaly in the O-arm O2 Imaging System s detector panel firmware.
2025-12-19	Canon Medical System, USA, INC.	It has been discovered that there is a possibility that the fixing screws of the ceiling movement gear in the ceiling suspended arm that are part of the X-ray …
2026-09-09	PHILIPS MEDICAL SYSTEMS NEDERLAND B.V. (Health Canada)	Air bubbles may develop in the monoblock oil when there is a prolonged period between manufacture and installation, for example due to storage, transportation, …
2024-12-18	GE OEC Medical Systems Inc. (Health Canada)	GE Healthcare has become aware that recently manufactured OEC Elite and OEC Mobile C-arms, as well as recently replaced x-ray tubes on certain OEC Elite, OEC …
2026-10-31	Philips (MHRA)	Field safety notice: Zenition 10, Zenition 30 (Interventional Fluoroscopic X-Ray System)
2026-10-20	Philips (MHRA)	Field safety notice: Allura Xper - UPDATE (Interventional Fluoroscopic X-Ray System)

Figures as retrieved from FDA and EUDAMED on 2026-10-13, from Health Canada and MHRA on 2026-10-04. These databases continue to add records for past dates, so later queries can return more.

Registries: no device registry with public device-level data was identified for this device type.

Manufacturer's assessment against the risk management file and clinical evaluation: [to be completed by the manufacturer].

d) Publicly available information about similar medical devices

Similar devices are those registered under EMDN Z11039009 (FLUOROSCOPY DEVICES) and, for third-country data, FDA product code OWB. In EUDAMED, 14 other manufacturers were identified in a sample of 52 registrations. In the USA, 1 new similar device received 510(k) clearance in the period; the 510(k) summaries describe indications for use and performance testing.

Decision	510(k)	Applicant	Device
2025-11-04	K250241	Siemens Medical Solutions USA, Inc.	Cios Select

Manufacturer of similar devices (EU market)	Country
GENORAY CO., Ltd.	South Korea
Siemens Healthcare GmbH	Germany
Philips Medical Systems DMC GmbH	Germany
Beijing Wandong Medical Technology Co., Ltd.	China

Summary of safety and clinical performance (SS(C)P): not applicable — the device is Class IIb and non-implantable (MDR Art. 32).

Manufacturer's assessment against the risk management file and clinical evaluation: [to be completed by the manufacturer].