Periodic Safety Update Report

Bittium MINI UL — extract: PMS information (b) to (d), MDCG 2022-21 Annex I

Cover page item	Entry
Manufacturer	Bittium Biosignals Oy (SRN FI-MF-000013983)
Medical device(s) covered	Bittium MINI UL — Basic UDI-DI 643003522Cardiac3Z; 2 UDI-DIs; Class IIa; EMDN Z12050403 ECG HOLTER RECORDERS
Notified body	NB 0537 — certificate CR-03-1030-811-22
PSUR reference number / version	[manufacturer]
Data collection period	2024-10-07 to 2026-10-08

Post-Market Surveillance: information including general PMCF information

b) Scientific literature review of relevant specialist or technical literature

A literature search was performed on 2026-10-08 in PubMed/MEDLINE, Europe PMC and Crossref for records first available between 2024-10-07 and 2026-10-08, using a structured search for the device type, its synonyms and subject headings (full search strings in the search documentation). 1,785 records were screened on title and abstract; 173 met the inclusion criteria (the device type is a subject of the publication and safety, performance or clinical-use findings are reported), of which one reports safety-related findings. Publications concerning the device(s) covered by this PSUR: 0.

#	Publication	Finding type
1	Jo NY et al. (2025). Prophylactic Intranasal Corticosteroids to Prevent Contact Dermatitis From Ambulatory ECG Monitor Adhesives: A Novel Strategy. Cureus.	Safety
2	Yamamoto H et al. (2026). Normal electrocardiography at presentation masks SCN5A-related long QT syndrome: the importance of Holter monitoring in children with a family history of Brugada syndrome. Cardiology in the young.	Performance
3	Fujino T et al. (2026). One-week patch ECG reveals residual atrial fibrillation missed by repeated Holter monitoring after apparently successful catheter ablation. Journal of electrocardiology.	Performance
4	Russo P et al. (2026). Clinical and economic outcomes in diagnostic-naïve Medicare fee-for-service beneficiaries managed with alert-based ambulatory cardiac monitors: The CAMELOT-ALERT study. American heart journal.	Performance
5	Pavluk D et al. (2026). The weakest link in digital atrial fibrillation screening: a randomized comparison of self-applied ECG patch systems in older adults. Europace.	Performance
6	Reed MJ et al. (2026). Immediate Ambulatory Electrocardiographic Monitoring in Syncope. The New England journal of medicine.	Performance

Table continues: 167 further publications.

Manufacturer's assessment against the risk management file and clinical evaluation: [to be completed by the manufacturer].

c) Public databases and/or registry data

Database reviewed	Type	Period	Similar devices: findings
FDA MAUDE (USA)	Adverse event reports	2024-10-07 – 2026-10-08	18 reports (death 0, injury 6, malfunction 11, other 1); product code MWJ, MLO
FDA Medical Device Recalls (USA)	Field safety corrective actions	5 years to 2026-10-08	2 recall events, 2 in the data collection period
Health Canada Recalls and Safety Alerts	Field safety corrective actions	5 years	0 recalls
MHRA Field Safety Notices (UK)	Field safety notices	5 years	0 notices
EUDAMED (EU)	Device registration	at 2026-10-13	262 UDI-DIs registered under EMDN Z12050403

Most frequent device problems, similar devices (FDA MAUDE)	Reports
Patient Data Problem	3
Overheating of Device	2
Patient Device Interaction Problem	2
Patient-Device Incompatibility	2
Adverse Event Without Identified Device or Use Problem	1

Most frequent health effects, similar devices (FDA MAUDE)	Reports
Burning Sensation	4
Itching Sensation	4
Blister	2
Erythema	2

Date	Firm or brand (source)	Field action (excerpt)
2024-11-25	Spacelabs Healthcare, Ltd.	Due to a manufacturing error the incorrect device serial number was potentially programmed into the holter recorder. The physical serial number label on the …
2024-10-11	Baxter Healthcare Corporation	There is the potential for exam files being assigned duplicate Unique Identifiers (UIDs). If the system receiving the DICOM file (e.g., Picture Archiving and …

Figures as retrieved from FDA and EUDAMED on 2026-10-13, from Health Canada and MHRA on 2026-10-04. These databases continue to add records for past dates, so later queries can return more.

Registries: no device registry with public device-level data was identified for this device type.

Manufacturer's assessment against the risk management file and clinical evaluation: [to be completed by the manufacturer].

d) Publicly available information about similar medical devices

Similar devices are those registered under EMDN Z12050403 (ECG HOLTER RECORDERS) and, for third-country data, FDA product code MWJ, MLO. In EUDAMED, 41 other manufacturers were identified in a sample of 228 registrations. In the USA, 9 new similar devices received 510(k) clearance in the period; the 510(k) summaries describe indications for use and performance testing.

Decision	510(k)	Applicant	Device
2026-09-11	K254119	Shenzhen Xinyan Medical Technology Co.,ltd.	Wearable ECG Recorder (YMHeart XY-12); Wearable ECG Recorder (YMHeart XY-1201); …
2026-09-11	K261580	Zbeats, Inc.	ZBPro Diagnostic (Version 2.0)
2026-09-04	K253529	Peerbridge Health, Inc.	Peerbridge COR MDx (Cor MDx)
2026-08-14	K253562	Schiller AG	medilogFD

Manufacturer of similar devices (EU market)	Country
Qingdao Weheal Technology co., Ltd	China
Biomedical Instruments Co., Ltd.	China
Shenzhen Viatom Technology Co., Ltd.	China
Suzhou Beneware Medical Equipment Co.,Ltd.	China

Summary of safety and clinical performance (SS(C)P): not applicable — the device is Class IIa and non-implantable (MDR Art. 32).

Manufacturer's assessment against the risk management file and clinical evaluation: [to be completed by the manufacturer].