Filename: 2025-03-13 - 11.52.58.gDNA



Default image (Contrast 100%)

Sample Info

Well	DIN	Conc. [ng/µl]	Sample Description	Alert	Observations
L1	-	86.8	Ladder	⚠	Caution! Expired ScreenTape device; Ladder
B1	5.6	82.3	1	⚠	Caution! Expired ScreenTape device
C1	5.5	76.7	2	⚠	Caution! Expired ScreenTape device
D1	3.1	16.4	3	⚠	Caution! Expired ScreenTape device
E1	3.9	21.7	4	⚠	Caution! Expired ScreenTape device
F1	5.7	17.0	6	⚠	Caution! Expired ScreenTape device
L2	-	84.3	Ladder	⚠	Caution! Expired ScreenTape device; Ladder
G1	5.3	25.9	8	⚠	Caution! Expired ScreenTape device
H1	4.5	7.62	9	⚠	Caution! Expired ScreenTape device; Sample concentration outside recommended range
A2	6.6	13.4	10	⚠	Caution! Expired ScreenTape device
B2	4.8	8.96	11	⚠	Caution! Expired ScreenTape device; Sample concentration outside recommended range

© Agilent Technologies, Inc. 2023

L1: Ladder



Sample Table

Well	DIN	Conc. [ng/µl]	Sample Description	Alert	Observations
L1	-	86.8	Ladder	⚠	Caution! Expired ScreenTape device; Ladder

Peak Table

Size [bp]	Calibrated Conc. [ng/µl]	Assigned Conc. [ng/µl]	% Integrated Area	From [bp]	To [bp]	Peak Comment	Observations
100	8.50	8.50	-	69	149		Lower Marker
250	6.67	-	7.88	181	312		
400	6.98	-	8.25	337	484		
600	7.24	-	8.56	527	725		
900	7.57	-	8.95	791	1077		
1200	7.37	-	8.71	1077	1372		
1500	7.54	-	8.91	1372	1780		
2000	7.04	-	8.32	1822	2293		
2500	6.62	-	7.82	2293	2793		
3000	6.48	-	7.66	2793	3517		
4000	6.18	-	7.30	3625	5407		
7000	6.07	-	7.17	5564	10781		
15000	5.27	-	6.23	10781	23114		
48500	3.58	-	4.23	23114	73067		

© Agilent Technologies, Inc. 2023

B1: 1



Sample Table

Well	DIN	Conc. [ng/µl]	Sample Description	Alert	Observations
B1	5.6	82.3	1	⚠	Caution! Expired ScreenTape device

Peak Table

Size [bp]	Calibrated Conc. [ng/µl]	Assigned Conc. [ng/µl]	% Integrated Area	From [bp]	To [bp]	Peak Comment	Observations
100	8.50	8.50	-	66	148		Lower Marker
1571	3.89	-	10.97	1420	1686		
12277	31.2	-	88.07	4790	57023		
-	-	-	-	-	-		Sample Well

© Agilent Technologies, Inc. 2023

C1: 2



Sample Table

Well	DIN	Conc. [ng/µl]	Sample Description	Alert	Observations
C1	5.5	76.7	2	⚠	Caution! Expired ScreenTape device

Peak Table

Size [bp]	Calibrated Conc. [ng/µl]	Assigned Conc. [ng/µl]	% Integrated Area	From [bp]	To [bp]	Peak Comment	Observations
100	8.50	8.50	-	67	146		Lower Marker
1247	5.67	-	14.82	1201	1624		
9655	32.6	-	85.05	3971	29898		
-	-	-	-	-	-		Sample Well

© Agilent Technologies, Inc. 2023

D1: 3



Sample Table

Well	DIN	Conc. [ng/µl]	Sample Description	Alert	Observations
D1	3.1	16.4	3	⚠	Caution! Expired ScreenTape device

Peak Table

Size [bp]	Calibrated Conc. [ng/µl]	Assigned Conc. [ng/µl]	% Integrated Area	From [bp]	To [bp]	Peak Comment	Observations
100	8.50	8.50	-	66	148		Lower Marker
780	4.22	-	46.95	294	931		
16010	4.77	-	53.05	5299	>60000		

© Agilent Technologies, Inc. 2023

E1: 4



Sample Table

Well	DIN	Conc. [ng/µl]	Sample Description	Alert	Observations
E1	3.9	21.7	4	⚠	Caution! Expired ScreenTape device

Peak Table

Size [bp]	Calibrated Conc. [ng/µl]	Assigned Conc. [ng/µl]	% Integrated Area	From [bp]	To [bp]	Peak Comment	Observations
100	8.50	8.50	-	66	148		Lower Marker
1363	1.79	-	25.81	1234	1579		
9428	5.08	-	73.24	4356	45390		
-	-	-	-	-	-		Sample Well

© Agilent Technologies, Inc. 2023

F1: 6



Sample Table

Well	DIN	Conc. [ng/µl]	Sample Description	Alert	Observations
F1	5.7	17.0	6	⚠	Caution! Expired ScreenTape device

Peak Table

Size [bp]	Calibrated Conc. [ng/µl]	Assigned Conc. [ng/µl]	% Integrated Area	From [bp]	To [bp]	Peak Comment	Observations
100	8.50	8.50	-	66	146		Lower Marker
9723	5.70	-	97.04	4748	34195		
-	-	-	-	-	-		Sample Well

© Agilent Technologies, Inc. 2023

L2: Ladder



Sample Table

Well	DIN	Conc. [ng/µl]	Sample Description	Alert	Observations
L2	-	84.3	Ladder	⚠	Caution! Expired ScreenTape device; Ladder

Peak Table

Size [bp]	Calibrated Conc. [ng/µl]	Assigned Conc. [ng/µl]	% Integrated Area	From [bp]	To [bp]	Peak Comment	Observations
100	8.50	8.50	-	69	148		Lower Marker
250	6.40	-	7.82	183	310		
400	6.44	-	7.87	340	474		
600	6.64	-	8.12	524	716		
900	6.73	-	8.22	789	1053		
1200	6.91	-	8.44	1053	1353		
1500	7.06	-	8.63	1353	1739		
2000	6.71	-	8.19	1808	2267		
2500	6.35	-	7.76	2299	2757		
3000	6.79	-	8.30	2757	3529		
4000	6.43	-	7.86	3555	5330		
7000	6.54	-	8.00	5564	11343		
15000	5.16	-	6.30	11343	22526		
48500	3.67	-	4.49	22526	71194		

© Agilent Technologies, Inc. 2023

G1: 8



Sample Table

Well	DIN	Conc. [ng/µl]	Sample Description	Alert	Observations
G1	5.3	25.9	8	⚠	Caution! Expired ScreenTape device

Peak Table

Size [bp]	Calibrated Conc. [ng/µl]	Assigned Conc. [ng/µl]	% Integrated Area	From [bp]	To [bp]	Peak Comment	Observations
100	8.50	8.50	-	67	156		Lower Marker
1350	2.24	-	19.43	1144	1498		
12582	9.27	-	80.57	4325	55610		

© Agilent Technologies, Inc. 2023

H1: 9



Sample Table

Well	DIN	Conc. [ng/µl]	Sample Description	Alert	Observations
H1	4.5	7.62	9	⚠	Caution! Expired ScreenTape device; Sample concentration outside recommended range

Peak Table

Size [bp]	Calibrated Conc. [ng/µl]	Assigned Conc. [ng/µl]	% Integrated Area	From [bp]	To [bp]	Peak Comment	Observations
100	8.50	8.50	-	64	152		Lower Marker
-	-	-	-	-	-		Sample Well

© Agilent Technologies, Inc. 2023

A2: 10



Sample Table

Well	DIN	Conc. [ng/µl]	Sample Description	Alert	Observations
A2	6.6	13.4	10	⚠	Caution! Expired ScreenTape device

Peak Table

Size [bp]	Calibrated Conc. [ng/µl]	Assigned Conc. [ng/µl]	% Integrated Area	From [bp]	To [bp]	Peak Comment	Observations
100	8.50	8.50	-	65	147		Lower Marker
28066	6.58	-	99.53	6566	>60000		
-	-	-	-	-	-		Sample Well

© Agilent Technologies, Inc. 2023

B2: 11



Sample Table

Well	DIN	Conc. [ng/µl]	Sample Description	Alert	Observations
B2	4.8	8.96	11	⚠	Caution! Expired ScreenTape device; Sample concentration outside recommended range

Peak Table

Size [bp]	Calibrated Conc. [ng/µl]	Assigned Conc. [ng/µl]	% Integrated Area	From [bp]	To [bp]	Peak Comment	Observations
100	8.50	8.50	-	66	147		Lower Marker
10713	1.21	-	82.95	6432	26242		
-	-	-	-	-	-		Sample Well

© Agilent Technologies, Inc. 2023