



sam_2026-10-08_09-19-38_Connect-VIPERIN-DNMT1.pcrd

10/8/2026 17:35

Report Information

User: BioRad/sam
Data File Name: sam_2026-10-08_09-19-38_Connect-VIPERIN-DNMT1.pcrd
Data File Path: C:\Users\Samb\Desktop
Well Group Name: All Wells
Report Differs from Last Save: Yes

Run Setup

Run Information

Run Date: 10/8/2026 09:20
Run User: sam
Run Type: User-defined
Plate File: 20261008-mgig-VIPERIN-DNMT1-polyIC-valentina.pltd
ID:
Notes:
Sample Volume: 20
Temperature Control Mode: Calculated
Lid Temperature: 105
Base Serial Number: BR006896
Optical Head Serial Number: 788BR07000

Protocol

- 1: 95.0°C for 0:30
- 2: 95.0°C for 0:03
- 3: 60.0°C for 0:05
- Plate Read
- 4: GOTO 2, 39 more times
- 5: Melt Curve 65.0°C to 95.0°C: Increment 0.5°C 0:05
- Plate Read

Plate Display


	1	2	3	4	5	6	7	8	9	10	11	12
A	*Unk-1 VIPERIN A1PC	*Unk-1 VIPERIN A1PC	*Unk-1 VIPERIN A1PC	*Unk-2 VIPERIN A2PC	*Unk-2 VIPERIN A2PC	*Unk-2 VIPERIN A2PC	*Unk-3 VIPERIN A3PC	*Unk-3 VIPERIN A3PC	*Unk-3 VIPERIN A3PC	*Unk-4 VIPERIN A4PC	*Unk-4 VIPERIN A4PC	*Unk-4 VIPERIN A4PC
B	*Unk-5 VIPERIN A5PC	*Unk-5 VIPERIN A5PC	*Unk-5 VIPERIN A5PC	*Unk-6 VIPERIN B1PC	*Unk-6 VIPERIN B1PC	*Unk-6 VIPERIN B1PC	*Unk-7 VIPERIN B2PC	*Unk-7 VIPERIN B2PC	*Unk-7 VIPERIN B2PC	*Unk-8 VIPERIN B3PC	*Unk-8 VIPERIN B3PC	*Unk-8 VIPERIN B3PC
C	*Unk-9 VIPERIN B4PC	*Unk-9 VIPERIN B4PC	*Unk-9 VIPERIN B4PC	*Unk-10 VIPERIN B5PC	*Unk-10 VIPERIN B5PC	*Unk-10 VIPERIN B5PC	*NTC-1 VIPERIN	*NTC-1 VIPERIN	*NTC-1 VIPERIN			
D	Unk-11 DNMT1 A1PC	Unk-11 DNMT1 A1PC	Unk-11 DNMT1 A1PC	Unk-12 DNMT1 A2PC	Unk-12 DNMT1 A2PC	Unk-12 DNMT1 A2PC	Unk-13 DNMT1 A3PC	Unk-13 DNMT1 A3PC	Unk-13 DNMT1 A3PC	Unk-14 DNMT1 A4PC	Unk-14 DNMT1 A4PC	Unk-14 DNMT1 A4PC

Plate Display

	1	2	3	4	5	6	7	8	9	10	11	12
E	Unk-15 DNMT1 A5PC	Unk-15 DNMT1 A5PC	Unk-15 DNMT1 A5PC	Unk-16 DNMT1 B1PC	Unk-16 DNMT1 B1PC	Unk-16 DNMT1 B1PC	Unk-17 DNMT1 B2PC	Unk-17 DNMT1 B2PC	Unk-17 DNMT1 B2PC	Unk-18 DNMT1 B3PC	Unk-18 DNMT1 B3PC	Unk-18 DNMT1 B3PC
F	Unk-19 DNMT1 B4PC	Unk-19 DNMT1 B4PC	Unk-19 DNMT1 B4PC	Unk-20 DNMT1 B5PC	Unk-20 DNMT1 B5PC	Unk-20 DNMT1 B5PC	NTC-2 DNMT1	NTC-2 DNMT1	NTC-2 DNMT1			
G												
H												

Quantification

Step #: 3
Analysis Mode: Target
Cq Determination: Single Threshold
Baseline Method:
DNMT1: Auto Calculated
Threshold Setting:
DNMT1: 5976.77, Auto Calculated



Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev
D01	SYBR	DNMT1	Unkn-11	A1PC	31.80	31.94	0.124
D02	SYBR	DNMT1	Unkn-11	A1PC	32.00	31.94	0.124
D03	SYBR	DNMT1	Unkn-11	A1PC	32.02	31.94	0.124
D04	SYBR	DNMT1	Unkn-12	A2PC	30.64	30.62	0.044
D05	SYBR	DNMT1	Unkn-12	A2PC	30.57	30.62	0.044
D06	SYBR	DNMT1	Unkn-12	A2PC	30.66	30.62	0.044
D07	SYBR	DNMT1	Unkn-13	A3PC	32.37	32.16	0.216
D08	SYBR	DNMT1	Unkn-13	A3PC	32.17	32.16	0.216
D09	SYBR	DNMT1	Unkn-13	A3PC	31.94	32.16	0.216
D10	SYBR	DNMT1	Unkn-14	A4PC	32.10	32.17	0.069

Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev
D11	SYBR	DNMT1	Unkn-14	A4PC	32.18	32.17	0.069
D12	SYBR	DNMT1	Unkn-14	A4PC	32.24	32.17	0.069
E01	SYBR	DNMT1	Unkn-15	A5PC	30.78	30.60	0.166
E02	SYBR	DNMT1	Unkn-15	A5PC	30.55	30.60	0.166
E03	SYBR	DNMT1	Unkn-15	A5PC	30.45	30.60	0.166
E04	SYBR	DNMT1	Unkn-16	B1PC	30.57	30.62	0.366
E05	SYBR	DNMT1	Unkn-16	B1PC	31.00	30.62	0.366
E06	SYBR	DNMT1	Unkn-16	B1PC	30.28	30.62	0.366
E07	SYBR	DNMT1	Unkn-17	B2PC	32.85	32.71	0.202
E08	SYBR	DNMT1	Unkn-17	B2PC	32.81	32.71	0.202
E09	SYBR	DNMT1	Unkn-17	B2PC	32.48	32.71	0.202
E10	SYBR	DNMT1	Unkn-18	B3PC	31.14	30.99	0.129
E11	SYBR	DNMT1	Unkn-18	B3PC	30.95	30.99	0.129
E12	SYBR	DNMT1	Unkn-18	B3PC	30.89	30.99	0.129
F01	SYBR	DNMT1	Unkn-19	B4PC	29.36	30.83	2.426
F02	SYBR	DNMT1	Unkn-19	B4PC	29.51	30.83	2.426
F03	SYBR	DNMT1	Unkn-19	B4PC	33.63	30.83	2.426
F04	SYBR	DNMT1	Unkn-20	B5PC	N/A	0.00	0.000
F05	SYBR	DNMT1	Unkn-20	B5PC	N/A	0.00	0.000
F06	SYBR	DNMT1	Unkn-20	B5PC	37.88	37.88	0.000
F07	SYBR	DNMT1	NTC-02		N/A	0.00	0.000
F08	SYBR	DNMT1	NTC-02		N/A	0.00	0.000
F09	SYBR	DNMT1	NTC-02		N/A	0.00	0.000

QC Parameters

Data

Description	Value	Use	Results	Exclude Wells	All excluded wells
Negative control with a Cq less than	38	True		False	
NTC with a Cq less than	38	True		False	
NRT with a Cq less than	38	True		False	
Positive control with a Cq greater than	30	True		False	
Unknown without a Cq	N/A	True	DNMT1:F4, F5.	False	
Standard without a Cq	N/A	True		False	

Data

Description	Value	Use	Results	Exclude Wells	All excluded wells
Efficiency greater than	110.0	True			
Efficiency less than	90.0	True			
Std Curve R^2 less than	0.980	True			
Replicate group Cq Std Dev greater than	0.20	True	DNMT1:D7, D8, D9, E4, E5, E6, E7, E8, E9, F1, F2, F3.	False	