



sam_2026-10-08_09-19-38_Connect-VIPERIN-DNMT1.pcrd

10/8/2026 17:33

Report Information

User: BioRad/sam
Data File Name: sam_2026-10-08_09-19-38_Connect-VIPERIN-DNMT1.pcrd
Data File Path: C:\Users\Samb\Desktop
Well Group Name: All Wells
Report Differs from Last Save: Yes

Run Setup

Run Information

Run Date: 10/8/2026 09:20
Run User: sam
Run Type: User-defined
Plate File: 20261008-mgig-VIPERIN-DNMT1-polyIC-valentina.pltd
ID:
Notes:
Sample Volume: 20
Temperature Control Mode: Calculated
Lid Temperature: 105
Base Serial Number: BR006896
Optical Head Serial Number: 788BR07000

Protocol

- 1: 95.0°C for 0:30
- 2: 95.0°C for 0:03
- 3: 60.0°C for 0:05
- Plate Read
- 4: GOTO 2, 39 more times
- 5: Melt Curve 65.0°C to 95.0°C: Increment 0.5°C 0:05
- Plate Read

Plate Display


	1	2	3	4	5	6	7	8	9	10	11	12
A	Unk-1 VIPERIN A1PC	Unk-1 VIPERIN A1PC	Unk-1 VIPERIN A1PC	Unk-2 VIPERIN A2PC	Unk-2 VIPERIN A2PC	Unk-2 VIPERIN A2PC	Unk-3 VIPERIN A3PC	Unk-3 VIPERIN A3PC	Unk-3 VIPERIN A3PC	Unk-4 VIPERIN A4PC	Unk-4 VIPERIN A4PC	Unk-4 VIPERIN A4PC
B	Unk-5 VIPERIN A5PC	Unk-5 VIPERIN A5PC	Unk-5 VIPERIN A5PC	Unk-6 VIPERIN B1PC	Unk-6 VIPERIN B1PC	Unk-6 VIPERIN B1PC	Unk-7 VIPERIN B2PC	Unk-7 VIPERIN B2PC	Unk-7 VIPERIN B2PC	Unk-8 VIPERIN B3PC	Unk-8 VIPERIN B3PC	Unk-8 VIPERIN B3PC
C	Unk-9 VIPERIN B4PC	Unk-9 VIPERIN B4PC	Unk-9 VIPERIN B4PC	Unk-10 VIPERIN B5PC	Unk-10 VIPERIN B5PC	Unk-10 VIPERIN B5PC	NTC-1 VIPERIN	NTC-1 VIPERIN	NTC-1 VIPERIN			
D	*Unk-11 DNMT1 A1PC	*Unk-11 DNMT1 A1PC	*Unk-11 DNMT1 A1PC	*Unk-12 DNMT1 A2PC	*Unk-12 DNMT1 A2PC	*Unk-12 DNMT1 A2PC	*Unk-13 DNMT1 A3PC	*Unk-13 DNMT1 A3PC	*Unk-13 DNMT1 A3PC	*Unk-14 DNMT1 A4PC	*Unk-14 DNMT1 A4PC	*Unk-14 DNMT1 A4PC

Plate Display

	1	2	3	4	5	6	7	8	9	10	11	12
E	*Unk-15 DNMT1 A5PC	*Unk-15 DNMT1 A5PC	*Unk-15 DNMT1 A5PC	*Unk-16 DNMT1 B1PC	*Unk-16 DNMT1 B1PC	*Unk-16 DNMT1 B1PC	*Unk-17 DNMT1 B2PC	*Unk-17 DNMT1 B2PC	*Unk-17 DNMT1 B2PC	*Unk-18 DNMT1 B3PC	*Unk-18 DNMT1 B3PC	*Unk-18 DNMT1 B3PC
F	*Unk-19 DNMT1 B4PC	*Unk-19 DNMT1 B4PC	*Unk-19 DNMT1 B4PC	*Unk-20 DNMT1 B5PC	*Unk-20 DNMT1 B5PC	*Unk-20 DNMT1 B5PC	*NTC-2 DNMT1	*NTC-2 DNMT1	*NTC-2 DNMT1			
G												
H												

Quantification

Step #: 3
Analysis Mode: Target
Cq Determination: Single Threshold
Baseline Method:
VIPERIN: Auto Calculated
Threshold Setting:
VIPERIN: 6458.13, Auto Calculated



Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev
A01	SYBR	VIPERIN	Unkn-01	A1PC	29.55	29.51	0.105
A02	SYBR	VIPERIN	Unkn-01	A1PC	29.60	29.51	0.105
A03	SYBR	VIPERIN	Unkn-01	A1PC	29.39	29.51	0.105
A04	SYBR	VIPERIN	Unkn-02	A2PC	29.74	29.94	0.244
A05	SYBR	VIPERIN	Unkn-02	A2PC	29.86	29.94	0.244
A06	SYBR	VIPERIN	Unkn-02	A2PC	30.21	29.94	0.244
A07	SYBR	VIPERIN	Unkn-03	A3PC	30.25	30.27	0.039
A08	SYBR	VIPERIN	Unkn-03	A3PC	30.25	30.27	0.039
A09	SYBR	VIPERIN	Unkn-03	A3PC	30.32	30.27	0.039
A10	SYBR	VIPERIN	Unkn-04	A4PC	31.33	31.59	0.268

Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev
A11	SYBR	VIPERIN	Unkn-04	A4PC	31.57	31.59	0.268
A12	SYBR	VIPERIN	Unkn-04	A4PC	31.87	31.59	0.268
B01	SYBR	VIPERIN	Unkn-05	A5PC	30.02	30.06	0.048
B02	SYBR	VIPERIN	Unkn-05	A5PC	30.04	30.06	0.048
B03	SYBR	VIPERIN	Unkn-05	A5PC	30.12	30.06	0.048
B04	SYBR	VIPERIN	Unkn-06	B1PC	31.64	31.65	0.180
B05	SYBR	VIPERIN	Unkn-06	B1PC	31.48	31.65	0.180
B06	SYBR	VIPERIN	Unkn-06	B1PC	31.84	31.65	0.180
B07	SYBR	VIPERIN	Unkn-07	B2PC	30.69	30.88	0.197
B08	SYBR	VIPERIN	Unkn-07	B2PC	31.08	30.88	0.197
B09	SYBR	VIPERIN	Unkn-07	B2PC	30.87	30.88	0.197
B10	SYBR	VIPERIN	Unkn-08	B3PC	31.08	31.00	0.087
B11	SYBR	VIPERIN	Unkn-08	B3PC	31.02	31.00	0.087
B12	SYBR	VIPERIN	Unkn-08	B3PC	30.91	31.00	0.087
C01	SYBR	VIPERIN	Unkn-09	B4PC	29.57	29.69	0.125
C02	SYBR	VIPERIN	Unkn-09	B4PC	29.68	29.69	0.125
C03	SYBR	VIPERIN	Unkn-09	B4PC	29.82	29.69	0.125
C04	SYBR	VIPERIN	Unkn-10	B5PC	29.15	29.11	0.176
C05	SYBR	VIPERIN	Unkn-10	B5PC	28.93	29.11	0.176
C06	SYBR	VIPERIN	Unkn-10	B5PC	29.27	29.11	0.176
C07	SYBR	VIPERIN	NTC-01		N/A	0.00	0.000
C08	SYBR	VIPERIN	NTC-01		N/A	0.00	0.000
C09	SYBR	VIPERIN	NTC-01		N/A	0.00	0.000

QC Parameters

Data

Description	Value	Use	Results	Exclude Wells	All excluded wells
Negative control with a Cq less than	38	True		False	
NTC with a Cq less than	38	True		False	
NRT with a Cq less than	38	True		False	
Positive control with a Cq greater than	30	True		False	
Unknown without a Cq	N/A	True		False	
Standard without a Cq	N/A	True		False	

Data

Description	Value	Use	Results	Exclude Wells	All excluded wells
Efficiency greater than	110.0	True			
Efficiency less than	90.0	True			
Std Curve R^2 less than	0.980	True			
Replicate group Cq Std Dev greater than	0.20	True	VIPERIN:A4, A5, A6, A10, A11, A12.	False	