



sam_2026-10-08_09-37-12_CFX96-HSP90-cGAS.pcrd

10/8/2026 17:38

Report Information

User: BioRad/sam
Data File Name: sam_2026-10-08_09-37-12_CFX96-HSP90-cGAS.pcrd
Data File Path: C:\Users\Samb\Desktop
Well Group Name: All Wells
Report Differs from Last Save: Yes

Run Setup

Run Information

Run Date: 10/8/2026 09:37
Run User: sam
Run Type: User-defined
Plate File: 20261008-mgig-HSP90-cGAS-polyIC-valentina.pltd
ID:
Notes:
Sample Volume: 20
Temperature Control Mode: Calculated
Lid Temperature: 105
Base Serial Number: CC009827
Optical Head Serial Number: 785BR3659

Protocol

- 1: 95.0°C for 0:30
- 2: 95.0°C for 0:03
- 3: 60.0°C for 0:05
- Plate Read
- 4: GOTO 2, 39 more times
- 5: Melt Curve 65.0°C to 95.0°C: Increment 0.5°C 0:05
- Plate Read

Plate Display


	1	2	3	4	5	6	7	8	9	10	11	12
A	*Unk-1 HSP90 A1PC	*Unk-1 HSP90 A1PC	*Unk-1 HSP90 A1PC	*Unk-2 HSP90 A2PC	*Unk-2 HSP90 A2PC	*Unk-2 HSP90 A2PC	*Unk-3 HSP90 A3PC	*Unk-3 HSP90 A3PC	*Unk-3 HSP90 A3PC	*Unk-4 HSP90 A4PC	*Unk-4 HSP90 A4PC	*Unk-4 HSP90 A4PC
B	*Unk-5 HSP90 A5PC	*Unk-5 HSP90 A5PC	*Unk-5 HSP90 A5PC	*Unk-6 HSP90 B1PC	*Unk-6 HSP90 B1PC	*Unk-6 HSP90 B1PC	*Unk-7 HSP90 B2PC	*Unk-7 HSP90 B2PC	*Unk-7 HSP90 B2PC	*Unk-8 HSP90 B3PC	*Unk-8 HSP90 B3PC	*Unk-8 HSP90 B3PC
C	*Unk-9 HSP90 B4PC	*Unk-9 HSP90 B4PC	*Unk-9 HSP90 B4PC	*Unk-10 HSP90 B5PC	*Unk-10 HSP90 B5PC	*Unk-10 HSP90 B5PC	*NTC-1 HSP90	*NTC-1 HSP90	*NTC-1 HSP90			
D	Unk-11 cGAS A1PC	Unk-11 cGAS A1PC	Unk-11 cGAS A1PC	Unk-12 cGAS A2PC	Unk-12 cGAS A2PC	Unk-12 cGAS A2PC	Unk-13 cGAS A3PC	Unk-13 cGAS A3PC	Unk-13 cGAS A3PC	Unk-14 cGAS A4PC	Unk-14 cGAS A4PC	Unk-14 cGAS A4PC

Plate Display

	1	2	3	4	5	6	7	8	9	10	11	12
E	Unk-15 cGAS A5PC	Unk-15 cGAS A5PC	Unk-15 cGAS A5PC	Unk-16 cGAS B1PC	Unk-16 cGAS B1PC	Unk-16 cGAS B1PC	Unk-17 cGAS B2PC	Unk-17 cGAS B2PC	Unk-17 cGAS B2PC	Unk-18 cGAS B3PC	Unk-18 cGAS B3PC	Unk-18 cGAS B3PC
F	Unk-19 cGAS B4PC	Unk-19 cGAS B4PC	Unk-19 cGAS B4PC	Unk-20 cGAS B5PC	Unk-20 cGAS B5PC	Unk-20 cGAS B5PC	NTC-2 cGAS	NTC-2 cGAS	NTC-2 cGAS			
G												
H												

Quantification

Step #: 3
Analysis Mode: Target
Cq Determination: Single Threshold
Baseline Method:
cGAS: Auto Calculated
Threshold Setting:
cGAS: 5079.66, Auto Calculated



Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev
D01	SYBR	cGAS	Unkn-11	A1PC	29.90	29.73	0.151
D02	SYBR	cGAS	Unkn-11	A1PC	29.67	29.73	0.151
D03	SYBR	cGAS	Unkn-11	A1PC	29.63	29.73	0.151
D04	SYBR	cGAS	Unkn-12	A2PC	27.84	27.96	0.193
D05	SYBR	cGAS	Unkn-12	A2PC	27.86	27.96	0.193
D06	SYBR	cGAS	Unkn-12	A2PC	28.19	27.96	0.193
D07	SYBR	cGAS	Unkn-13	A3PC	29.79	29.85	0.176
D08	SYBR	cGAS	Unkn-13	A3PC	29.71	29.85	0.176
D09	SYBR	cGAS	Unkn-13	A3PC	30.04	29.85	0.176
D10	SYBR	cGAS	Unkn-14	A4PC	31.11	31.27	0.136

Quantification Data

Well	Fluor	Target	Content	Sample	Cq	Cq Mean	Cq Std. Dev
D11	SYBR	cGAS	Unkn-14	A4PC	31.34	31.27	0.136
D12	SYBR	cGAS	Unkn-14	A4PC	31.36	31.27	0.136
E01	SYBR	cGAS	Unkn-15	A5PC	29.48	29.74	0.226
E02	SYBR	cGAS	Unkn-15	A5PC	29.84	29.74	0.226
E03	SYBR	cGAS	Unkn-15	A5PC	29.89	29.74	0.226
E04	SYBR	cGAS	Unkn-16	B1PC	31.44	31.40	0.190
E05	SYBR	cGAS	Unkn-16	B1PC	31.57	31.40	0.190
E06	SYBR	cGAS	Unkn-16	B1PC	31.19	31.40	0.190
E07	SYBR	cGAS	Unkn-17	B2PC	29.91	29.94	0.112
E08	SYBR	cGAS	Unkn-17	B2PC	29.86	29.94	0.112
E09	SYBR	cGAS	Unkn-17	B2PC	30.07	29.94	0.112
E10	SYBR	cGAS	Unkn-18	B3PC	29.70	29.65	0.059
E11	SYBR	cGAS	Unkn-18	B3PC	29.59	29.65	0.059
E12	SYBR	cGAS	Unkn-18	B3PC	29.66	29.65	0.059
F01	SYBR	cGAS	Unkn-19	B4PC	30.66	30.57	0.188
F02	SYBR	cGAS	Unkn-19	B4PC	30.35	30.57	0.188
F03	SYBR	cGAS	Unkn-19	B4PC	30.69	30.57	0.188
F04	SYBR	cGAS	Unkn-20	B5PC	30.56	30.57	0.019
F05	SYBR	cGAS	Unkn-20	B5PC	30.59	30.57	0.019
F06	SYBR	cGAS	Unkn-20	B5PC	30.55	30.57	0.019
F07	SYBR	cGAS	NTC-02		N/A	0.00	0.000
F08	SYBR	cGAS	NTC-02		39.16	39.16	0.000
F09	SYBR	cGAS	NTC-02		N/A	0.00	0.000

QC Parameters

Data

Description	Value	Use	Results	Exclude Wells	All excluded wells
Negative control with a Cq less than	38	True		False	
NTC with a Cq less than	38	True		False	
NRT with a Cq less than	38	True		False	
Positive control with a Cq greater than	30	True		False	
Unknown without a Cq	N/A	True		False	
Standard without a Cq	N/A	True		False	

Data

Description	Value	Use	Results	Exclude Wells	All excluded wells
Efficiency greater than	110.0	True			
Efficiency less than	90.0	True			
Std Curve R^2 less than	0.980	True			
Replicate group Cq Std Dev greater than	0.20	True	cGAS:E1, E2, E3.	False	