Component	Your Value	Standard Range			Flag
COMPLEMENT ACTIVITY ENZYME IMMUNOASSAY,	60 U/mL	31 - 60 U/mL			
C3 COMPLEMENT	111 mg/dL	90 - 180 mg/dL			
C4 COMPLEMENT	22 mg/dL	10 - 40 mg/dL			
COLOR, UA	Yellow				
(Note) Reference Range: dark yellow, pale yellow, straw, yel	low				
APPEARANCE	Cloudy		Clear		Α
GLUCOSE, TIMED URINE	Negative		Negative		
BILIRUBIN URINE	Negative		Negative		
KETONES UA	Negative		Negative		
SPECIFIC GRAVITY, URINE	1.010				
Reference Range: 1.001-1.035					
PH URINE	6.0		5.0 - 8.0		
PROTEIN, URINE	Negative		Negative		
NITRITE	Negative		Negative		
BLOOD UA	Trace		Negative		А
LEUKOCYTES ESTERASE	1+		Negative		Α
SQUAMOUS EPITHELIAL	10-20 /[HPF]		0 - 5 /[HPF]		Α
WBC, URINE	0-5 /[HPF]		0 - 5 /[HPF]		
RBC, URINE	0-2 /[HPF]		0 - 2 /[HPF]		
BACTERIA	NONE SEEN /[H	IPF]	NONE SEEN	/[HPF]	
HYALINE CASTS	1-5 /[LPF]		NONE SEEN	/[LPF]	Α
NOTE	SEE BELOW				
This urine was analyzed for the presence of WBC, RB other formed elements. Only those elements seen were reported.	C, bacteria, casts, a	and			
HLA-B27 ANTIGEN		Negative		Negative	
ANA SCREEN		Negative		Negative	

Component	Your Value Standard Range	Flag
SED RATE	3 mm/h	0 - 20 mm/h
C-REACTIVE PROTEIN	<0.3 mg/dL	0.0 - 0.4 mg/dL
RHEUMATOID FACTOR	<10 IU/mL	0 - 13 IU/mL
CYCLIC CITRULLIN PEPTIDE AB	Negative	Negative
CYCLIC CITRULL. PEPTIDE QUANT	< 0.5 U/mL	0.0 - 2.9 U/mL
TSH	0.925 uIU/mL	
FREE T4	1.32 ng/dL 0.93 - 1.70 ng/dL	
Biotin, present in some multivitamins and dietary interfere with this test, causing falsely high or low		
FREE T3	3.15 pg/mL 2.00 - 4.40 pg/mL	
ANTI-THYROID-PEROXIDASE AB	11.70 IU/mL	0.00 - 34.00 IU/mL
Biotin, present in some multivitamins and dietary interfere with this test, causing falsely high or low		
THYROGLOBULIN AB	15 IU/mL	0 - 115 IU/mL
Biotin, present in some multivitamins and dietary interfere with this test, causing falsely high or low		
COLOR, UA	Yellow	
(Note) Reference Range: dark yellow, pale yellow, straw,	yellow	
APPEARANCE	Clear	Clear
GLUCOSE, TIMED URINE	Negative	Negative
BILIRUBIN URINE	Negative	Negative
KETONES UA	Negative	Negative
SPECIFIC GRAVITY, URINE	1.016	
Reference Range: 1.001-1.035		
PH URINE	6.5	5.0 - 8.0
PROTEIN, URINE	Negative	Negative
NITRITE	Negative	Negative
BLOOD UA	Negative	Negative

Component	Your Value	Standard Range		Flag
LEUKOCYTES ESTERASE	Negative		Negative	
SQUAMOUS EPITHELIAL	0-5 /[HPF]		0 - 5 /[HPF]	
WBC, URINE	0-5 /[HPF]		0 - 5 /[HPF]	
RBC, URINE	0-2 /[HPF]		0 - 2 /[HPF]	
BACTERIA	NONE SEEN /[HPF	:]	NONE SEEN /[HPF]	
HYALINE CASTS	NONE SEEN /[LPF]]	NONE SEEN /[LPF]	
NOTE	SEE BELOW			
This urine was analyzed for the presence of other formed elements. Only those elements seen were reported. VITAMIN D, 25 (HYDROXY)	of WBC, RBC, bacteria, casts, a		30.00 - 80.00 ng/mL	L
VITAMIN B12	1,456 pg/mL		232 - 1,245 pg/mL	Н
FERRITIN	22.8 ng/mL		9.0 - 150.0 ng/mL	
Biotin, present in some multivitamins and interfere with this test, causing falsely high				
IRON	44 ug/dL	37 - 14	15 ug/dL	
PERCENT SATURATION	13 %	15 - 50	15 - 50 %	
IRON BINDING CAPACITY TOTAL	342 ug/dL	250 - 4	150 ug/dL	
UIBC	298.00 ug/dL	112.00	0 - 347.00 ug/dL	
WBC	4.3 10*3/uL	3.8 - 1	1.8 10*3/uL	
RBC	4.3 10*6/uL	4.0 - 5	.3 10*6/uL	
HEMOGLOBIN	12.9 g/dL	11.5 -	15.5 g/dL	
HEMATOCRIT	39.3 %	35.0 -	45.0 %	
MCV	91.7 fL	80.0 -	100.0 fL	
мсн	30.2 pg	27.0 -	34.0 pg	
мснс	32.9 g/dL	32.0 -	35.5 g/dL	
RDW	13.1 %	11.0 -	15.0 %	
PLATELET COUNT	243 10*3/uL	130 - 4	400 10*3/uL	
NEUTROS PCT	61.2 %	40.0 -	74.0 %	

Component	Your Value	Standard Range	Flag
LYMPH %	27.5 %	19.0 - 48.0 %	
MONO %	9.1 %	4.0 - 13.0 %	
EOS PCT	1.5 %	0.0 - 7.0 %	
BASOS PCT	0.7 %	0.0 - 2.0 %	
NEUTROS ABS	2.6 10*3/uL	1.9 - 8.2 10*3/uL	
LYMPHS ABS	1.2 10*3/uL	1.1 - 3.1 10*3/uL	
MONOS ABS	0.4 10*3/uL	0.2 - 0.9 10*3/uL	
ABSOLUTE EOSINOPHIL COUNT	0.1 10*3/uL	0.0 - 0.5 10*3/uL	
BASOS ABS	0.0 10*3/uL	0.0 - 0.1 10*3/uL	
GLUCOSE	112 mg/dL	70 - 99 mg/dL	Н

A normal fasting glucose concentration is less than 100 mg/dL. An impaired fasting glucose concentration is 100-125 mg/dL. A provisional diagnosis of diabetes mellitus can be made when a fasting glucose concentration is greater than 125 mg/dL.

BUN	11 mg/dL	6 - 20 mg/dL	
CREATININE	0.64 mg/dL	0.53 - 1.04 mg/dL	
ESTIMATED GLOMERULAR FILTRATION RATE	116 mL/min/1.73m2	>=60 mL/min/1.73m2	
BUN / CREAT RATIO	17	7 - 25	
SODIUM	139 meq/L	136 - 145 meq/L	
POTASSIUM	4.6 meq/L	3.5 - 5.3 meq/L	
CHLORIDE	106 meq/L	98 - 107 meq/L	
CARBON DIOXIDE	25 meq/L	23 - 31 meq/L	
CALCIUM	9.4 mg/dL	8.7 - 10.4 mg/dL	
PROTEIN, TOTAL	6.6 g/dL	6.1 - 7.9 g/dL	
ALBUMIN	4.4 g/dL	3.9 - 5.2 g/dL	
GLOBULIN, CALCULATED	2.2 g/dL	1.7 - 3.3 g/dL	
A/G RATIO	2.0	1.3 - 2.8	
BILIRUBIN, TOTAL	< 0.2 mg/dL	<1.2 mg/dL	
ALKALINE PHOSPHATASE	60 ∪/L	40 - 135 U/L	

Component	Your Value	Standard Range		Flag
SGOT (AST)	22 U/L		0 - 32 U/L	
SGPT (ALT)	19 U/L		0 - 33 U/L	
TOTAL ESTROGENS	278.3 pg/mL			pg/mL
Follicular Phase (1-12 days): 90-590 pg/mL Luteal Phase: 130-460 pg/mL Postmenopausal: 50-170 pg/mL				
PROGESTERONE	5.57 ng/mL		ng/	mL
PROGESTERONE Male: <0.050-0.149 ng/mL				
Female: Cycle phase Follicular: <0.050-0.193 ng/mL Ovulation: 0.055-4.140 ng/mL				

Intermediate Luteal: 4.76-20.90 ng/mL

Late luteal: 0.537-13.500 ng/mL

Postmenopausal females: <0.050-0.126 ng/mL

Pregnant females:

First trimester: 11.00-44.30 ng/mL Second trimester: 25.40-83.40 ng/mL Third trimester: 58.70-214.00 ng/mL

Biotin, present in some multivitamins and dietary supplements, may

interfere with this test, causing falsely high or low results.

TESTOSTERONE FREE 0.1 - 6.4 pg/mL 0.1 - 6.4 pg/mL

(Note)

This test was developed and its analytical performance

characteristics have been determined by medfusion. It has not been

	ndard Range	Flag
cleared or approved by the FDA. This assay has been validated		
oursuant to the CLIA regulations and is used for clinical purposes.		
ULTRASENSITIVE TESTOSTERONE, FEMALES AND CHILDREN	53 ng/dL	2 - 45 ng/dL
For additional information, please refer to		
nttps://education.questdiagnostics.com/faq/FAQ165		
This link is being provided for informational/educational purposes		
only.)		
Note) This test was developed and its analytical performance		
characteristics have been determined by medfusion. It has not		
peen cleared or approved by the FDA. This assay has been validated		
oursuant to the CLIA regulations and is used for clinical purposes.		
DOUBLE STRAND DNA ANTIBODY (DS DNA) ANTIBODY	< 1 IU/mL	<=4 IU/mL
NTERPRETIVE INFORMATION: ds DNA Multiplex Flow AutoAB		
Negative: <= 4 IU/mL		
ndeterminate: 5-9 IU/mL		
Positive: >= 10 IU/mL		
T .:		
Testing performed on Bioplex 2200 utilizing multiplexed bead immunoassay.		No
CENTROMERE AUTO AB	Negative	Negative
CENTROMERE AUTO AB Testing performed on Bioplex 2200 utilizing multiplexed bead immunoassay.	Negative	
CENTROMERE AUTO AB	Negative	Negative
CENTROMERE AUTO AB Testing performed on Bioplex 2200 utilizing multiplexed bead immunoassay. CENTROMERE AUTOAB QUANT Al has not been correlated with disease state or intensity of disease. It	_	
CENTROMERE AUTO AB Testing performed on Bioplex 2200 utilizing multiplexed bead immunoassay. CENTROMERE AUTOAB QUANT	_	
Testing performed on Bioplex 2200 utilizing multiplexed bead immunoassay. CENTROMERE AUTOAB QUANT Al has not been correlated with disease state or intensity of disease. It is not supported or documented for use other than for pos/neg.	_	
CENTROMERE AUTO AB Testing performed on Bioplex 2200 utilizing multiplexed bead immunoassay. CENTROMERE AUTOAB QUANT Al has not been correlated with disease state or intensity of disease. It is not supported or documented for use other than for pos/neg. Testing performed on Bioplex 2200 utilizing multiplexed bead immunoassay.	< 0.2 AI	<1.0 AI
CENTROMERE AUTO AB Testing performed on Bioplex 2200 utilizing multiplexed bead immunoassay. CENTROMERE AUTOAB QUANT Al has not been correlated with disease state or intensity of disease. It is not supported or documented for use other than for pos/neg. Testing performed on Bioplex 2200 utilizing multiplexed bead immunoassay. CHROMATIN	_	
Testing performed on Bioplex 2200 utilizing multiplexed bead immunoassay. CENTROMERE AUTOAB QUANT Al has not been correlated with disease state or intensity of disease. It is not supported or documented for use other than for pos/neg. Testing performed on Bioplex 2200 utilizing multiplexed bead immunoassay. CHROMATIN Testing performed on Bioplex 2200 utilizing multiplexed bead immunoassay.	<0.2 AI Negative	<1.0 AI Negative
CENTROMERE AUTO AB Testing performed on Bioplex 2200 utilizing multiplexed bead immunoassay. CENTROMERE AUTOAB QUANT Al has not been correlated with disease state or intensity of disease. It is not supported or documented for use other than for pos/neg. Testing performed on Bioplex 2200 utilizing multiplexed bead immunoassay. CHROMATIN	< 0.2 AI	<1.0 AI
Testing performed on Bioplex 2200 utilizing multiplexed bead immunoassay. CENTROMERE AUTOAB QUANT All has not been correlated with disease state or intensity of disease. It is not supported or documented for use other than for pos/neg. Testing performed on Bioplex 2200 utilizing multiplexed bead immunoassay. CHROMATIN Testing performed on Bioplex 2200 utilizing multiplexed bead immunoassay. CHROMATIN AUTOAB QUANT All has not been correlated with disease state or intensity of disease. It	<0.2 AI Negative	<1.0 AI Negative
Testing performed on Bioplex 2200 utilizing multiplexed bead immunoassay. CENTROMERE AUTOAB QUANT All has not been correlated with disease state or intensity of disease. It is not supported or documented for use other than for pos/neg. Testing performed on Bioplex 2200 utilizing multiplexed bead immunoassay. CHROMATIN Testing performed on Bioplex 2200 utilizing multiplexed bead immunoassay. CHROMATIN AUTOAB QUANT	<0.2 AI Negative	<1.0 AI Negative
Testing performed on Bioplex 2200 utilizing multiplexed bead immunoassay. CENTROMERE AUTOAB QUANT All has not been correlated with disease state or intensity of disease. It is not supported or documented for use other than for pos/neg. Testing performed on Bioplex 2200 utilizing multiplexed bead immunoassay. CHROMATIN Testing performed on Bioplex 2200 utilizing multiplexed bead immunoassay. CHROMATIN AUTOAB QUANT All has not been correlated with disease state or intensity of disease. It	<0.2 AI Negative	<1.0 AI Negative
Testing performed on Bioplex 2200 utilizing multiplexed bead immunoassay. CENTROMERE AUTOAB QUANT All has not been correlated with disease state or intensity of disease. It is not supported or documented for use other than for pos/neg. Testing performed on Bioplex 2200 utilizing multiplexed bead immunoassay. CHROMATIN Testing performed on Bioplex 2200 utilizing multiplexed bead immunoassay. CHROMATIN AUTOAB QUANT All has not been correlated with disease state or intensity of disease. It is not supported or documented for use other than for pos/neg.	<0.2 AI Negative	<1.0 AI Negative

Component Yo	our Value Standard Range	Flag
JO-1 AB IGG QUANT	< 0.2 AI	<1.0 AI
Al has not been correlated with disease state or intensity is not supported or documented for use other than for p	•	
Testing performed on Bioplex 2200 utilizing multiplexed	bead immunoassay.	
RIBOSOMAL P AUTOAB	Negative	Negative
Testing performed on Bioplex 2200 utilizing multiplexed	bead immunoassay.	
RIBOSOMAL P AUTOAB QUANT	< 0.2 AI	<1.0 AI
Al has not been correlated with disease state or intensity is not supported or documented for use other than for p	oos/neg.	
Testing performed on Bioplex 2200 utilizing multiplexed RNP EXTRACTABLE NUCLEAR AB	Negative	Negative
Testing performed on Bioplex 2200 utilizing multiplexed	н	
RNP EXTRACTABLE NUCLEAR AB QUANT	< 0.2 AI	<1.0 AI
Al has not been correlated with disease state or intensition is not supported or documented for use other than for particles are the properties of the state of t	oos/neg.	
SCL-70 IGG AUTOAB	Negative	Negative
Testing performed on Bioplex 2200 utilizing multiplexed	bead immunoassay.	
SCL-70 IGG AUTOAB QUANT	<0.2 AI	<1.0 AI
Al has not been correlated with disease state or intensity is not supported or documented for use other than for particles. Testing performed on Bioplex 2200 utilizing multiplexed	oos/neg.	
SMITH AUTOANTIBODY	Negative	Negative
Testing performed on Bioplex 2200 utilizing multiplexed	bead immunoassay.	
SMITH AUTOAB QUANT	< 0.2 AI	<1.0 AI
Al has not been correlated with disease state or intensition is not supported or documented for use other than for p	•	
Testing performed on Bioplex 2200 utilizing multiplexed	bead immunoassay.	
SM RNP AUTOANTIBODY	Negative	Negative

Component	Your Value	Standard Range	Flag
Testing performed on Bioplex 2200 utilizing multip	olexed bead immunoassay.		
SM RNP AUTOANTIBODY QUANT		< 0.2 AI	<1.0 AI
AI has not been correlated with disease state or in its not supported or documented for use other that	•		
Testing performed on Bioplex 2200 utilizing multip	lexed bead immunoassay.		
SSA/RO IGG AUTOAB		Negative	Negative
Testing performed on Bioplex 2200 utilizing multip	lexed bead immunoassay.		
SSA/RO IGG AUTOAB QUANT		< 0.2 AI	<1.0 AI
AI has not been correlated with disease state or in some supported or documented for use other that Testing performed on Bioplex 2200 utilizing multip	n for pos/neg.		
SSB/LA IGG AUTOAB	•	Negative	Negative
Testing performed on Bioplex 2200 utilizing multip	plexed bead immunoassay.		
SSB/LA IGG AUTOAB QUANT		< 0.2 AI	<1.0 AI
AI has not been correlated with disease state or in is not supported or documented for use other that Testing performed on Bioplex 2200 utilizing multip	n for pos/neg.		
ENA INTERPRETATION		SEE COMMENT	
All antibody levels for systemic autoimmune disease pre-established cutoffs. MDSS outputs of "Negative not rule out autoimmune disease. Patients with Rhresult in an SLE association from MDSS, thus MDSS patients with RA should be interpreted with caution	e" or "No Association" do neumatoid Arthritis may S associations from		
URINE CULTURE	10,000 - 50,000 cfu/n	nl, Mixed urogenital flora	Α
URINE CULTURE	Escherichia coli 50,000 - 100,000 cfu/	'ml	Α
URINE CULTURE	Enterococcus faecalis 50,000 - 100,000 cfu/		Α

Comments from the Doctor's Office

You urine dip shows only trace blood and leukocytes (white blood cells). This could be a sign on an infection or contamination. I am waiting the final urine culture results which will let us know if this is a UTI and which antibiotics will treat it. I will follow up once those results are received.

Thanks, Megan Hubbard, PA-C

Component Results

Component	Your Value	Standard Range	Flag
APPEARANCE	Clear	Clear	
COLOR UA	Yellow	Colorless, Straw, Yellow, Amber	
SPEC GRAV UA	1.020	1.005 - 1.035	
PH URINE	5.5	5 - 9	
PROTEIN, URINE, SEMIQUANTITATIVE (DIPSTICK)	Negative	Negative, 5-9, 0-4, 10-14, 15-19, 20-29, 30-49, >50	
GLUCOSE UA	Negative	Negative, 5-9, 0-4, 10-14, 15-19, 20-29, 30-49, >50	
KETONES UA	Negative	Negative	
BILIRUBIN URINE	Negative	Negative	
BLOOD	Trace	Negative, 5-9, 0-4, 10-14, 15-19, 20-29, 30-49, >50	A
UROBILINOGEN	0.2	0.2-1.0, 0.2, 1.0	
NITRITE	Negative	Negative	
LEUKOCYTE ESTERASE	Trace	Negative	A