

Schedule

KSA Metrology Pte Ltd
60 Kaki Bukit Place
#06-16 Eunos Techpark (Lobby A - Exit B)
Singapore 415979

Certificate No. : LA-2009-0444-C
Issue No. : 21
Date : 26 June 2026
Expiry of Certificate : 22 August 2028
Page : 1 of 4

FIELD OF TESTING: Calibration and Measurement

MEASURED QUANTITIES/ INSTRUMENTS/RANGE TO BE CALIBRATED	METHOD	CALIBRATION AND MEASUREMENT CAPABILITY (CMC*)
A Temperature		
1. Temperature Sensor (RTD)		
-80 °C	KSA23/014A (Rev 0)	0.30 °C
-30 °C to 140 °C	KSA23/001A (Rev 0)	0.34 °C
>140 °C to 200 °C	KSA23/001A (Rev 0)	0.36 °C
2. Electrical simulation for RTD controller / display / recorder transmitter / etc.		
-200 °C to 800 °C	KSA23/003A (Rev 0)	0.41 °C
3. Thermometer (digital display)		
-80 °C	KSA23/014A (Rev 0)	0.30 °C
-30 °C to 140 °C	KSA23/007A (Rev 0) & KSA23/006A (Rev 0)	0.34 °C
>140 °C to 200 °C		0.36 °C
4. Thermometer (analog display)	KSA23/029A (Rev 0)	
-30 °C to 140 °C		1.1 °C
>140 °C to 200 °C		1.1 °C
5. Loop line calibration (sensor & readout)		
-80 °C	KSA23/014A (Rev 0)	0.30 °C
-30 °C to 140 °C	KSA23/004A (Rev 0)	0.34 °C
>140 °C to 200 °C	KSA23/004A (Rev 0)	0.36 °C

Schedule



Certificate No.: LA-2009-0444-C

Issue No.: 21

Date : 26 June 2026

Page : 2 of 4

MEASURED QUANTITIES/ INSTRUMENTS/RANGE TO BE CALIBRATED	METHOD	CALIBRATION AND MEASUREMENT CAPABILITY (CMC*)
6. Dry block heater 25 °C to 140 °C	KSA23/025A (Rev 0)	0.30 °C
7. Temperature mapping (Autoclave, Climatic chamber, Fridge & Freezer, Oven, Incubator, Storage area, etc....) -80 °C -30 °C to 140 °C >140 °C to 200 °C	KSA23/008A (Rev 0) & KSA23/009A (Rev 0) & KSA23/010A (Rev 0)	0.70 °C 0.44 °C 0.45 °C
8. Humidity instruments (sensors, loggers, thermo hygrometer) a) at 20 to 40 °C (30 to 90) % relative humidity b) Salt solutions: saturated at 23 +/- 3 °C (10 and 35) % relative humidity (50; 75 and 95) % relative humidity	KSA23/021A (Rev 0) & KSA23/030A (Rev 0) & KSA23/027A (Rev 0) & KSA23/032A (Rev 0) KSA23/018A (Rev 0) & KSA23/032A (Rev 0)	0.60 °C 2.5 % relative humidity 1.2 % relative humidity 1.6 % relative humidity
9. Humidity mapping (20 to 90) % relative humidity	KSA23/010A (Rev 0)	2.2 % relative humidity
10. Liquid in Glass Thermometer Minimum scale division: 0.1 °C Partial Immersion Full Immersion 0 °C 30 °C to 200 °C	KSA23/028A (Rev 0) KSA23/022A (Rev 0)	0.21 °C 0.42 °C

Schedule



Certificate No.: LA-2009-0444-C

Issue No.: 21

Date : 26 June 2026

Page : 3 of 4

MEASURED QUANTITIES/ INSTRUMENTS/RANGE TO BE CALIBRATED	METHOD	CALIBRATION AND MEASUREMENT CAPABILITY (CMC*)
B Mechanical		
1. Pressure gauges (digital display)	KSA23/011A (Rev 0)	
(-900 to 0) mbar		3.0 mbar
(0 to 10 000) mbar		4.1 mbar
(10 000 to 20 000) mbar		6.0 mbar
2. Pressure gauges (analog display)	KSA23/017A (Rev 0)	
(-900 to 0) mbar		13 mbar
(0 to 4 000) mbar		14 mbar
3. Absolute pressure instruments	KSA23/016A (Rev 0)	
(100 to 1 000) mbar abs		4.7 mbar
(1 000 to 11 000) mbar abs		5.2 mbar
(11 000 to 21 000) mbar abs		7.0 mbar
4. Differential gauges	KSA23/012A (Rev 0)	
(0 to 250) Pascal		5 Pascal
(250 to 2 500) Pascal		23 Pascal
5. Balances and Weighing Scales	KSA23/039A (Rev 1)	
(up to 220) g		0.0006 g
(>220 to 500) g		0.002 g
(>500 to 6200) g		0.02 g
(>6200 to 12000) g		0.2 g
(>12000 to 40000) g		0.4 g
(>40 to 150) kg		0.04 kg
(>150 to 1000) kg		0.3 kg
6. Standard Weights		
OIML Class M1	KSA23/040A (Rev 1)	
1kg		17 mg
2kg		30 mg
5kg		80 mg
10kg		160 mg

Schedule



Certificate No.: LA-2009-0444-C

Issue No.: 21

Date : 26 June 2026

Page : 4 of 4

MEASURED QUANTITIES/ INSTRUMENTS/RANGE TO BE CALIBRATED	METHOD	CALIBRATION AND MEASUREMENT CAPABILITY (CMC*)
C TIME Stopwatch & Bench Timers Equipment's Timer's (5 to 3600) s	KSA23/013.1A (Rev 0) KSA23/013.2A (Rev 1)	0.25 s

* CMC is expressed as an expanded uncertainty estimated at a level of confidence of approximately 95 %.

Approved signatories:

Mr. Roedil Philip Alquerro GELERA B5 & B6

Mr. Eugene Dumali BALDOMAR For all items (except for B5 & B6)

Note:

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025. A laboratory's fulfilment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and **management system requirements** that are necessary for it to consistently deliver technically valid calibrations. The **management system requirements** in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001.