

DATA SOURCE: NINTENDO WII BALANCE BOARD (MODEL RVL-WBC-01) — LIVE HARDWARE

PATIENT INFORMATION

Patient Name: Eleanor Vance
Patient ID: pat-001
Date of Session: 9/30/2026
Start Time: 6:30:00 PM

SESSION METADATA

Session ID: session-20260930-testhw
Active Duration: 00:30 (30.0s)
Hardware Model: RVL-WBC-01
Data Source: Wii Balance Board

1. SESSION SUMMARY

Duration	00:30	Average Total Weight	64.60 kg
Valid Measurements	301	Average Left Loading	49.8 %
Invalid Measurements	0	Average Right Loading	50.2 %
Data Transmission Gaps	0	Average Anterior Loading	50.9 %
Nominal Sampling Rate	10 Hz	Average Posterior Loading	49.1 %

2. WEIGHT DISTRIBUTION & BODY MASS STABILITY


Average Weight	64.60 kg	Bilateral Symmetry (L/R)	49.8% / 50.2%
Minimum Weight	64.60 kg	Anteroposterior Balance (Ant/Post)	50.9% / 49.1%
Maximum Weight	64.60 kg	Weight Excursion Range	0.00 kg

3. DATA QUALITY ASSURANCE

Signal Integrity Status: 100% Valid Packet Stream — No sensor dropouts or numerical anomalies detected.
Transmission Continuity: Continuous high-frequency stream with zero transmission dropouts.
Effective Frequency: Sampling cadence operating at approximately 10 Hz throughout session.

4. CENTER OF PRESSURE (COP) DISPERSION & TRAJECTORY

Lateral COP X Range	[-0.350, 0.350]	Mean COP X (Lateral Bias)	+0.041
Anteroposterior COP Y Range	[-0.280, 0.280]	Mean COP Y (A/P Bias)	+0.012
Total COP Coordinate Samples	301	Coordinate Bounds	Normalized [-1.0 to +1.0]




2D Trajectory Interpretation Guide:


- The green marker indicates the patient's initial stance (START) at elapsed time 00:00.
- The red marker indicates the final recorded position (END) when the session stopped.
- The central dashed circle represents the 10% equilibrium balance zone around absolute neutral (0, 0).

5. TIME-SERIES DYNAMICS & POSTURAL SWAY TRAJECTORIES


Graph 1: Center of Pressure — Lateral Movement (COP X) Left (-X) | Center (0.0) | Right (+X)



Graph 2: Center of Pressure — Anteroposterior Movement (COP Y) Front / Anterior (+Y) | Center (0.0) | Back / Posterior (-Y)



Graph 3: Weight Distribution Over Time (Left % vs Right %) Y: 0% to 100% | Dashed Center: 50/50 Balanced Stance



6. FOUR-CHANNEL LOAD-CELL SENSOR BREAKDOWN

Channel	Anatomical Position	Mean Force	Min Force	Max Force	Load Share
Front-Left (FL)	Anterior-Left	16.24 kg	15.85 kg	16.55 kg	25.1%
Front-Right (FR)	Anterior-Right	16.26 kg	15.95 kg	16.65 kg	25.2%
Back-Left (BL)	Posterior-Left	16.04 kg	15.65 kg	16.35 kg	24.8%
Back-Right (BR)	Posterior-Right	16.06 kg	15.75 kg	16.45 kg	24.9%

7. OBSERVED SESSION CHARACTERISTICS (NON-DIAGNOSTIC)

- Bilateral weight distribution remained symmetric within 3% (Left: 49.8%, Right: 50.2%).
- Anteroposterior weight distribution remained centered within neutral parameters (Anterior: 50.9%, Posterior: 49.1%).
- The recorded Center of Pressure trajectory had an excursion span of 0.70 on the lateral axis and 0.56 on the anteroposterior axis.
- Total applied force stability remained within a range of 0.00 kg (Min: 64.60 kg, Max: 64.60 kg).

8. CLINICAL & REHABILITATION REVIEW NOTES

Clinician / Researcher observational notes & follow-up rehabilitation directives:

Reviewing Clinician / Investigator: _____

Date: _____