

OIML R-76 METROLOGY LABORATORY

NON-AUTOMATIC WEIGHING INSTRUMENT (NAWI) VERIFICATION

Report Ref: R76-20260911-5A65A9
Cert No: CERT-2026-RET-0192
Date: 2026-09-11
Test Type: In-Service Verification

OVERALL METROLOGICAL
RESULT: PASS

Standard: OIML R-76-1:2006 (E)
Turning Point Method: Enabled (Small Weights)
Instrument meets all tested requirements of OIML R-76-1.

1. NAWI & METROLOGICAL SPECIFICATIONS

Manufacturer:	Mettler Toledo / Avery	Accuracy Class:	Class III (Medium Accuracy)
Model / Type:	bRite Standard Retail NAWI	Max Capacity (Max):	15 kg
Serial Number:	BR-2026-88341	Min Capacity (Min):	0.1 kg
Device Type:	Electronic Counter-Top Scale	Scale Interval (e):	0.005 kg
Customer / Location:	Fresh Mart Hypermarket Ltd / Central Inspection Bay #2, Mumbai	Actual Interval (d):	0.005 kg
Inspector / Verifier:	Er. R. Sharma (Senior Metrologist)	Divisions n = Max/e:	3,000 divisions

Temperature:	22 °C	Humidity:	52 % RH	Pressure:	1012.5 hPa	Weights Ref:	OIML Class M1 Standard Weights (1g - 20kg) (NABL-CAL-2026-00912)
--------------	-------	-----------	---------	-----------	------------	--------------	---

2. SUMMARY OF METROLOGICAL EVALUATIONS (OIML R-76-1)

OIML R-76 Clause / Evaluation Subject	Requirement / Criterion	Result
Instrument Metrological Limits (Table 3)	divisions n within class range; Min capacity verified	PASS
Zero Setting Accuracy (A.4.2)	Zero error E0 <= 0.25 e	PASS
Tare Balancing & Net Weighing (A.4.6)	Net indication corrected error <= MPE	PASS
Weighing Performance & Hysteresis (A.4.4)	Increasing & Decreasing load errors <= MPE	PASS
Repeatability (A.4.10)	Span delta (Imax - Imin) <= MPE	PASS
Eccentricity (A.4.7)	Errors at 5 load receptor positions <= MPE	PASS

3. DETAILED METROLOGICAL TEST PROTOCOLS

Test Description	Nominal Load	Indication	Small Wt (ΔL)	Error (Ec)	MPE / Limit	Status
Zero Setting (A.4.2)	0.00 kg	0.0 kg	0.0025 kg	0.0 kg	± 0.00125 kg (0.25e)	PASS
Tare Balancing (A.4.6)	2.0 kg	2.0 kg	0.0025 kg	0.0 kg	± 0.0013 kg	PASS
Tare Net Weighing (A.4.6)	Net: 5.0 kg	5.0 kg	0.0025 kg	0.0 kg	± 0.01 kg	PASS

Eccentricity Test (clause A.4.7) — Test Load: 5.0 kg | MPE: ± 0.01 kg | Max Diff: 0.001 kg

Pos #	Position Name / Description	Indication	Small Wt (ΔL)	Error (E)	Corrected (Ec)	Status
1	Center (1)	5.0 kg	0.0025 kg	0.0 kg	0.0 kg	PASS
2	Front-Left (2)	5.0 kg	0.002 kg	0.0005 kg	0.0005 kg	PASS
3	Back-Left (3)	5.0 kg	0.003 kg	-0.0005 kg	-0.0005 kg	PASS
4	Back-Right (4)	5.0 kg	0.0025 kg	0.0 kg	0.0 kg	PASS
5	Front-Right (5)	5.0 kg	0.0025 kg	0.0 kg	0.0 kg	PASS

Repeatability Test (clause A.4.10) — Requirement: Δ = (I_max - I_min) ≤ |MPE|

Load Level	Observed Indications (Runs 1..N)	I_max	I_min	Spread (Δ)	MPE	Status
7.5 kg	7.5, 7.5, 7.505	7.505 kg	7.5 kg	0.005 kg	± 0.01 kg	PASS
15.0 kg	15, 15.005, 15	15.005 kg	15.0 kg	0.005 kg	± 0.015 kg	PASS

Weighing Performance Test (clause A.4.4) — Max Hysteresis: 0.0 kg

Load (L)	I (Inc)	Ec (Inc)	I (Dec)	Ec (Dec)	Hysteresis	MPE ($\pm$)	Status
0.1 kg	0.1 kg	0.0	0.1 kg	0.0 kg	0.0 kg	± 0.005 kg	PASS
2.5 kg	2.5 kg	0.0	2.5 kg	0.0 kg	0.0 kg	± 0.005 kg	PASS
5.0 kg	5.0 kg	0.0	5.0 kg	0.0 kg	0.0 kg	± 0.01 kg	PASS
10.0 kg	10.0 kg	0.0	10.0 kg	0.0 kg	0.0 kg	± 0.01 kg	PASS
15.0 kg	15.0 kg	0.0	15.0 kg	0.0 kg	0.0 kg	± 0.015 kg	PASS

Tested / Verified By:	Approved / Authorized By:	Official Seal & Verification Date:
		Date: 2026-09-11
Technical Metrologist / Inspector	Quality Manager / Signatory	Place: Central Inspection Bay #2, Mumbai
		Ref: R76-20260911-5A65A9

METROLOGICAL NOTICE & STATUTORY DISCLAIMER:
This verification test report was generated automatically by the NAWI Test Report Generation System based on computational evaluation of Non-Automatic Weighing Instruments conforming strictly to OIML R-76-1 (2006). This report is intended for college / academic prototyping, metrology software demonstrations, laboratory pre-compliance, and technical verification. Unless stamped, signed, and validated by an accredited legal metrology officer in accordance with national legislation, this report does not confer statutory legal metrology certification.