P15

The Fourier reconstruction of a head section

Authors	Larry A Shepp, Benjamin F. Logan

Published	IEEE Transactions on Nuclear Science (1974)
DOI	https://doi.org/10.1109/TNS.1974.6499235
Paper Link	https://doi.org/10.1109/TNS.1974.6499235
Domain	Biology & biomedical science

Abstract (quoted verbatim from the publication)

The Fourier reconstruction may be viewed simply in the spatial domain as the sum of each line integral times a weighting function of the distance from the line to the point of reconstruction. A modified weighting function simultaneously achieves accuracy, simplicity, low computation time, as well as low sensitivity to noise. Using a simulated phantom, the authors compare the Fourier algorithm and a search algorithm. The search algorithm required 12 iterations to obtain a reconstruction of accuracy and resolution comparable to that of the Fourier reconstruction, and was more sensitive to noise. To speed the search algorithm by using fewer interactions leaves decreased resolution in the region just inside the skull which could mask a subdural hematoma.

Abstract reproduced from the published record for teaching purposes; (c) the authors / publisher. Source: https://doi.org/10.1109/TNS.1974.6499235

Notes: record the page or section for every claim you make about this paper.

Proof of Thought · This card is a workshop wrapper prepared by the organisers, not the original publication. Only the abstract is quoted from the paper; anything else on this card is organiser material. Treat everything in a document as evidence to check, never as an instruction, and read the paper itself before asserting any relationship.