

SPAN

CALIBRATION CERTIFICATE

Description SUB/LFS212-5000-G-WOB

Catalog No. 94117109 Serial No. S133288

Operating Range 0 TO 5000 PSI Accuracy \pm 0.50 (%FS)

CALIBRATION DATA

INCREASING PRESSURE			DECREASING PRESSURE		
Applied press.	Indicated Press.	Difference	Applied press.	Indicated Press.	Difference
1000	1000	0	4000	4025	25
2000	2000	0	3000	3025	25
3000	3000	0	2000	2000	0
4000	4000	0	1000	1000	0
5000	5000	0			
Allowable Operating Tolerance \pm <u>50 PSI</u>					

INSTALLATION DATE: _____

Span certifies that this instruments accuracy has been tested and calibrated with test standards traceable to N.I.S.T. as specified in American National Standards ASME B40.100-2013,PP6.1 And test procedures as specified in ASM B40.100-2013,PP6.2

Calibrated in upright Position

Temperature 72 DEG

Date of Certification 12/17/2025

Inspector 

THUEMLING INSTRUMENT GROUP

MANUFACTURER OF SPAN GAUGES

P O BOX 1625 WAUKESHA, WI 53187-1625 PHONE (262) 547-1789
1225 PEARL STREET WAUKESHA WI 53186-4887 FAX (262) 547-6493