

SPAN

CALIBRATION CERTIFICATE

Description SUB/LFS212-1000-G-CERT

Catalog No. 94112664 Serial No. S133294

Operating Range 0 TO 1000 PSI Accuracy \pm 0.50 (%FS)


CALIBRATION DATA

INCREASING PRESSURE			DECREASING PRESSURE		
Applied press.	Indicated Press.	Difference	Applied press.	Indicated Press.	Difference
100	95	-5	700	702	2
300	295	-5	500	502	2
500	495	-5	300	302	2
700	695	-5	100	102	2
1000	995	-5			
Allowable Operating Tolerance +/- <u>10 PSI</u>					

INSTALLATION DATE: _____

Span certifies that this instruments accuracy has been tested and calibrated with test standards traceable to N.I.S.T. as specified in American National Standards ASME B40.100-2013,PP6.1 And test procedures as specified in ASM B40.100-2013,PP6.2

Calibrated in upright Position Temperature 72 DEG

Date of Certification 12/17/2025 Inspector 

THUEMLING INSTRUMENT GROUP

MANUFACTURER OF SPAN GAUGES
P O BOX 1625 WAUKESHA, WI 53187-1625 PHONE (262) 547-1789
1225 PEARL STREET WAUKESHA WI 53186-4887 FAX (262) 547-6493