

SPAN

CALIBRATION CERTIFICATE

Description SUB/LFS212-3000-G-WOB

Catalog No. 94115160 Serial No. S133286

Operating Range 0 TO 3000 PSI Accuracy \pm 0.67 (%FS)

CALIBRATION DATA

INCREASING PRESSURE			DECREASING PRESSURE		
Applied press.	Indicated Press.	Difference	Applied press.	Indicated Press.	Difference
500	500	0	2000	2020	20
1000	1000	0	1500	1510	10
1500	1490	-10	1000	1000	0
2000	2000	0	500	500	0
3000	3000	0			
Allowable Operating Tolerance \pm 30 PSI					

INSTALLATION DATE: _____

Span certifies that this instruments accuracy has been tested and calibrated with test standards traceable to N.I.S.T. as specified in American National Standards ASME B40.100-2013,PP6.1 And test procedures as specified in ASM B40.100-2013,PP6.2

Calibrated in upright Position

Temperature 72 DEG

Date of Certification 12/17/2025

Inspector 

THUEMLING INSTRUMENT GROUP

MANUFACTURER OF SPAN GAUGES

P O BOX 1625 WAUKESHA, WI 53187-1625 PHONE (262) 547-1789
1225 PEARL STREET WAUKESHA WI 53186-4887 FAX (262) 547-6493