

SPAN

CALIBRATION CERTIFICATE

Description SUB/LFS212-3000-G-WOB

Catalog No. 94115160 Serial No. S133287

Operating Range 0 TO 3000 PSI Accuracy \pm 1.00 (%FS)

CALIBRATION DATA

| INCREASING PRESSURE | | | DECREASING PRESSURE | | |
|--|------------------|------------|---------------------|------------------|------------|
| Applied press. | Indicated Press. | Difference | Applied press. | Indicated Press. | Difference |
| 500 | 470 | -30 | 2000 | 2000 | 0 |
| 1000 | 970 | -30 | 1500 | 1500 | 0 |
| 1500 | 1470 | -30 | 1000 | 990 | -10 |
| 2000 | 2000 | 0 | 500 | 480 | -20 |
| 3000 | 3000 | 0 | | | |
| | | | | | |
| Allowable Operating Tolerance \pm 30 PSI | | | | | |

INSTALLATION DATE: _____

Span certifies that this instruments accuracy has been tested and calibrated with test standards traceable to N.I.S.T. as specified in American National Standards ASME B40.100-2013,PP6.1 And test procedures as specified in ASM B40.100-2013,PP6.2

Calibrated in upright Position

Temperature 72 DEG

Date of Certification 12/17/2025

Inspector 

THUEMLING INSTRUMENT GROUP

MANUFACTURER OF SPAN GAUGES

P O BOX 1625 WAUKESHA, WI 53187-1625 PHONE (262) 547-1789
1225 PEARL STREET WAUKESHA WI 53186-4887 FAX (262) 547-6493