Daniel Siragusa

Diagnosis: Epilepsy

Patient ID: **Gender:** Male **Date Of Birth:** Apr 5, 2000

DEVICE

NEUROSTIMULATOR
Percept PC

Serial Number
NPI732770H

Implanted
Sep 6, 2023

Product Application ID:
4201

Product Version:
07.08.00

Core Application ID:
4001

Core Version:
07.08.00

Device Application ID:
4101

Device Version:
07.08.00

Model Number
B35200

Components
Left Chest

Device Name
--

Battery Level
77%

Estimated Battery Remaining
(based on last 7 days)
6 years, 4 months

Feature Information Code
60d

STIMULATION SETTINGS

	Initial	Final
STIMULATION	On	On
CHECK BATTERY REMINDER	Off	Off
CLOSE COMMUNICATION REQUIRED FOR THERAPY OFF	ON	ON

Lead Configuration

◑ Left ANT



Contacts: 0–7
Lead Model: B33005

◐ Right ANT



Contacts: 8–15
Lead Model: B33005

◑ Left ANT



Contacts: 0–7
Lead Model: B33005

◐ Right ANT



Contacts: 8–15
Lead Model: B33005

Group A (Better BrainSense group)

Left ANT



BrainSense: Sensing Only

3.5 mA
90 µs
145 Hz

2a: 1.1 mA
2b: 1.1 mA
2c: 1.1 mA

SENSING SETUP
LFP Frequency: 11.72 Hz

Upper LFP Threshold:	Lower LFP Threshold:
- -	- -

Group A (Better BrainSense group)

Left ANT



BrainSense: Sensing Only

3.5 mA
90 µs
145 Hz

2a: 1.1 mA
2b: 1.1 mA
2c: 1.1 mA

SENSING SETUP
LFP Frequency: 11.72 Hz

Upper LFP Threshold:	Lower LFP Threshold:
- -	- -

Right ANT



BrainSense: Sensing Only

3.5 mA
90 µs
145 Hz

10a: 1.1 mA
10b: 1.1 mA
10c: 1.1 mA

SENSING SETUP
LFP Frequency: 9.77 Hz

Upper LFP Threshold:	Lower LFP Threshold:
- -	- -

Right ANT



BrainSense: Sensing Only

3.5 mA
90 µs
145 Hz

10a: 1.1 mA
10b: 1.1 mA
10c: 1.1 mA

SENSING SETUP
LFP Frequency: 9.77 Hz

Upper LFP Threshold:	Lower LFP Threshold:
- -	- -

Group B (BrainSense group)

Left ANT



BrainSense: Sensing Only

3.5 mA
90 µs
145 Hz

1a: 1.1 mA
1b: 1.1 mA
1c: 1.1 mA

SENSING SETUP
LFP Frequency: 11.72 Hz

Upper LFP Threshold:	Lower LFP Threshold:
- -	- -

Group B (BrainSense group)

Left ANT



BrainSense: Sensing Only

3.5 mA
90 µs
145 Hz

1a: 1.1 mA
1b: 1.1 mA
1c: 1.1 mA

SENSING SETUP
LFP Frequency: 11.72 Hz

Upper LFP Threshold:	Lower LFP Threshold:
- -	- -

Right ANT



BrainSense: Sensing Only

3.5 mA
90 µs
145 Hz

9a: 1.1 mA
9b: 1.1 mA
9c: 1.1 mA

SENSING SETUP
LFP Frequency: 10.74 Hz

Upper LFP Threshold:	Lower LFP Threshold:
- -	- -

Right ANT



BrainSense: Sensing Only

3.5 mA
90 µs
145 Hz

9a: 1.1 mA
9b: 1.1 mA
9c: 1.1 mA

SENSING SETUP
LFP Frequency: 10.74 Hz

Upper LFP Threshold:	Lower LFP Threshold:
- -	- -

Group C

Left ANT



BrainSense: Sensing Only

2.0 mA
90 µs
145 Hz

2a: 0.7 mA
2b: 0.7 mA
2c: 0.7 mA
1a: 0.7 mA
1b: 0.7 mA
1c: 0.7 mA

SENSING SETUP
LFP Frequency: 11.72 Hz

Upper LFP Threshold: - -
Lower LFP Threshold: - -

Group C

Left ANT



BrainSense: Sensing Only

2.0 mA
90 µs
145 Hz

2a: 0.7 mA
2b: 0.7 mA
2c: 0.7 mA
1a: 0.7 mA
1b: 0.7 mA
1c: 0.7 mA

SENSING SETUP
LFP Frequency: 11.72 Hz

Upper LFP Threshold: - -
Lower LFP Threshold: - -

Right ANT



BrainSense: Sensing Only

2.0 mA
90 µs
145 Hz

10a: 0.7 mA
10b: 0.7 mA
10c: 0.7 mA
9a: 0.7 mA
9b: 0.7 mA
9c: 0.7 mA

SENSING SETUP
LFP Frequency: 10.74 Hz

Upper LFP Threshold: - -
Lower LFP Threshold: - -

Right ANT



BrainSense: Sensing Only

2.0 mA
90 µs
145 Hz

10a: 0.7 mA
10b: 0.7 mA
10c: 0.7 mA
9a: 0.7 mA
9b: 0.7 mA
9c: 0.7 mA

SENSING SETUP
LFP Frequency: 10.74 Hz

Upper LFP Threshold: - -
Lower LFP Threshold: - -

Group D

Left ANT



BrainSense: Sensing Only

0.0 mA
90 µs
145 Hz

2a: 0.0 mA
2b: 0.0 mA
2c: 0.0 mA
1a: 0.0 mA
1b: 0.0 mA
1c: 0.0 mA

SENSING SETUP
LFP Frequency: 11.72 Hz

Upper LFP Threshold: - -
Lower LFP Threshold: - -

Group D

Left ANT



BrainSense: Sensing Only

0.0 mA
90 µs
145 Hz

2a: 0.0 mA
2b: 0.0 mA
2c: 0.0 mA
1a: 0.0 mA
1b: 0.0 mA
1c: 0.0 mA

SENSING SETUP
LFP Frequency: 11.72 Hz

Upper LFP Threshold: - -
Lower LFP Threshold: - -

Right ANT



BrainSense: Sensing Only

0.0 mA
90 µs
145 Hz

10a: 0.0 mA
10b: 0.0 mA
10c: 0.0 mA
9a: 0.0 mA
9b: 0.0 mA
9c: 0.0 mA

SENSING SETUP
LFP Frequency: 10.74 Hz

Upper LFP Threshold: - -
Lower LFP Threshold: - -

Right ANT



BrainSense: Sensing Only

0.0 mA
90 µs
145 Hz

10a: 0.0 mA
10b: 0.0 mA
10c: 0.0 mA
9a: 0.0 mA
9b: 0.0 mA
9c: 0.0 mA

SENSING SETUP
LFP Frequency: 10.74 Hz

Upper LFP Threshold: - -
Lower LFP Threshold: - -

PATIENT PROGRAMMER SETTINGS

Initial

Group A (Better BrainSense group)

Left ANT

Mode	Parameter: View Amplitude
Advanced View	

Right ANT

Mode	Parameter: View Amplitude
Advanced View	

Group B (BrainSense group)

Left ANT

Mode	Parameter: View Amplitude
Advanced View	

Right ANT

Mode	Parameter: View Amplitude
Advanced View	

Final

Group A (Better BrainSense group)

Left ANT

Mode	Parameter: View Amplitude
Advanced View	

Right ANT

Mode	Parameter: View Amplitude
Advanced View	

Group B (BrainSense group)

Left ANT

Mode	Parameter: View Amplitude
Advanced View	

Right ANT

Mode	Parameter: View Amplitude
Advanced View	

Group C

Left ANT

Mode	Parameter: View Amplitude
Advanced View	

Right ANT

Mode	Parameter: View Amplitude
Advanced View	

Group C

Left ANT

Mode	Parameter: View Amplitude
Advanced View	

Right ANT

Mode	Parameter: View Amplitude
Advanced View	

Group D

Left ANT

Mode	Parameter: View Amplitude
Advanced View	

Right ANT

Mode	Parameter: View Amplitude
Advanced View	

Group D

Left ANT

Mode	Parameter: View Amplitude
Advanced View	

Right ANT

Mode	Parameter: View Amplitude
Advanced View	

GROUP SETTINGS

Initial

Group A (Better BrainSense group)

SoftStart/Stop: 8 Seconds

Cycling: On

Cycling On Duration: 1 Minute

Cycling Off Duration: 5 Minutes

Left ANT

SENSING SETUP
High Pass Filter: 1 Hz
Averaging Duration: 3.0 secs
Sensing Blanking Duration: 1.9 ms

Right ANT

SENSING SETUP
High Pass Filter: 1 Hz
Averaging Duration: 3.0 secs
Sensing Blanking Duration: 1.9 ms

Group B (BrainSense group)

SoftStart/Stop: 8 Seconds

Cycling: On

Cycling On Duration: 1 Minute

Cycling Off Duration: 5 Minutes

Left ANT

SENSING SETUP
High Pass Filter: 1 Hz
Averaging Duration: 3.0 secs
Sensing Blanking Duration: 1.9 ms

Right ANT

SENSING SETUP
High Pass Filter: 1 Hz
Averaging Duration: 3.0 secs
Sensing Blanking Duration: 1.9 ms

Group C

Final

Group A (Better BrainSense group)

SoftStart/Stop: 8 Seconds

Cycling: On

Cycling On Duration: 1 Minute

Cycling Off Duration: 5 Minutes

Left ANT

SENSING SETUP
High Pass Filter: 1 Hz
Averaging Duration: 3.0 secs
Sensing Blanking Duration: 1.9 ms

Right ANT

SENSING SETUP
High Pass Filter: 1 Hz
Averaging Duration: 3.0 secs
Sensing Blanking Duration: 1.9 ms

Group B (BrainSense group)

SoftStart/Stop: 8 Seconds

Cycling: On

Cycling On Duration: 1 Minute

Cycling Off Duration: 5 Minutes

Left ANT

SENSING SETUP
High Pass Filter: 1 Hz
Averaging Duration: 3.0 secs
Sensing Blanking Duration: 1.9 ms

Right ANT

SENSING SETUP
High Pass Filter: 1 Hz
Averaging Duration: 3.0 secs
Sensing Blanking Duration: 1.9 ms

Group C

SoftStart/Stop: 8 Seconds

Cycling: On

Cycling On Duration: 1 Minute

Cycling Off Duration: 5 Minutes

Left ANT

SENSING SETUP
High Pass Filter: 1 Hz
Averaging Duration: 3.0 secs
Sensing Blanking Duration: 1.9 ms

Right ANT

SENSING SETUP
High Pass Filter: 1 Hz
Averaging Duration: 3.0 secs
Sensing Blanking Duration: 1.9 ms

SoftStart/Stop: 8 Seconds

Cycling: On

Cycling On Duration: 1 Minute

Cycling Off Duration: 5 Minutes

Left ANT

SENSING SETUP
High Pass Filter: 1 Hz
Averaging Duration: 3.0 secs
Sensing Blanking Duration: 1.9 ms

Right ANT

SENSING SETUP
High Pass Filter: 1 Hz
Averaging Duration: 3.0 secs
Sensing Blanking Duration: 1.9 ms

Group D

SoftStart/Stop: 8 Seconds

Cycling: Off

Left ANT

SENSING SETUP
High Pass Filter: 1 Hz
Averaging Duration: 3.0 secs
Sensing Blanking Duration: 1.9 ms

Right ANT

SENSING SETUP
High Pass Filter: 1 Hz
Averaging Duration: 3.0 secs
Sensing Blanking Duration: 1.9 ms

Group D

SoftStart/Stop: 8 Seconds

Cycling: Off

Left ANT

SENSING SETUP
High Pass Filter: 1 Hz
Averaging Duration: 3.0 secs
Sensing Blanking Duration: 1.9 ms

Right ANT

SENSING SETUP
High Pass Filter: 1 Hz
Averaging Duration: 3.0 secs
Sensing Blanking Duration: 1.9 ms

SYSTEM INTEGRITY INFORMATION

ELECTRODE IMPEDANCE

Impedance measured at 1.0 mA

Left ANT

MONOPOLAR

3	✓ 2,140
2c	✓ 2,621
2b	✓ 2,684
2a	✓ 2,913
1c	✓ 2,389
1b	✓ 2,233
1a	✓ 2,398
0	✓ 1,314

BIPOLAR

	0	1a	1b	1c	2a	2b	2c
3	✓ 3,170	✓ 4,167	✓ 4,041	✓ 4,236	✓ 4,534	✓ 4,410	✓ 4,366
2c	✓ 3,624	✓ 4,446	✓ 4,321	✓ 4,384	✓ 4,402	✓ 4,189	
2b	✓ 3,686	✓ 4,531	✓ 4,325	✓ 4,623	✓ 4,380		
2a	✓ 3,912	✓ 4,660	✓ 4,627	✓ 4,907			
1c	✓ 3,284	✓ 3,981	✓ 3,786				
1b	✓ 3,136	✓ 3,826					
1a	✓ 3,317						

Right ANT

MONOPOLAR

11	✓ 1,787
10c	✓ 2,930
10b	✓ 2,918
10a	✓ 2,706
9c	✓ 2,483
9b	✓ 3,005
9a	✓ 2,691
8	✓ 1,723

BIPOLAR

	8	9a	9b	9c	10a	10b	10c
11	✓ 3,101	✓ 3,985	✓ 4,413	✓ 3,749	✓ 3,814	✓ 4,214	✓ 4,166
10c	✓ 4,228	✓ 4,851	✓ 5,325	✓ 4,526	✓ 4,265	✓ 4,752	
10b	✓ 4,243	✓ 4,925	✓ 5,256	✓ 4,664	✓ 4,393		
10a	✓ 3,975	✓ 4,509	✓ 5,088	✓ 4,286			
9c	✓ 3,624	✓ 3,867	✓ 4,239				
9b	✓ 4,225	✓ 4,301					
9a	✓ 3,848						

GROUP HISTORY

Mar 5, 2026

A	BrainSense: Sensing Only 3.5 mA 90 µs 145 Hz	BrainSense: Sensing Only 3.5 mA 90 µs 145 Hz
B	BrainSense: Sensing Only 3.5 mA 90 µs 145 Hz	BrainSense: Sensing Only 3.5 mA 90 µs 145 Hz
C	BrainSense: Sensing Only 2.0 mA 90 µs 145 Hz	BrainSense: Sensing Only 2.0 mA 90 µs 145 Hz
D	--	

Jan 30, 2026

A	BrainSense: Sensing Only 3.5 mA 90 µs 145 Hz	BrainSense: Sensing Only 3.5 mA 90 µs 145 Hz
B	BrainSense: Sensing Only 3.5 mA 90 µs 145 Hz	BrainSense: Sensing Only 3.5 mA 90 µs 145 Hz
C	1.5 mA 90 µs 145 Hz	1.5 mA 90 µs 145 Hz
D	--	

Oct 30, 2025

Jul 25, 2025

A



BrainSense: Sensing Only
3.5 mA
90 µs
145 Hz



BrainSense: Sensing Only
3.5 mA
90 µs
145 Hz

B



BrainSense: Sensing Only
3.5 mA
90 µs
145 Hz



BrainSense: Sensing Only
3.5 mA
90 µs
145 Hz

C

--

D

--

A



BrainSense: Sensing Only
3.5 mA
90 µs
145 Hz



BrainSense: Sensing Only
3.5 mA
90 µs
145 Hz

B



BrainSense: Sensing Only
3.5 mA
90 µs
145 Hz



BrainSense: Sensing Only
3.5 mA
90 µs
145 Hz

C

--

D

--

Apr 18, 2025

A



3.5 mA
90 µs
145 Hz



3.5 mA
90 µs
145 Hz

B



BrainSense:
Sensing Only

3.5 mA
90 µs
145 Hz



BrainSense:
Sensing Only

3.5 mA
90 µs
145 Hz

C

--

D

--

GROUP USAGE



Electrode Survey

No Data Available

Electrode Identifier

No Data Available

PATIENT EVENTS

Initial		Final	
Name:	Feeling Good	Name:	Feeling Good
Additional Behavior:	LFP	Additional Behavior:	LFP
Name:	Clusters	Name:	Clusters
Additional Behavior:	LFP & Cycling Restart	Additional Behavior:	LFP & Cycling Restart
Name:	Seizure	Name:	Seizure
Additional Behavior:	LFP & Cycling Restart	Additional Behavior:	LFP & Cycling Restart
Name:	Aura	Name:	Aura
Additional Behavior:	LFP & Cycling Restart	Additional Behavior:	LFP & Cycling Restart

Event Summary

Group C Left ANT	Mar 05 – Mar 05
Above Threshold	N/A
Between Thresholds	N/A
Below Threshold	N/A
Average Amplitude	0.00

Group C Right ANT	Mar 05 – Mar 05
Above Threshold	N/A
Between Thresholds	N/A
Below Threshold	N/A
Average Amplitude	0.00

END OF REPORT