

# Measurement Report

24.04.2024,10:16 UTC+02:00,created by testo 184 H1



RESULT: ☒ OK ☐ ALARM

## General Information

Title	
E-Mail to	
Sender	
Receiver	
Comment	

## Device Configuration

Type	testo 184 H1 V01.51 2018-11-15	Channels activated	Temperature Humidity
Meas. Interval	1 min	Current Mode	End
Start/Stop	Button/Button	Record Duration	20.7 h
Start Date/Time	23.04.2024 12:58 UTC+02:00	No. of Values	1244
Stop Date/Time	24.04.2024 09:41	Remaining Battery	281 d
Start Delay	0 min	Production Date	26.10.2020
Configured by	23.04.2024 11:58 UTC+02:00	Check Sum	434F7983691D9845DA874C7CDBAD28CC

## Channel Statistics

	Maximum/Time	Minimum/Time	Average	MKT
Temperature	18.8°C 23.04.2024 18:11	14.6°C 23.04.2024 12:58	18.3°C	18.3°C
Humidity	71.2% 23.04.2024 23:49	47.5% 23.04.2024 12:59	66.8%	--

## Alarm Settings & Results

	Limit Value	Type	Allowed Time	Status	No. of Violations	Total Time	Longest Time
Temp.	8.0°C	disable	--	--	--	--	--
Temp.	2.0°C	disable	--	--	--	--	--
Temp.	---°C	disable	--	--	--	--	--
Temp.	---°C	disable	--	--	--	--	--
MKT	---°C	disable	--	--	--	--	--
Hum.	40.0%	disable	--	--	--	--	--
Hum.	20.0%	disable	--	--	--	--	--

## Graph

