

Measurement Report

23.04.2024,09:54 UTC+02:00,created by testo 184 H1



RESULT: ☒ OK ☐ ALARM

General Information

Title	
E-Mail to	
Sender	
Receiver	
Comment	

Device Configuration

Type	testo 184 H1 V01.51 2018-11-15	Channels activated	Temperature Humidity
Meas. Interval	1 min	Current Mode	End
Start/Stop	Button/Button	Record Duration	22.8 h
Start Date/Time	22.04.2024 11:03 UTC+02:00	No. of Values	1369
Stop Date/Time	23.04.2024 09:51	Remaining Battery	282 d
Start Delay	0 min	Production Date	26.10.2020
Configured by	17.04.2024 14:38 UTC+02:00	Check Sum	4E93377F4E7877667D8DCC85866B09E8

Channel Statistics

	Maximum/Time	Minimum/Time	Average	MKT
Temperature	17.1°C 22.04.2024 11:04	15.0°C 23.04.2024 08:43	15.6°C	15.6°C
Humidity	76.0% 22.04.2024 11:14	50.1% 22.04.2024 16:04	66.5%	--

Alarm Settings & Results

	Limit Value	Type	Allowed Time	Status	No. of Violations	Total Time	Longest Time
Temp.	8.0°C	disable	--	--	--	--	--
Temp.	2.0°C	disable	--	--	--	--	--
Temp.	---°C	disable	--	--	--	--	--
Temp.	---°C	disable	--	--	--	--	--
MKT	---°C	disable	--	--	--	--	--
Hum.	40.0%	disable	--	--	--	--	--
Hum.	20.0%	disable	--	--	--	--	--

Graph

