



Tel: +1.805.964.3615 info@fwt.com
 Fax: +1.805.964.3162 www.fwt.com
 330 Kellogg Avenue, Suite D, Goleta, CA 93117 U.S.A.

CERTIFICATE OF CONFORMANCE

We do hereby certify that all of the items listed below were manufactured in full accordance with our established standards and procedures as well as in compliance with every detail of your

Purchase Order : CC on 03/18/2026

We further state that all records of test and inspection reflect acceptable data and that they, along with material traceability records, are on file at our facility and available for review at our facility.

<u>ITEM</u>	<u>MODEL OR PART#</u>	<u>DESCRIPTION</u>
1	FWT-70-40M	Opti-Chromic Dosimeters
		Lot 1189
		Quantity 36 sets
		50 dosimeters/set

Official Signature: *Scott A. Juroso*

Title: Vice President

Date: 18-Mar-2026



FWT-70 Characterization

Batch 1189

Typical Calibration Curve

Dose, kGy	ΔA at 600 nm	ΔA at 656 nm
0.1008	0.592	0.021
0.301	1.652	0.060
0.502	2.169	0.099
3.01		0.530
10.01		1.344

NOTE: This typical calibration curve is provided as a guide to the response of FWT-70 Opti-Chromic Detectors to ionizing radiation. Actual response also depends on the instrumentation used to measure optical densities.

Kate A. Larson
Authorization for Release

2025-08-13
Date