JEFFERSON LAB
RADIATION CONTROL
SUITE 2
111 HADRON DRIVE
NEWPORT NEWS, VA 23606

Received Date / Reported Date	2025-05-02 / 2025-06-17
Page	1 of 24
Analytical Work Order / QC Release	2512100010 / LCA
Copy / Version	0 / 1

The following dose calculation algorithms were used for assessing the reported doses		
Dosimeter Model	Dosimeter Type	Algorithm Version
InLight	Whole Body	00101
CR-39--Thermal	Whole Body	03001

LANDAUER®
LANDAUER, Inc., 2 Science Road
Glenwood, Illinois 60425-1586
landauer.com
Telephone: (708) 755-7000
Facsimile: (708) 755-7016
Customer Service: (800) 323-8830
Technical: (800) 438-3241

Radiation Dosimetry Report

Account : 738211

Participant Number	Name		Dosimeter	Use	Rad. Type	Rad. Quality	Equivalent Dose (mrem) for Periods Shown Below												Inception Date	Serial Number
							Period Shown Below			Quarter to Date			Year to Date			Lifetime to Date				
	ID Number	Birth Date					Whole Body	Lens	Skin	Whole Body	Lens	Skin	Whole Body	Lens	Skin	Whole Body	Lens	Skin		
For Monitoring Period:							2025-01-01 to 2025-06-30			QUARTER 1			2025			LIFETIME				
	Historical Customer Avg Control Dose		L02TN																	
01057			L02TN	WHBODY	*PN		**79665**	**79665**	**79665**										2025/01	XA00877430F
					*P	H	**79115**	**79115**	**79115**											
					*N	T	**30**	**30**	**30**											
					*N	F	**520**	**520**	**520**											
01058			L02TN	WHBODY	*PN		**249900**	**249900**	**249900**										2025/01	XA00076910G
					*P	M	**249660**	**249660**	**249660**											
					*N	T	**20**	**20**	**20**											
					*N	F	**220**	**220**	**220**											
01059			L02TN	WHBODY	*P		**952**	**952**	**920**										2025/01	XA018087194
					*P	M	**952**	**952**	**920**											
					*N	T	**M**	**M**	**M**											
					*N	F	**M**	**M**	**M**											
01060			L02TN	WHBODY	*PN		**265801**	**257229**	**213040**										2025/01	XA007347682
					*P	M	**265721**	**257149**	**212960**											
					*N	T	**M**	**M**	**M**											
					*N	F	**80**	**80**	**80**											
01061			L02TN	WHBODY	*PN		**4434**	**4434**	**4434**										2025/01	XA00647347A
					*P	H	**4334**	**4334**	**4334**											
					*N	T	**M**	**M**	**M**											
					*N	F	**100**	**100**	**100**											
01062			L02TN	WHBODY	*PN		**109590**	**109590**	**109590**										2025/01	XA01047842E
					*P	H	**109380**	**109380**	**109380**											
					*N	T	**10**	**10**	**10**											
					*N	F	**200**	**200**	**200**											

* - Customer average background dose used for control subtraction

This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government.

JEFFERSON LAB
RADIATION CONTROL
SUITE 2
111 HADRON DRIVE
NEWPORT NEWS, VA 23606

Received Date / Reported Date	2025-05-02 / 2025-06-17
Page	2 of 24
Analytical Work Order / QC Release	2512100010 / LCA
Copy / Version	0 / 1

The following dose calculation algorithms were used for assessing the reported doses		
Dosimeter Model	Dosimeter Type	Algorithm Version
InLight	Whole Body	00101
CR-39--Thermal	Whole Body	03001

LANDAUER®
LANDAUER, Inc., 2 Science Road
Glenwood, Illinois 60425-1586
landauer.com
Telephone: (708) 755-7000
Facsimile: (708) 755-7016
Customer Service: (800) 323-8830
Technical: (800) 438-3241

Radiation Dosimetry Report

Account : 738211

Participant Number	Name: ID Number	Name: Birth Date	Dosimeter	Use	Rad. Type	Rad. Quality	Equivalent Dose (mrem) for Periods Shown Below: Period Shown Below: Whole Body	Period Shown Below: Lens	Period Shown Below: Skin	Quarter to Date: Whole Body	Quarter to Date: Lens	Quarter to Date: Skin	Year to Date: Whole Body	Year to Date: Lens	Year to Date: Skin	Lifetime to Date: Whole Body	Lifetime to Date: Lens	Lifetime to Date: Skin	Inception Date	Serial Number
For Monitoring Period:							2025-01-01 to 2025-06-30			QUARTER 1			2025			LIFETIME				
01063			L02TN	WHBODY															2025/01	XA029134893
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													
01064			L02TN	WHBODY	*PN		**252657**	**244512**	**202524**										2025/01	XA00789617W
					*P	M	**252487**	**244342**	**202354**											
					*N	T	**10**	**10**	**10**											
					*N	F	**160**	**160**	**160**											
01065			L02TN	WHBODY	*PBN		**22628**	**21944**	**18421**										2025/01	XA01023226S
					*P		**21188**	**20504**	**9750**											
					*B				**7231**											
					*N	T	**120**	**120**	**120**											
					*N	F	**1320**	**1320**	**1320**											
01066			L02TN	WHBODY	*PN		**6754**	**6754**	**6754**										2025/01	XA04105401V
					*P	H	**6624**	**6624**	**6624**											
					*N	T	**10**	**10**	**10**											
					*N	F	**120**	**120**	**120**											
01067			L02TN	WHBODY															2025/01	XA02896962R
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													
01068			L02TN	WHBODY															2025/01	XA023495184
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													

* - Customer average background dose used for control subtraction

This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government.

JEFFERSON LAB
RADIATION CONTROL
SUITE 2
111 HADRON DRIVE
NEWPORT NEWS, VA 23606

Received Date / Reported Date	2025-05-02 / 2025-06-17
Page	3 of 24
Analytical Work Order / QC Release	2512100010 / LCA
Copy / Version	0 / 1

The following dose calculation algorithms were used for assessing the reported doses		
Dosimeter Model	Dosimeter Type	Algorithm Version
InLight	Whole Body	00101
CR-39--Thermal	Whole Body	03001

LANDAUER®
LANDAUER, Inc., 2 Science Road
Glenwood, Illinois 60425-1586
landauer.com
Telephone: (708) 755-7000
Facsimile: (708) 755-7016
Customer Service: (800) 323-8830
Technical: (800) 438-3241

Radiation Dosimetry Report

Account : 738211

Participant Number	Name: ID Number	Name: Birth Date	Dosimeter	Use	Rad. Type	Rad. Quality	Equivalent Dose (mrem) for Periods Shown Below: Period Shown Below: Whole Body	Period Shown Below: Lens	Period Shown Below: Skin	Quarter to Date: Whole Body	Quarter to Date: Lens	Quarter to Date: Skin	Year to Date: Whole Body	Year to Date: Lens	Year to Date: Skin	Lifetime to Date: Whole Body	Lifetime to Date: Lens	Lifetime to Date: Skin	Inception Date	Serial Number
For Monitoring Period:							2025-01-01 to 2025-06-30			QUARTER 1			2025			LIFETIME				
01069			L02TN	WHBODY	*PN		**147344**	**142594**	**118107**										2025/01	XA00555126J
					*P	M	**147244**	**142494**	**118007**											
					*N	T	**M**	**M**	**M**											
					*N	F	**100**	**100**	**100**											
01070			L02TN	WHBODY	*PN		**154449**	**149470**	**123800**										2025/01	XA00718119H
					*P	M	**154359**	**149380**	**123710**											
					*N	T	**M**	**M**	**M**											
					*N	F	**90**	**90**	**90**											
01071			L02TN	WHBODY	*P		**1053**	**1053**	**1033**										2025/01	XA00647082M
					*P	M	**1053**	**1053**	**1033**											
					*N	T	**M**	**M**	**M**											
					*N	F	**M**	**M**	**M**											
01072			L02TN	WHBODY	*PN		**1201**	**1201**	**1185**										2025/01	XA01817751F
					*P	M	**581**	**581**	**565**											
					*N	T	**50**	**50**	**50**											
					*N	F	**570**	**570**	**570**											
01073			L02TN	WHBODY	*PN		**4826**	**4826**	**4826**										2025/01	XA02528328A
					*P	H	**866**	**866**	**866**											
					*N	T	**290**	**290**	**290**											
					*N	F	**3670**	**3670**	**3670**											
01074			L02TN	WHBODY	*PN		**89571**	**89571**	**89571**										2025/01	XA02543365D
					*P	H	**89221**	**89221**	**89221**											
					*N	T	**20**	**20**	**20**											
					*N	F	**330**	**330**	**330**											

* - Customer average background dose used for control subtraction

This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government.

JEFFERSON LAB
RADIATION CONTROL
SUITE 2
111 HADRON DRIVE
NEWPORT NEWS, VA 23606

Received Date / Reported Date	2025-05-02 / 2025-06-17
Page	4 of 24
Analytical Work Order / QC Release	2512100010 / LCA
Copy / Version	0 / 1

The following dose calculation algorithms were used for assessing the reported doses		
Dosimeter Model	Dosimeter Type	Algorithm Version
InLight	Whole Body	00101
CR-39--Thermal	Whole Body	03001

LANDAUER®
LANDAUER, Inc., 2 Science Road
Glenwood, Illinois 60425-1586
landauer.com
Telephone: (708) 755-7000
Facsimile: (708) 755-7016
Customer Service: (800) 323-8830
Technical: (800) 438-3241

Radiation Dosimetry Report

Account : 738211

Participant Number	Name: ID Number	Name: Birth Date	Dosimeter	Use	Rad. Type	Rad. Quality	Equivalent Dose (mrem) for Periods Shown Below: Period Shown Below: Whole Body	Period Shown Below: Lens	Period Shown Below: Skin	Quarter to Date: Whole Body	Quarter to Date: Lens	Quarter to Date: Skin	Year to Date: Whole Body	Year to Date: Lens	Year to Date: Skin	Lifetime to Date: Whole Body	Lifetime to Date: Lens	Lifetime to Date: Skin	Inception Date	Serial Number
For Monitoring Period:							2025-01-01 to 2025-06-30			QUARTER 1			2025			LIFETIME				
01075			L02TN	WHBODY	*PN		**151054**	**151054**	**151054**										2025/01	XA00010729J
					*P	H	**150634**	**150634**	**150634**											
					*N	T	**30**	**30**	**30**											
					*N	F	**390**	**390**	**390**											
01076			L02TN	WHBODY	*P		**770**	**770**	**732**										2025/01	XA009505478
					*P	M	**770**	**770**	**732**											
					*N	T	**M**	**M**	**M**											
					*N	F	**M**	**M**	**M**											
01077			L02TN	WHBODY	*PN		**307590**	**297696**	**246691**										2025/01	XA000986643
					*P	M	**306710**	**296816**	**245811**											
					*N	T	**100**	**100**	**100**											
					*N	F	**780**	**780**	**780**											
01078			L02TN	WHBODY	*PN		**118657**	**118657**	**118657**										2025/01	XA01879853V
					*P	H	**118037**	**118037**	**118037**											
					*N	T	**90**	**90**	**90**											
					*N	F	**530**	**530**	**530**											
01079			L02TN	WHBODY	*P		**664**	**664**	**645**										2025/01	XA01353630M
					*P	M	**664**	**664**	**645**											
					*N	T	**M**	**M**	**M**											
					*N	F	**M**	**M**	**M**											
01080			L02TN	WHBODY															2025/01	XA02537977U
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													

* - Customer average background dose used for control subtraction

This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government.

JEFFERSON LAB
RADIATION CONTROL
SUITE 2
111 HADRON DRIVE
NEWPORT NEWS, VA 23606

Received Date / Reported Date	2025-05-02 / 2025-06-17
Page	5 of 24
Analytical Work Order / QC Release	2512100010 / LCA
Copy / Version	0 / 1

The following dose calculation algorithms were used for assessing the reported doses		
Dosimeter Model	Dosimeter Type	Algorithm Version
InLight	Whole Body	00101
CR-39--Thermal	Whole Body	03001

LANDAUER®
LANDAUER, Inc., 2 Science Road
Glenwood, Illinois 60425-1586
landauer.com
Telephone: (708) 755-7000
Facsimile: (708) 755-7016
Customer Service: (800) 323-8830
Technical: (800) 438-3241

Radiation Dosimetry Report

Account : 738211

Participant Number	Name: ID Number	Name: Birth Date	Dosimeter	Use	Rad. Type	Rad. Quality	Equivalent Dose (mrem) for Periods Shown Below: Period Shown Below: Whole Body	Period Shown Below: Lens	Period Shown Below: Skin	Quarter to Date: Whole Body	Quarter to Date: Lens	Quarter to Date: Skin	Year to Date: Whole Body	Year to Date: Lens	Year to Date: Skin	Lifetime to Date: Whole Body	Lifetime to Date: Lens	Lifetime to Date: Skin	Inception Date	Serial Number
For Monitoring Period:							2025-01-01 to 2025-06-30			QUARTER 1			2025			LIFETIME				
01081			L02TN	WHBODY	*PN		**118137**	**118137**	**118137**										2025/01	XA00813523O
					*P	H	**117207**	**117207**	**117207**											
					*N	T	**80**	**80**	**80**											
					*N	F	**850**	**850**	**850**											
01082			L02TN	WHBODY															2025/01	XA00916631L
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													
01083			L02TN	WHBODY															2025/01	XA02726405G
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													
01084			L02TN	WHBODY															2025/01	XA007949602
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													
01085			L02TN	WHBODY	*P		**744**	**744**	**711**										2025/01	XA017860088
					*P	M	**744**	**744**	**711**											
					*N	T	**M**	**M**	**M**											
					*N	F	**M**	**M**	**M**											
01086			L02TN	WHBODY	*PN		**189939**	**183814**	**152241**										2025/01	XA01824418C
					*P	M	**189859**	**183734**	**152161**											
					*N	T	**M**	**M**	**M**											
					*N	F	**80**	**80**	**80**											

* - Customer average background dose used for control subtraction

This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government.

JEFFERSON LAB
RADIATION CONTROL
SUITE 2
111 HADRON DRIVE
NEWPORT NEWS, VA 23606

Received Date / Reported Date	2025-05-02 / 2025-06-17
Page	6 of 24
Analytical Work Order / QC Release	2512100010 / LCA
Copy / Version	0 / 1

The following dose calculation algorithms were used for assessing the reported doses		
Dosimeter Model	Dosimeter Type	Algorithm Version
InLight	Whole Body	00101
CR-39--Thermal	Whole Body	03001

LANDAUER®
LANDAUER, Inc., 2 Science Road
Glenwood, Illinois 60425-1586
landauer.com
Telephone: (708) 755-7000
Facsimile: (708) 755-7016
Customer Service: (800) 323-8830
Technical: (800) 438-3241

Radiation Dosimetry Report

Account : 738211

Participant Number	Name: ID Number	Name: Birth Date	Dosimeter	Use	Rad. Type	Rad. Quality	Equivalent Dose (mrem) for Periods Shown Below: Period Shown Below: Whole Body	Period Shown Below: Lens	Period Shown Below: Skin	Quarter to Date: Whole Body	Quarter to Date: Lens	Quarter to Date: Skin	Year to Date: Whole Body	Year to Date: Lens	Year to Date: Skin	Lifetime to Date: Whole Body	Lifetime to Date: Lens	Lifetime to Date: Skin	Inception Date	Serial Number
For Monitoring Period:							2025-01-01 to 2025-06-30			QUARTER 1			2025			LIFETIME				
01087			L02TN	WHBODY	*PN		**18098**	**17554**	**14751**										2025/01	XA01660940D
					*P	M	**16858**	**16314**	**13511**											
					*N	T	**90**	**90**	**90**											
					*N	F	**1150**	**1150**	**1150**											
01088			L02TN	WHBODY	*PN		**260267**	**260267**	**260267**										2025/01	XA00806454K
					*P	M	**259867**	**259867**	**259867**											
					*N	T	**20**	**20**	**20**											
					*N	F	**380**	**380**	**380**											
01089			L02TN	WHBODY	*PN		**165708**	**160397**	**133022**										2025/01	XA01435335J
					*P	M	**164618**	**159307**	**131932**											
					*N	T	**110**	**110**	**110**											
					*N	F	**980**	**980**	**980**											
01090			L02TN	WHBODY	*P		**1031**	**1031**	**1047**										2025/01	XA007641878
					*P	M	**1031**	**1031**	**1047**											
					*N	T	**M**	**M**	**M**											
					*N	F	**M**	**M**	**M**											
01091			L02TN	WHBODY															2025/01	XA007493758
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													
01092			L02TN	WHBODY	*PBN		**188715**	**182654**	**151409**										2025/01	XA01285966U
					*P		**187885**	**181824**	**49960**											
					*B				**100619**											
					*N	T	**90**	**90**	**90**											
					*N	F	**740**	**740**	**740**											

* - Customer average background dose used for control subtraction

This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government.

JEFFERSON LAB
RADIATION CONTROL
SUITE 2
111 HADRON DRIVE
NEWPORT NEWS, VA 23606

Received Date / Reported Date	2025-05-02 / 2025-06-17
Page	7 of 24
Analytical Work Order / QC Release	2512100010 / LCA
Copy / Version	0 / 1

The following dose calculation algorithms were used for assessing the reported doses		
Dosimeter Model	Dosimeter Type	Algorithm Version
InLight	Whole Body	00101
CR-39--Thermal	Whole Body	03001

LANDAUER®
LANDAUER, Inc., 2 Science Road
Glenwood, Illinois 60425-1586
landauer.com
Telephone: (708) 755-7000
Facsimile: (708) 755-7016
Customer Service: (800) 323-8830
Technical: (800) 438-3241

Radiation Dosimetry Report

Account : 738211

Participant Number	Name: ID Number	Name: Birth Date	Dosimeter	Use	Rad. Type	Rad. Quality	Equivalent Dose (mrem) for Periods Shown Below: Period Shown Below: Whole Body	Period Shown Below: Lens	Period Shown Below: Skin	Quarter to Date: Whole Body	Quarter to Date: Lens	Quarter to Date: Skin	Year to Date: Whole Body	Year to Date: Lens	Year to Date: Skin	Lifetime to Date: Whole Body	Lifetime to Date: Lens	Lifetime to Date: Skin	Inception Date	Serial Number
For Monitoring Period:							2025-01-01 to 2025-06-30			QUARTER 1			2025			LIFETIME				
01093			L02TN	WHBODY	*PN		**171564**	**171564**	**171564**										2025/01	XA00820416M
					*P	H	**170834**	**170834**	**170834**											
					*N	T	**70**	**70**	**70**											
					*N	F	**660**	**660**	**660**											
01094			L02TN	WHBODY	*PN		**261390**	**252964**	**209528**										2025/01	XA02690629V
					*P	M	**261190**	**252764**	**209328**											
					*N	T	**20**	**20**	**20**											
					*N	F	**180**	**180**	**180**											
01095			L02TN	WHBODY	*PN		**8049**	**8049**	**7836**										2025/01	XA008828037
					*P	M	**7569**	**7569**	**7356**											
					*N	T	**40**	**40**	**40**											
					*N	F	**440**	**440**	**440**											
01096			L02TN	WHBODY	*PBN		**261672**	**253261**	**209902**										2025/01	XA01358214J
					*P		**260732**	**252321**	**19134**											
					*B				**189828**											
					*N	T	**70**	**70**	**70**											
					*N	F	**870**	**870**	**870**											
01097			L02TN	WHBODY															2025/01	XA009557742
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													
01098			L02TN	WHBODY	*PN		**216168**	**209201**	**173284**										2025/01	XA00600179M
					*P	M	**215978**	**209011**	**173094**											
					*N	T	**20**	**20**	**20**											
					*N	F	**170**	**170**	**170**											

* - Customer average background dose used for control subtraction

This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government.

JEFFERSON LAB
RADIATION CONTROL
SUITE 2
111 HADRON DRIVE
NEWPORT NEWS, VA 23606

Received Date / Reported Date	2025-05-02 / 2025-06-17
Page	8 of 24
Analytical Work Order / QC Release	2512100010 / LCA
Copy / Version	0 / 1

The following dose calculation algorithms were used for assessing the reported doses		
Dosimeter Model	Dosimeter Type	Algorithm Version
InLight	Whole Body	00101
CR-39--Thermal	Whole Body	03001

LANDAUER®
LANDAUER, Inc., 2 Science Road
Glenwood, Illinois 60425-1586
landauer.com
Telephone: (708) 755-7000
Facsimile: (708) 755-7016
Customer Service: (800) 323-8830
Technical: (800) 438-3241

Radiation Dosimetry Report

Account : 738211

Participant Number	Name: ID Number	Name: Birth Date	Dosimeter	Use	Rad. Type	Rad. Quality	Equivalent Dose (mrem) for Periods Shown Below: Period Shown Below: Whole Body	Period Shown Below: Lens	Period Shown Below: Skin	Quarter to Date: Whole Body	Quarter to Date: Lens	Quarter to Date: Skin	Year to Date: Whole Body	Year to Date: Lens	Year to Date: Skin	Lifetime to Date: Whole Body	Lifetime to Date: Lens	Lifetime to Date: Skin	Inception Date	Serial Number
For Monitoring Period:							2025-01-01 to 2025-06-30			QUARTER 1			2025			LIFETIME				
01099			L02TN	WHBODY	*PN		**91290**	**91290**	**91290**										2025/01	XA00742673D
					*P	H	**90570**	**90570**	**90570**											
					*N	T	**90**	**90**	**90**											
					*N	F	**630**	**630**	**630**											
01100			L02TN	WHBODY	*PN		**17538**	**17538**	**17203**										2025/01	XA01033444O
					*P	M	**17018**	**17018**	**16683**											
					*N	T	**60**	**60**	**60**											
					*N	F	**460**	**460**	**460**											
01101			L02TN	WHBODY	*PN		**10864**	**10864**	**10864**										2025/01	XA012973456
					*P	H	**10614**	**10614**	**10614**											
					*N	T	**20**	**20**	**20**											
					*N	F	**230**	**230**	**230**											
01102			L02TN	WHBODY	*PN		**112274**	**112274**	**112274**										2025/01	XA02501385L
					*P	H	**111924**	**111924**	**111924**											
					*N	T	**20**	**20**	**20**											
					*N	F	**330**	**330**	**330**											
01103			L02TN	WHBODY	*PN		**208380**	**208380**	**208380**										2025/01	XA00096391E
					*P	M	**207700**	**207700**	**207700**											
					*N	T	**100**	**100**	**100**											
					*N	F	**580**	**580**	**580**											
01104			L02TN	WHBODY	*PN		**141706**	**141706**	**141706**										2025/01	XA00763950A
					*P	H	**141546**	**141546**	**141546**											
					*N	T	**10**	**10**	**10**											
					*N	F	**150**	**150**	**150**											

* - Customer average background dose used for control subtraction

This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government.

JEFFERSON LAB
RADIATION CONTROL
SUITE 2
111 HADRON DRIVE
NEWPORT NEWS, VA 23606

Received Date / Reported Date	2025-05-02 / 2025-06-17
Page	9 of 24
Analytical Work Order / QC Release	2512100010 / LCA
Copy / Version	0 / 1

The following dose calculation algorithms were used for assessing the reported doses		
Dosimeter Model	Dosimeter Type	Algorithm Version
InLight	Whole Body	00101
CR-39--Thermal	Whole Body	03001

LANDAUER®
LANDAUER, Inc., 2 Science Road
Glenwood, Illinois 60425-1586
landauer.com
Telephone: (708) 755-7000
Facsimile: (708) 755-7016
Customer Service: (800) 323-8830
Technical: (800) 438-3241

Radiation Dosimetry Report

Account : 738211

Participant Number	Name: ID Number	Name: Birth Date	Dosimeter	Use	Rad. Type	Rad. Quality	Equivalent Dose (mrem) for Periods Shown Below: Period Shown Below: Whole Body	Period Shown Below: Lens	Period Shown Below: Skin	Quarter to Date: Whole Body	Quarter to Date: Lens	Quarter to Date: Skin	Year to Date: Whole Body	Year to Date: Lens	Year to Date: Skin	Lifetime to Date: Whole Body	Lifetime to Date: Lens	Lifetime to Date: Skin	Inception Date	Serial Number
For Monitoring Period:							2025-01-01 to 2025-06-30			QUARTER 1			2025			LIFETIME				
01105			L02TN	WHBODY	*PBN		**4200**	**4200**	**4915**										2025/01	XA00804421V
					*P		**3680**	**3680**	**3660**											
					*B				**735**											
					*N	T	**70**	**70**	**70**											
					*N	F	**450**	**450**	**450**											
01106			L02TN	WHBODY	*PN		**244954**	**244954**	**244954**										2025/01	XA024650886
					*P	M	**240484**	**240484**	**240484**											
					*N	T	**330**	**330**	**330**											
					*N	F	**4140**	**4140**	**4140**											
01107			L02TN	WHBODY	*PBN		**141509**	**137464**	**116614**										2025/01	XA007813493
					*P		**125379**	**121334**	**24007**											
					*B				**76477**											
					*N	T	**1040**	**1040**	**1040**											
					*N	F	**15090**	**15090**	**15090**											
01108			L02TN	WHBODY	*PN		**158921**	**158921**	**158921**										2025/01	XA00762050T
					*P	H	**151481**	**151481**	**151481**											
					*N	T	**510**	**510**	**510**											
					*N	F	**6930**	**6930**	**6930**											
01109			L02TN	WHBODY															2025/01	XA00136861I
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													
01110			L02TN	WHBODY															2025/01	XA00921831K
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													

* - Customer average background dose used for control subtraction

This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government.

JEFFERSON LAB
RADIATION CONTROL
SUITE 2
111 HADRON DRIVE
NEWPORT NEWS, VA 23606

Received Date / Reported Date	2025-05-02 / 2025-06-17
Page	10 of 24
Analytical Work Order / QC Release	2512100010 / LCA
Copy / Version	0 / 1

The following dose calculation algorithms were used for assessing the reported doses		
Dosimeter Model	Dosimeter Type	Algorithm Version
InLight	Whole Body	00101
CR-39--Thermal	Whole Body	03001

LANDAUER®
LANDAUER, Inc., 2 Science Road
Glenwood, Illinois 60425-1586
landauer.com
Telephone: (708) 755-7000
Facsimile: (708) 755-7016
Customer Service: (800) 323-8830
Technical: (800) 438-3241

Radiation Dosimetry Report

Account : 738211

Participant Number	Name: ID Number	Name: Birth Date	Dosimeter	Use	Rad. Type	Rad. Quality	Equivalent Dose (mrem) for Periods Shown Below: Period Shown Below: Whole Body	Period Shown Below: Lens	Period Shown Below: Skin	Quarter to Date: Whole Body	Quarter to Date: Lens	Quarter to Date: Skin	Year to Date: Whole Body	Year to Date: Lens	Year to Date: Skin	Lifetime to Date: Whole Body	Lifetime to Date: Lens	Lifetime to Date: Skin	Inception Date	Serial Number
For Monitoring Period:							2025-01-01 to 2025-06-30			QUARTER 1			2025			LIFETIME				
01111			L02TN	WHBODY															2025/01	XA00552203S
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													
01112			L02TN	WHBODY															2025/01	XA00137873C
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													
01113			L02TN	WHBODY															2025/01	XA02731744B
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													
01114			L02TN	WHBODY	*PN		**35603**	**35603**	**35603**										2025/01	XA01974937R
					*P	H	**34213**	**34213**	**34213**											
					*N	T	**130**	**130**	**130**											
					*N	F	**1260**	**1260**	**1260**											
01115			L02TN	WHBODY	*P	H	**509803**	**509803**	**509803**										2025/01	XA01407363O
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													
01116			L02TN	WHBODY	*PBN		**251734**	**251734**	**251734**										2025/01	XA00636841F
					*P		**245784**	**245784**	**170491**											
					*B				**75293**											
					*N	T	**520**	**520**	**520**											
					*N	F	**5430**	**5430**	**5430**											

* - Customer average background dose used for control subtraction

This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government.

JEFFERSON LAB
RADIATION CONTROL
SUITE 2
111 HADRON DRIVE
NEWPORT NEWS, VA 23606

Received Date / Reported Date	2025-05-02 / 2025-06-17
Page	11 of 24
Analytical Work Order / QC Release	2512100010 / LCA
Copy / Version	0 / 1

The following dose calculation algorithms were used for assessing the reported doses		
Dosimeter Model	Dosimeter Type	Algorithm Version
InLight	Whole Body	00101
CR-39--Thermal	Whole Body	03001

LANDAUER®
LANDAUER, Inc., 2 Science Road
Glenwood, Illinois 60425-1586
landauer.com
Telephone: (708) 755-7000
Facsimile: (708) 755-7016
Customer Service: (800) 323-8830
Technical: (800) 438-3241

Radiation Dosimetry Report

Account : 738211

Participant Number	Name: ID Number	Name: Birth Date	Dosimeter	Use	Rad. Type	Rad. Quality	Equivalent Dose (mrem) for Periods Shown Below: Period Shown Below: Whole Body	Period Shown Below: Lens	Period Shown Below: Skin	Quarter to Date: Whole Body	Quarter to Date: Lens	Quarter to Date: Skin	Year to Date: Whole Body	Year to Date: Lens	Year to Date: Skin	Lifetime to Date: Whole Body	Lifetime to Date: Lens	Lifetime to Date: Skin	Inception Date	Serial Number
For Monitoring Period:							2025-01-01 to 2025-06-30			QUARTER 1			2025			LIFETIME				
01117			L02TN	WHBODY	*PBN		**5896**	**6361**	**6928**										2025/01	XA00850422N
					*P		**5346**	**5811**	**5346**											
					*B				**1032**											
					*N	T	**50**	**50**	**50**											
					*N	F	**500**	**500**	**500**											
01118			L02TN	WHBODY	*PN		**202953**	**202953**	**202953**										2025/01	XA009537950
					*P	H	**201063**	**201063**	**201063**											
					*N	T	**110**	**110**	**110**											
					*N	F	**1780**	**1780**	**1780**											
01119			L02TN	WHBODY	*PN		**46895**	**46895**	**46895**										2025/01	XA00740882C
					*P	H	**45025**	**45025**	**45025**											
					*N	T	**170**	**170**	**170**											
					*N	F	**1700**	**1700**	**1700**											
01120			L02TN	WHBODY	*PN		**29209**	**29209**	**29209**										2025/01	XA028857800
					*P	M	**28669**	**28669**	**28669**											
					*N	T	**50**	**50**	**50**											
					*N	F	**490**	**490**	**490**											
01121			L02TN	WHBODY															2025/01	XA00008746H
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													
01122			L02TN	WHBODY															2025/01	XA01795839M
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													

* - Customer average background dose used for control subtraction

This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government.

JEFFERSON LAB
RADIATION CONTROL
SUITE 2
111 HADRON DRIVE
NEWPORT NEWS, VA 23606

Received Date / Reported Date	2025-05-02 / 2025-06-17
Page	12 of 24
Analytical Work Order / QC Release	2512100010 / LCA
Copy / Version	0 / 1

The following dose calculation algorithms were used for assessing the reported doses		
Dosimeter Model	Dosimeter Type	Algorithm Version
InLight	Whole Body	00101
CR-39--Thermal	Whole Body	03001

LANDAUER®
LANDAUER, Inc., 2 Science Road
Glenwood, Illinois 60425-1586
landauer.com
Telephone: (708) 755-7000
Facsimile: (708) 755-7016
Customer Service: (800) 323-8830
Technical: (800) 438-3241

Radiation Dosimetry Report

Account : 738211

Participant Number	Name: ID Number	Name: Birth Date	Dosimeter	Use	Rad. Type	Rad. Quality	Equivalent Dose (mrem) for Periods Shown Below: Period Shown Below: Whole Body	Period Shown Below: Lens	Period Shown Below: Skin	Quarter to Date: Whole Body	Quarter to Date: Lens	Quarter to Date: Skin	Year to Date: Whole Body	Year to Date: Lens	Year to Date: Skin	Lifetime to Date: Whole Body	Lifetime to Date: Lens	Lifetime to Date: Skin	Inception Date	Serial Number
For Monitoring Period:							2025-01-01 to 2025-06-30			QUARTER 1			2025			LIFETIME				
01123			L02TN	WHBODY	*PBN		**107030**	**103724**	**86683**										2025/01	XA02011930R
					*P		**102470**	**99164**	**36438**											
					*B				**45685**											
					*N	T	**290**	**290**	**290**											
					*N	F	**4270**	**4270**	**4270**											
01124			L02TN	WHBODY															2025/01	XA012956452
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													
01125			L02TN	WHBODY	*PN		**103883**	**103883**	**103883**										2025/01	XA00581474A
					*P	H	**102243**	**102243**	**102243**											
					*N	T	**140**	**140**	**140**											
					*N	F	**1500**	**1500**	**1500**											
01126			L02TN	WHBODY															2025/01	XA024418185
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													
01127			L02TN	WHBODY	*PN		**42175**	**42175**	**42175**										2025/01	XA00803297H
					*P	M	**38155**	**38155**	**38155**											
					*N	T	**270**	**270**	**270**											
					*N	F	**3750**	**3750**	**3750**											
01128			L02TN	WHBODY	*PN		**7882**	**7882**	**7882**										2025/01	XA00772759V
					*P	H	**1432**	**1432**	**1432**											
					*N	T	**480**	**480**	**480**											
					*N	F	**5970**	**5970**	**5970**											

* - Customer average background dose used for control subtraction

This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government.

JEFFERSON LAB
RADIATION CONTROL
SUITE 2
111 HADRON DRIVE
NEWPORT NEWS, VA 23606

Received Date / Reported Date	2025-05-02 / 2025-06-17
Page	13 of 24
Analytical Work Order / QC Release	2512100010 / LCA
Copy / Version	0 / 1

The following dose calculation algorithms were used for assessing the reported doses		
Dosimeter Model	Dosimeter Type	Algorithm Version
InLight	Whole Body	00101
CR-39--Thermal	Whole Body	03001

LANDAUER®
LANDAUER, Inc., 2 Science Road
Glenwood, Illinois 60425-1586
landauer.com
Telephone: (708) 755-7000
Facsimile: (708) 755-7016
Customer Service: (800) 323-8830
Technical: (800) 438-3241

Radiation Dosimetry Report

Account : 738211

Participant Number	Name: ID Number	Name: Birth Date	Dosimeter	Use	Rad. Type	Rad. Quality	Equivalent Dose (mrem) for Periods Shown Below: Period Shown Below: Whole Body	Period Shown Below: Lens	Period Shown Below: Skin	Quarter to Date: Whole Body	Quarter to Date: Lens	Quarter to Date: Skin	Year to Date: Whole Body	Year to Date: Lens	Year to Date: Skin	Lifetime to Date: Whole Body	Lifetime to Date: Lens	Lifetime to Date: Skin	Inception Date	Serial Number
For Monitoring Period:							2025-01-01 to 2025-06-30			QUARTER 1			2025			LIFETIME				
01129			L02TN	WHBODY	*PBN		**84167**	**81497**	**67733**										2025/01	XA007799403
					*P		**82767**	**80097**	**39865**											
					*B				**26468**											
					*N	T	**120**	**120**	**120**											
					*N	F	**1280**	**1280**	**1280**											
01130			L02TN	WHBODY	*PBN		**263540**	**255089**	**255881**										2025/01	XA02887669K
					*P		**261990**	**253539**	**M**											
					*B				**254331**											
					*N	T	**120**	**120**	**120**											
					*N	F	**1430**	**1430**	**1430**											
01131			L02TN	WHBODY															2025/01	XA02581351G
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													
01132			L02TN	WHBODY	*PN		**63603**	**63603**	**63603**										2025/01	XA009185684
					*P	M	**58703**	**58703**	**58703**											
					*N	T	**270**	**270**	**270**											
					*N	F	**4630**	**4630**	**4630**											
01133			L02TN	WHBODY	*PN		**574139**	**574139**	**574139**										2025/01	XA023098847
					*P	H	**569539**	**569539**	**569539**											
					*N	T	**380**	**380**	**380**											
					*N	F	**4220**	**4220**	**4220**											
01134			L02TN	WHBODY	*PN		**67317**	**67317**	**67317**										2025/01	XA007631895
					*P	H	**65177**	**65177**	**65177**											
					*N	T	**160**	**160**	**160**											
					*N	F	**1980**	**1980**	**1980**											

* - Customer average background dose used for control subtraction

This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government.

JEFFERSON LAB
RADIATION CONTROL
SUITE 2
111 HADRON DRIVE
NEWPORT NEWS, VA 23606

Received Date / Reported Date	2025-05-02 / 2025-06-17
Page	14 of 24
Analytical Work Order / QC Release	2512100010 / LCA
Copy / Version	0 / 1

The following dose calculation algorithms were used for assessing the reported doses		
Dosimeter Model	Dosimeter Type	Algorithm Version
InLight	Whole Body	00101
CR-39--Thermal	Whole Body	03001

LANDAUER®
LANDAUER, Inc., 2 Science Road
Glenwood, Illinois 60425-1586
landauer.com
Telephone: (708) 755-7000
Facsimile: (708) 755-7016
Customer Service: (800) 323-8830
Technical: (800) 438-3241

Radiation Dosimetry Report

Account : 738211

Participant Number	Name: ID Number	Name: Birth Date	Dosimeter	Use	Rad. Type	Rad. Quality	Equivalent Dose (mrem) for Periods Shown Below: Period Shown Below: Whole Body	Period Shown Below: Lens	Period Shown Below: Skin	Quarter to Date: Whole Body	Quarter to Date: Lens	Quarter to Date: Skin	Year to Date: Whole Body	Year to Date: Lens	Year to Date: Skin	Lifetime to Date: Whole Body	Lifetime to Date: Lens	Lifetime to Date: Skin	Inception Date	Serial Number
For Monitoring Period:							2025-01-01 to 2025-06-30			QUARTER 1			2025			LIFETIME				
01135			L02TN	WHBODY	*PN		**156949**	**156949**	**156949**										2025/01	XA02388763Y
					*P	H	**153559**	**153559**	**153559**											
					*N	T	**200**	**200**	**200**											
					*N	F	**3190**	**3190**	**3190**											
01136			L02TN	WHBODY	*PBN		**37643**	**36467**	**30401**										2025/01	XA007780783
					*P		**36473**	**35297**	**15994**											
					*B				**13237**											
					*N	T	**90**	**90**	**90**											
					*N	F	**1080**	**1080**	**1080**											
01137			L02TN	WHBODY	*PN		**96636**	**96636**	**96636**										2025/01	XA028926225
					*P	H	**87356**	**87356**	**87356**											
					*N	T	**660**	**660**	**660**											
					*N	F	**8620**	**8620**	**8620**											
01138			L02TN	WHBODY	*P	H	**6224**	**6224**	**6224**										2025/01	XA008493666
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													
01139			L02TN	WHBODY	*PN		**170507**	**170507**	**170507**										2025/01	XA005799215
					*P	H	**168277**	**168277**	**168277**											
					*N	T	**230**	**230**	**230**											
					*N	F	**2000**	**2000**	**2000**											
01140			L02TN	WHBODY															2025/01	XA041356194
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													

* - Customer average background dose used for control subtraction

This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government.

JEFFERSON LAB
RADIATION CONTROL
SUITE 2
111 HADRON DRIVE
NEWPORT NEWS, VA 23606

Received Date / Reported Date	2025-05-02 / 2025-06-17
Page	15 of 24
Analytical Work Order / QC Release	2512100010 / LCA
Copy / Version	0 / 1

The following dose calculation algorithms were used for assessing the reported doses		
Dosimeter Model	Dosimeter Type	Algorithm Version
InLight	Whole Body	00101
CR-39--Thermal	Whole Body	03001

LANDAUER®
LANDAUER, Inc., 2 Science Road
Glenwood, Illinois 60425-1586
landauer.com
Telephone: (708) 755-7000
Facsimile: (708) 755-7016
Customer Service: (800) 323-8830
Technical: (800) 438-3241

Radiation Dosimetry Report

Account : 738211

Participant Number	Name: ID Number	Name: Birth Date	Dosimeter	Use	Rad. Type	Rad. Quality	Equivalent Dose (mrem) for Periods Shown Below: Period Shown Below: Whole Body	Period Shown Below: Lens	Period Shown Below: Skin	Quarter to Date: Whole Body	Quarter to Date: Lens	Quarter to Date: Skin	Year to Date: Whole Body	Year to Date: Lens	Year to Date: Skin	Lifetime to Date: Whole Body	Lifetime to Date: Lens	Lifetime to Date: Skin	Inception Date	Serial Number
For Monitoring Period:							2025-01-01 to 2025-06-30			QUARTER 1			2025			LIFETIME				
01141			L02TN	WHBODY	*PN		**209077**	**209077**	**209077**										2025/01	XA02519150S
					*P	M	**199197**	**199197**	**199197**											
					*N	T	**700**	**700**	**700**											
					*N	F	**9180**	**9180**	**9180**											
01142			L02TN	WHBODY															2025/01	XA02110336R
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													
01143			L02TN	WHBODY	*PN		**83982**	**83982**	**83982**										2025/01	XA04105754E
					*P	H	**81902**	**81902**	**81902**											
					*N	T	**160**	**160**	**160**											
					*N	F	**1920**	**1920**	**1920**											
01144			L02TN	WHBODY															2025/01	XA00760960C
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													
01145			L02TN	WHBODY	*PN		**42268**	**42268**	**42268**										2025/01	XA00553485B
					*P	H	**40198**	**40198**	**40198**											
					*N	T	**190**	**190**	**190**											
					*N	F	**1880**	**1880**	**1880**											
01146			L02TN	WHBODY	*PN		**32673**	**32673**	**32673**										2025/01	XA01865818V
					*P	H	**28823**	**28823**	**28823**											
					*N	T	**310**	**310**	**310**											
					*N	F	**3540**	**3540**	**3540**											

* - Customer average background dose used for control subtraction

This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government.

JEFFERSON LAB
RADIATION CONTROL
SUITE 2
111 HADRON DRIVE
NEWPORT NEWS, VA 23606

Received Date / Reported Date	2025-05-02 / 2025-06-17
Page	16 of 24
Analytical Work Order / QC Release	2512100010 / LCA
Copy / Version	0 / 1

The following dose calculation algorithms were used for assessing the reported doses		
Dosimeter Model	Dosimeter Type	Algorithm Version
InLight	Whole Body	00101
CR-39--Thermal	Whole Body	03001

LANDAUER®
LANDAUER, Inc., 2 Science Road
Glenwood, Illinois 60425-1586
landauer.com
Telephone: (708) 755-7000
Facsimile: (708) 755-7016
Customer Service: (800) 323-8830
Technical: (800) 438-3241

Radiation Dosimetry Report

Account : 738211

Participant Number	Name: ID Number	Name: Birth Date	Dosimeter	Use	Rad. Type	Rad. Quality	Equivalent Dose (mrem) for Periods Shown Below: Period Shown Below: Whole Body	Period Shown Below: Lens	Period Shown Below: Skin	Quarter to Date: Whole Body	Quarter to Date: Lens	Quarter to Date: Skin	Year to Date: Whole Body	Year to Date: Lens	Year to Date: Skin	Lifetime to Date: Whole Body	Lifetime to Date: Lens	Lifetime to Date: Skin	Inception Date	Serial Number
For Monitoring Period:							2025-01-01 to 2025-06-30			QUARTER 1			2025			LIFETIME				
01147			L02TN	WHBODY	*PN		**20404**	**20404**	**20404**										2025/01	XA01031838B
					*P	H	**17624**	**17624**	**17624**											
					*N	T	**190**	**190**	**190**											
					*N	F	**2590**	**2590**	**2590**											
01148			L02TN	WHBODY	*PB		**377590**	**365410**	**302617**										2025/01	XA00741395E
					*P		**377590**	**365410**	**2549**											
					*B				**300068**											
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													
01149			L02TN	WHBODY	*PN		**7807**	**7807**	**7807**										2025/01	XA009876142
					*P	H	**1397**	**1397**	**1397**											
					*N	T	**810**	**810**	**810**											
					*N	F	**5600**	**5600**	**5600**											
01150			L02TN	WHBODY	*P	H	**6454**	**6454**	**6454**										2025/01	XA006986621
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													
01151			L02TN	WHBODY															2025/01	XA00120686I
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													
01152			L02TN	WHBODY	*PN		**58432**	**58432**	**58432**										2025/01	XA00743842D
					*P	H	**50292**	**50292**	**50292**											
					*N	T	**520**	**520**	**520**											
					*N	F	**7620**	**7620**	**7620**											

* - Customer average background dose used for control subtraction

This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government.

JEFFERSON LAB
RADIATION CONTROL
SUITE 2
111 HADRON DRIVE
NEWPORT NEWS, VA 23606

Received Date / Reported Date	2025-05-02 / 2025-06-17
Page	17 of 24
Analytical Work Order / QC Release	2512100010 / LCA
Copy / Version	0 / 1

The following dose calculation algorithms were used for assessing the reported doses		
Dosimeter Model	Dosimeter Type	Algorithm Version
InLight	Whole Body	00101
CR-39--Thermal	Whole Body	03001

LANDAUER®
LANDAUER, Inc., 2 Science Road
Glenwood, Illinois 60425-1586
landauer.com
Telephone: (708) 755-7000
Facsimile: (708) 755-7016
Customer Service: (800) 323-8830
Technical: (800) 438-3241

Radiation Dosimetry Report

Account : 738211

Participant Number	Name: ID Number	Name: Birth Date	Dosimeter	Use	Rad. Type	Rad. Quality	Equivalent Dose (mrem) for Periods Shown Below: Period Shown Below: Whole Body	Period Shown Below: Lens	Period Shown Below: Skin	Quarter to Date: Whole Body	Quarter to Date: Lens	Quarter to Date: Skin	Year to Date: Whole Body	Year to Date: Lens	Year to Date: Skin	Lifetime to Date: Whole Body	Lifetime to Date: Lens	Lifetime to Date: Skin	Inception Date	Serial Number
For Monitoring Period:							2025-01-01 to 2025-06-30			QUARTER 1			2025			LIFETIME				
01153			L02TN	WHBODY	*PBN		**136007**	**131756**	**109845**										2025/01	XA00828689X
					*P		**131757**	**127506**	**66391**											
					*B				**39204**											
					*N	T	**230**	**230**	**230**											
					*N	F	**4020**	**4020**	**4020**											
01154			L02TN	WHBODY	*PN		**36536**	**36536**	**36536**										2025/01	XA02895407Y
					*P	H	**30346**	**30346**	**30346**											
					*N	T	**510**	**510**	**510**											
					*N	F	**5680**	**5680**	**5680**											
01155			L02TN	WHBODY															2025/01	XA00578631B
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													
01156			L02TN	WHBODY															2025/01	XA008499820
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													
01157			L02TN	WHBODY	*PN		**33227**	**33227**	**33227**										2025/01	XA01110376P
					*P	H	**27087**	**27087**	**27087**											
					*N	T	**650**	**650**	**650**											
					*N	F	**5490**	**5490**	**5490**											
01158			L02TN	WHBODY	*PB		**110017**	**106468**	**88173**										2025/01	XA008842566
					*P		**110017**	**106468**	**31586**											
					*B				**56587**											
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													

* - Customer average background dose used for control subtraction

This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government.

JEFFERSON LAB
RADIATION CONTROL
SUITE 2
111 HADRON DRIVE
NEWPORT NEWS, VA 23606

Received Date / Reported Date	2025-05-02 / 2025-06-17
Page	18 of 24
Analytical Work Order / QC Release	2512100010 / LCA
Copy / Version	0 / 1

The following dose calculation algorithms were used for assessing the reported doses		
Dosimeter Model	Dosimeter Type	Algorithm Version
InLight	Whole Body	00101
CR-39--Thermal	Whole Body	03001

LANDAUER®
LANDAUER, Inc., 2 Science Road
Glenwood, Illinois 60425-1586
landauer.com
Telephone: (708) 755-7000
Facsimile: (708) 755-7016
Customer Service: (800) 323-8830
Technical: (800) 438-3241

Radiation Dosimetry Report

Account : 738211

Participant Number	Name: ID Number	Name: Birth Date	Dosimeter	Use	Rad. Type	Rad. Quality	Equivalent Dose (mrem) for Periods Shown Below: Period Shown Below: Whole Body	Period Shown Below: Lens	Period Shown Below: Skin	Quarter to Date: Whole Body	Quarter to Date: Lens	Quarter to Date: Skin	Year to Date: Whole Body	Year to Date: Lens	Year to Date: Skin	Lifetime to Date: Whole Body	Lifetime to Date: Lens	Lifetime to Date: Skin	Inception Date	Serial Number
For Monitoring Period:							2025-01-01 to 2025-06-30			QUARTER 1			2025			LIFETIME				
01159			L02TN	WHBODY	*PN		**39403**	**39403**	**39403**										2025/01	XA00377957U
					*P	H	**37813**	**37813**	**37813**											
					*N	T	**140**	**140**	**140**											
					*N	F	**1450**	**1450**	**1450**											
01160			L02TN	WHBODY	*PBN		**190282**	**184191**	**152788**										2025/01	XA00821901M
					*P		**188832**	**182741**	**59895**											
					*B				**91443**											
					*N	T	**100**	**100**	**100**											
					*N	F	**1350**	**1350**	**1350**											
01161			L02TN	WHBODY															2025/01	XA00694018A
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													
01162			L02TN	WHBODY															2025/01	XA02218551M
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													
01163			L02TN	WHBODY	*PN		**85344**	**85344**	**85344**										2025/01	XA02006005Z
					*P	H	**83564**	**83564**	**83564**											
					*N	T	**140**	**140**	**140**											
					*N	F	**1640**	**1640**	**1640**											
01164			L02TN	WHBODY															2025/01	XA04105517G
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													

* - Customer average background dose used for control subtraction

This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government.

JEFFERSON LAB
RADIATION CONTROL
SUITE 2
111 HADRON DRIVE
NEWPORT NEWS, VA 23606

Received Date / Reported Date	2025-05-02 / 2025-06-17
Page	19 of 24
Analytical Work Order / QC Release	2512100010 / LCA
Copy / Version	0 / 1

The following dose calculation algorithms were used for assessing the reported doses

Dosimeter Model	Dosimeter Type	Algorithm Version
InLight	Whole Body	00101
CR-39--Thermal	Whole Body	03001

LANDAUER®
LANDAUER, Inc., 2 Science Road
Glenwood, Illinois 60425-1586
landauer.com
Telephone: (708) 755-7000
Facsimile: (708) 755-7016
Customer Service: (800) 323-8830
Technical: (800) 438-3241

Radiation Dosimetry Report

Account : 738211

Participant Number	Name		Dosimeter	Use	Rad. Type	Rad. Quality	Equivalent Dose (mrem) for Periods Shown Below												Inception Date	Serial Number
							Period Shown Below			Quarter to Date			Year to Date			Lifetime to Date				
	ID Number	Birth Date					Whole Body	Lens	Skin	Whole Body	Lens	Skin	Whole Body	Lens	Skin	Whole Body	Lens	Skin		
For Monitoring Period:							2025-01-01 to 2025-06-30			QUARTER 1			2025			LIFETIME				
01165			L02TN	WHBODY	*PN		**503665**	**503665**	**503665**										2025/01	XA01359057C
					*P	H	**487835**	**487835**	**487835**											
					*N	T	**1070**	**1070**	**1070**											
					*N	F	**14760**	**14760**	**14760**											
01166			L02TN	WHBODY	*PN		**39690**	**39690**	**39690**										2025/01	XA01809651H
					*P	H	**32940**	**32940**	**32940**											
					*N	T	**350**	**350**	**350**											
					*N	F	**6400**	**6400**	**6400**											
01167			L02TN	WHBODY	*PN		**22689**	**22689**	**22689**										2025/01	XA04105396E
					*P	H	**18009**	**18009**	**18009**											
					*N	T	**250**	**250**	**250**											
					*N	F	**4430**	**4430**	**4430**											
01168			L02TN	WHBODY	*PBN		**43723**	**42345**	**35240**										2025/01	XA01431915D
					*P		**42723**	**41345**	**25500**											
					*B				**8740**											
					*N	T	**80**	**80**	**80**											
					*N	F	**920**	**920**	**920**											
01169			L02TN	WHBODY	*PBN		**76056**	**73672**	**61381**										2025/01	XA011990964
					*P		**73906**	**71522**	**22296**											
					*B				**36935**											
					*N	T	**150**	**150**	**150**											
					*N	F	**2000**	**2000**	**2000**											

* - Customer average background dose used for control subtraction

This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government.

JEFFERSON LAB
RADIATION CONTROL
SUITE 2
111 HADRON DRIVE
NEWPORT NEWS, VA 23606

Received Date / Reported Date	2025-05-02 / 2025-06-17
Page	20 of 24
Analytical Work Order / QC Release	2512100010 / LCA
Copy / Version	0 / 1

The following dose calculation algorithms were used for assessing the reported doses		
Dosimeter Model	Dosimeter Type	Algorithm Version
InLight	Whole Body	00101
CR-39--Thermal	Whole Body	03001

LANDAUER®
LANDAUER, Inc., 2 Science Road
Glenwood, Illinois 60425-1586
landauer.com
Telephone: (708) 755-7000
Facsimile: (708) 755-7016
Customer Service: (800) 323-8830
Technical: (800) 438-3241

Radiation Dosimetry Report

Account : 738211

Participant Number	Name: ID Number	Name: Birth Date	Dosimeter	Use	Rad. Type	Rad. Quality	Equivalent Dose (mrem) for Periods Shown Below: Period Shown Below: Whole Body	Period Shown Below: Lens	Period Shown Below: Skin	Quarter to Date: Whole Body	Quarter to Date: Lens	Quarter to Date: Skin	Year to Date: Whole Body	Year to Date: Lens	Year to Date: Skin	Lifetime to Date: Whole Body	Lifetime to Date: Lens	Lifetime to Date: Skin	Inception Date	Serial Number
For Monitoring Period:							2025-01-01 to 2025-06-30			QUARTER 1			2025			LIFETIME				
01170			L02TN	WHBODY	*PBN		**53503**	**53503**	**53503**										2025/01	XA01008795C
					*P		**51413**	**51413**	**41712**											
					*B				**9701**											
					*N	T	**180**	**180**	**180**											
					*N	F	**1910**	**1910**	**1910**											
01171			L02TN	WHBODY															2025/01	XA003089288
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													
01292			L02TN	WHBODY	*PN		**1097**	**1097**	**1090**										2025/01	XA02194729U
					*P	M	**537**	**537**	**530**											
					*N	T	**50**	**50**	**50**											
					*N	F	**510**	**510**	**510**											
01293			L02TN	WHBODY															2025/01	XA012849722
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													
01294			L02TN	WHBODY	*PN		**4217**	**4217**	**4217**										2025/01	XA002967047
					*P	H	**907**	**907**	**907**											
					*N	T	**260**	**260**	**260**											
					*N	F	**3050**	**3050**	**3050**											
01295			L02TN	WHBODY	*PN		**1249**	**1249**	**1238**										2025/01	XA027364559
					*P	M	**749**	**749**	**738**											
					*N	T	**40**	**40**	**40**											
					*N	F	**460**	**460**	**460**											

* - Customer average background dose used for control subtraction

This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government.

JEFFERSON LAB
RADIATION CONTROL
SUITE 2
111 HADRON DRIVE
NEWPORT NEWS, VA 23606

Received Date / Reported Date	2025-05-02 / 2025-06-17
Page	21 of 24
Analytical Work Order / QC Release	2512100010 / LCA
Copy / Version	0 / 1

The following dose calculation algorithms were used for assessing the reported doses		
Dosimeter Model	Dosimeter Type	Algorithm Version
InLight	Whole Body	00101
CR-39--Thermal	Whole Body	03001

LANDAUER®
LANDAUER, Inc., 2 Science Road
Glenwood, Illinois 60425-1586
landauer.com
Telephone: (708) 755-7000
Facsimile: (708) 755-7016
Customer Service: (800) 323-8830
Technical: (800) 438-3241

Radiation Dosimetry Report

Account : 738211

Participant Number	Name: ID Number	Name: Birth Date	Dosimeter	Use	Rad. Type	Rad. Quality	Equivalent Dose (mrem) for Periods Shown Below: Period Shown Below: Whole Body	Period Shown Below: Lens	Period Shown Below: Skin	Quarter to Date: Whole Body	Quarter to Date: Lens	Quarter to Date: Skin	Year to Date: Whole Body	Year to Date: Lens	Year to Date: Skin	Lifetime to Date: Whole Body	Lifetime to Date: Lens	Lifetime to Date: Skin	Inception Date	Serial Number
For Monitoring Period:							2025-01-01 to 2025-06-30			QUARTER 1			2025			LIFETIME				
01300			L02TN	WHBODY	*PN		**10258**	**10258**	**10392**										2025/01	XA02181142N
					*P	M	**9898**	**9898**	**10032**											
					*N	T	**30**	**30**	**30**											
					*N	F	**330**	**330**	**330**											
01301			L02TN	WHBODY	*PN		**9091**	**9091**	**8918**										2025/01	XA02295370D
					*P	M	**9001**	**9001**	**8828**											
					*N	T	**M**	**M**	**M**											
					*N	F	**90**	**90**	**90**											
01302			L02TN	WHBODY															2025/01	XA00103954K
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													
01303			L02TN	WHBODY															2025/01	XA007666157
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													
01304			L02TN	WHBODY	*PN		**7428**	**7428**	**7418**										2025/01	XA00008186P
					*P	M	**868**	**868**	**858**											
					*N	T	**330**	**330**	**330**											
					*N	F	**6230**	**6230**	**6230**											
01305			L02TN	WHBODY	*PN		**773**	**773**	**773**										2025/01	XA02131006V
					*P	H	**493**	**493**	**493**											
					*N	T	**40**	**40**	**40**											
					*N	F	**240**	**240**	**240**											

* - Customer average background dose used for control subtraction

This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government.

JEFFERSON LAB
RADIATION CONTROL
SUITE 2
111 HADRON DRIVE
NEWPORT NEWS, VA 23606

Received Date / Reported Date	2025-05-02 / 2025-06-17
Page	22 of 24
Analytical Work Order / QC Release	2512100010 / LCA
Copy / Version	0 / 1

The following dose calculation algorithms were used for assessing the reported doses		
Dosimeter Model	Dosimeter Type	Algorithm Version
InLight	Whole Body	00101
CR-39--Thermal	Whole Body	03001

LANDAUER®
LANDAUER, Inc., 2 Science Road
Glenwood, Illinois 60425-1586
landauer.com
Telephone: (708) 755-7000
Facsimile: (708) 755-7016
Customer Service: (800) 323-8830
Technical: (800) 438-3241

Radiation Dosimetry Report

Account : 738211

Participant Number	Name: ID Number	Name: Birth Date	Dosimeter	Use	Rad. Type	Rad. Quality	Equivalent Dose (mrem) for Periods Shown Below: Period Shown Below: Whole Body	Period Shown Below: Lens	Period Shown Below: Skin	Quarter to Date: Whole Body	Quarter to Date: Lens	Quarter to Date: Skin	Year to Date: Whole Body	Year to Date: Lens	Year to Date: Skin	Lifetime to Date: Whole Body	Lifetime to Date: Lens	Lifetime to Date: Skin	Inception Date	Serial Number
For Monitoring Period:							2025-01-01 to 2025-06-30			QUARTER 1			2025			LIFETIME				
01306			L02TN	WHBODY	*PN		**1120**	**1120**	**1099**										2025/01	XA017828424
					*P	M	**660**	**660**	**639**											
					*N	T	**60**	**60**	**60**											
					*N	F	**400**	**400**	**400**											
01307			L02TN	WHBODY	*PN		**15939**	**15939**	**15939**										2025/01	XA01273284E
					*P	H	**15719**	**15719**	**15719**											
					*N	T	**20**	**20**	**20**											
					*N	F	**200**	**200**	**200**											
01308			L02TN	WHBODY	*P		**768**	**768**	**729**										2025/01	XA02833164I
					*P	M	**768**	**768**	**729**											
					*N	T	**M**	**M**	**M**											
					*N	F	**M**	**M**	**M**											
01309			L02TN	WHBODY	*PN		**1088**	**1088**	**1078**										2025/01	XA00865041O
					*P	H	**588**	**588**	**578**											
					*N	T	**40**	**40**	**40**											
					*N	F	**460**	**460**	**460**											
01310			L02TN	WHBODY	*P		**1100**	**1100**	**1100**										2025/01	XA00951300T
					*P	H	**1100**	**1100**	**1100**											
					*N	T	**M**	**M**	**M**											
					*N	F	**M**	**M**	**M**											
01311			L02TN	WHBODY															2025/01	XA02799476Q
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													

* - Customer average background dose used for control subtraction

This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government.

JEFFERSON LAB
RADIATION CONTROL
SUITE 2
111 HADRON DRIVE
NEWPORT NEWS, VA 23606

Received Date / Reported Date	2025-05-02 / 2025-06-17
Page	23 of 24
Analytical Work Order / QC Release	2512100010 / LCA
Copy / Version	0 / 1

The following dose calculation algorithms were used for assessing the reported doses

Dosimeter Model	Dosimeter Type	Algorithm Version
InLight	Whole Body	00101
CR-39--Thermal	Whole Body	03001

LANDAUER®
LANDAUER, Inc., 2 Science Road
Glenwood, Illinois 60425-1586
landauer.com
Telephone: (708) 755-7000
Facsimile: (708) 755-7016
Customer Service: (800) 323-8830
Technical: (800) 438-3241

Radiation Dosimetry Report

Account : 738211

Participant Number	Name: ID Number	Name: Birth Date	Dosimeter	Use	Rad. Type	Rad. Quality	Equivalent Dose (mrem) for Periods Shown Below: Period Shown Below: Whole Body	Period Shown Below: Lens	Period Shown Below: Skin	Quarter to Date: Whole Body	Quarter to Date: Lens	Quarter to Date: Skin	Year to Date: Whole Body	Year to Date: Lens	Year to Date: Skin	Lifetime to Date: Whole Body	Lifetime to Date: Lens	Lifetime to Date: Skin	Inception Date	Serial Number
For Monitoring Period:							2025-01-01 to 2025-06-30			QUARTER 1			2025			LIFETIME				
01312			L02TN	WHBODY	*PN		**4051**	**4051**	**4051**										2025/01	XA00636763B
					*P	H	**831**	**831**	**831**											
					*N	T	**280**	**280**	**280**											
					*N	F	**2940**	**2940**	**2940**											
01313			L02TN	WHBODY	*PN		**242086**	**234307**	**194203**										2025/01	XA02065969S
					*P	M	**241156**	**233377**	**193273**											
					*N	T	**80**	**80**	**80**											
					*N	F	**850**	**850**	**850**											
01314			L02TN	WHBODY	*PN		**282209**	**273108**	**226192**										2025/01	XA007814954
					*P	M	**282119**	**273018**	**226102**											
					*N	T	**M**	**M**	**M**											
					*N	F	**90**	**90**	**90**											
01315			L02TN	WHBODY	*P		**1169**	**1169**	**1169**										2025/01	XA022638644
					*P	H	**1169**	**1169**	**1169**											
					*N	T	**M**	**M**	**M**											
					*N	F	**M**	**M**	**M**											
01316			L02TN	WHBODY															2025/01	XA00290482F
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													
01317			L02TN	WHBODY	*P		**875**	**875**	**828**										2025/01	XA012660889
					*P	M	**875**	**875**	**828**											
					*N	T	**M**	**M**	**M**											
					*N	F	**M**	**M**	**M**											

* - Customer average background dose used for control subtraction

This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government.

JEFFERSON LAB
RADIATION CONTROL
SUITE 2
111 HADRON DRIVE
NEWPORT NEWS, VA 23606

Received Date / Reported Date	2025-05-02 / 2025-06-17
Page	24 of 24
Analytical Work Order / QC Release	2512100010 / LCA
Copy / Version	0 / 1

The following dose calculation algorithms were used for assessing the reported doses

Dosimeter Model	Dosimeter Type	Algorithm Version
InLight	Whole Body	00101
CR-39--Thermal	Whole Body	03001

LANDAUER®
LANDAUER, Inc., 2 Science Road
Glenwood, Illinois 60425-1586
landauer.com
Telephone: (708) 755-7000
Facsimile: (708) 755-7016
Customer Service: (800) 323-8830
Technical: (800) 438-3241

Radiation Dosimetry Report

Account : 738211

Participant Number	Name		Dosimeter	Use	Rad. Type	Rad. Quality	Equivalent Dose (mrem) for Periods Shown Below												Inception Date	Serial Number
							Period Shown Below			Quarter to Date			Year to Date			Lifetime to Date				
	ID Number	Birth Date					Whole Body	Lens	Skin	Whole Body	Lens	Skin	Whole Body	Lens	Skin	Whole Body	Lens	Skin		
For Monitoring Period:							2025-01-01 to 2025-06-30			QUARTER 1			2025			LIFETIME				
01319			L02TN	WHBODY															2025/01	XA01132942I
				NOTE			**Dosimeter has exceeded the reporting capabilities of 1000 rads. Dosimeter reprocessed, second read agrees with reported dose.**													
				NOTE			**Neutron component has exceeded the fast and thermal reporting capabilities.**													
01320			L02TN	WHBODY	*PN		**6104**	**6104**	**6090**										2025/01	XA00008884F
					*P	H	**914**	**914**	**900**											
					*N	T	**330**	**330**	**330**											
					*N	F	**4860**	**4860**	**4860**											
01321			L02TN	WHBODY	*PN		**3597**	**3597**	**3585**										2025/01	XA00276998Q
					*P	H	**907**	**907**	**895**											
					*N	T	**200**	**200**	**200**											
					*N	F	**2490**	**2490**	**2490**											

* - Customer average background dose used for control subtraction

This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government.

Radiation Dosimetry Report

Annual Radiation Exposure Limits (mrem) :

Whole body, blood forming organs, gonads	5,000
Lens of Eye	15,000
Extremities and Skin	50,000
Fetal (Gestation period)	500
General Public	100

Based on the US NRC Regulations, Title 10, Part 20, Code of Federal Regulations and adopted by many states. Certain state and other regulatory agencies may adhere to different limits.
Control Dosimeter: A control dosimeter is included with each shipment of dosimeters for monitoring radiation exposure received during transit. At the customer's facility, store the control in a radiation free area during the wear period.
Minimal Dose Equivalent Reported: Dose equivalents below the minimum measurable quantity for the current monitoring period are recorded as "M." The minimal reporting levels vary by the dosimeter type and radiation quality. "SL" is an elective option for the minimal dose equivalent reported where exposures less than 10 mrem report as "SL" (excludes fetal dosimeters), and/or exposures at or more than 10 mrem begin reporting at 10 mrem and report in increments of 10 mrem.

Dosimeter Type	M (DDE,LDE,SDE)	M (SDE Only)	SL
Luxel+®	1	-	10
InLight®	5	-	10
Whole Body Beta	-	10	10
U Ring	-	30	-
Neutrak® Neutron Fast	20	-	-
Neutrak® Neutron Thermal/Fast	10	-	-
Saturn Ring	-	10	10

Special Calculations: Special dose calculations can be applied to radiation workers who wear lead aprons.
EDE 1 - two dosimeters: one worn at the waist level under lead apron and one worn at the collar level outside lead apron. 1.5 (Waist DDE) + 0.04 (Collar DDE) = Assigned Deep Dose Equivalent.
EDE 2 - one dosimeter: one worn at the collar level outside lead apron. 0.3 (Collar DDE) = Assigned Deep Dose Equivalent.
EDE 122 - one dosimeter: one worn at the collar level outside lead apron. Collar DDE / 5.6 = Assigned Deep Dose Equivalent.
Calc3 - Lens of Eye dosimeter. 0.5 (Lens of Eye LDE) = Assigned Lens of Eye Dose Equivalent.
Lens.175 - Lens of Eye dosimeter. 0.175 (Lens of Eye LDE) = Assigned Lens of Eye Dose Equivalent.
EDE1-NTC EDE1 without Thyroid Collar assigned deep dose equivalent = 0.06 × (collar dose - waist dose) + waist dose
EDE1-TC EDE1 with Thyroid Collar assigned deep dose equivalent = 0.02 × (collar dose - waist dose) + waist dose
The "ASSIGNED" line follows all of the original whole body dosimeter doses with the EDE 1 or EDE 2 calculation results or LANDAUER's standard Dose Assessment Protocol (deep and shallow whole body dose from the highest reading whole body dosimeter, lens dose from dosimeter closest to the eye).

Ring Dosimeter Reading: Ring dosimeter readings report as a shallow dose.

Fetal Dosimeter: A declared pregnant worker will possess a fetal exposure on an extra page of the report based upon the whole body dosimeter worn closest to the fetus. The fetal dose is reported for the current wear period, plus the estimated dose from conception to declaration (if provided by customer), and the total dose from declaration to present.

Use	Description	Use	Description
AREA	Area Monitor	OEXTRM	Other Extremity
CHEST	Chest	OWHBDY	Other Whole Body
CNTRL	Control	RANKLE	Right Ankle
COLLAR	Collar	RFINGR	Right Hand Ring
EYE	Eye	RUARM	Right Upper Arm
FETAL	Fetal	RULEG	Right Upper Leg
LANKLE	Left Ankle	RWRIST	Right Wrist
LFINGR	Left Hand Ring	SPCPUR	Special Purpose
LUARM	Left Upper Arm	UPBACK	Upper Back
LULEG	Left Upper Leg	WAIST	Waist
LWBACK	Lower Back	WHBODY	Whole Body
LWRIST	Left Wrist		

Code	Radiation Quality Description (Type and/or Energy)
B	beta
BH	beta high energy, e.g. Strontium, Phosphorus
BL	beta low energy e.g. Thallium, Krypton
BS	Strontium beta
BT	Thallium beta
BU	Uranium beta
BN	beta, neutron mixture
NF	neutron fast
NT	neutron thermal
P	photon (x or gamma ray)
PB	photon, beta mixture
PBN	photon, beta, neutron mixture
PH	photon high energy greater than 200 keV
PL	photon low energy less than 40 keV
PM	photon medium energy 40 keV to 200 keV
PN	photon, neutron mixture

First Line Explanation
Participant Number: Unique number assigned by LANDAUER.
Name: Participant to whom the dosimeter is assigned.
Dosimeter: Badge type according to radiation monitoring needs.

Dosimeter	Code	Type of Radiation Monitored				
		Photons			Neutrons	
		X	Gamma	Beta	Fast	Thermal
InLight Model 2	L02NN	Yes	Yes	Yes		
InLight Model 2J	L02JN	Yes	Yes	Yes	Yes	
InLight Model 2T	L02TN	Yes	Yes	Yes	Yes	Yes
Luxel+	Pa	Yes	Yes	Yes		
Luxel+	Ja	Yes	Yes	Yes	Yes	
Luxel+	Ta	Yes	Yes	Yes	Yes	Yes
Luxel+ Escort	Pa	Yes	Yes			
Neutrak	N				Yes	
Neutrak	E				Yes	Yes
Ring, Single TLD	U or S	Yes	Yes	Yes		

Deep, Eye and Shallow Dose Equivalents: Deep dose equivalent (DDE) applies to external whole body exposure at a tissue depth of 1 cm (1000 mg/cm^2).
Eye dose equivalent (LDE) applies to external exposure of the lens at a tissue depth of 0.3 cm (300 mg/cm^2).
Shallow dose equivalent (SDE) applies to the external exposure of the skin or extremity at a tissue depth of 0.007 cm (7 mg/cm^2) averaged over an area 1 cm^2.

Deep, eye and shallow dose equivalents report for the time frame indicated by "For Monitoring Period." These doses represent the dose received only for the account/subaccount specified. Individual radiation component results and combined totals report in separate lines.

Quarterly accumulated results reflect total dose received within a calendar 3-months time frame and the customer defined start day. (Note: Quarterly accumulated columns are eliminated for bimonthly service or display "Not applicable.") Year to date accumulation totals dose received from the beginning of the current year to report date. Lifetime accumulation totals all dose received from inception date of dosimeter service to report date, and could include earlier dose history if supplied by customer. Reported quarterly, annual and lifetime dose accumulations represent the doses totaling from all account/subaccount dosimeters to be reported at the customer level.

Inception Date: The date LANDAUER began keeping dosimeter records for a given dosimeter for a badging participant on the current customer.

Serial Number: Dosimeter serial number.

Second Line Explanation
Participant's personal information consisting of ID number and birth date. This information can be suppressed on "Duplicate and Original Reports" for privacy and/or posting needs.

Notes: Text messages explaining any abnormalities or comments. The notes with message appears on a separate line below all dosimeter exposure information.

U.S. Patents
6,316,702; 6,127,685; 5,892,234

LANDAUER®
© 2018 LANDAUER