



Tel: +1.805.964.3615 info@fwt.com
Fax: +1.805.964.3162 www.fwt.com
330 Kellogg Avenue, Suite D, Gokula, CA 93117 U.S.A.

CERTIFICATE OF CONFORMANCE

We do hereby certify that all of the items listed below were manufactured in full accordance with our established standards and procedures as well as in compliance with every detail of your

Purchase Order : CC on 03/18/2026

We further state that all records of test and inspection reflect acceptable data and that they, along with material traceability records, are on file at our facility and available for review at our facility.

ITEM MODEL OR PART#

DESCRIPTION

1

FWT-70-83M

Opti-Chromic Dosimeters

Lot

1190

Quantity

36 sets

50 dosimeters/set

Official Signature: *Scott A. Jarvis*

Title: Vice President

Date: 18-Mar-2026



FWT-70 Characterization

Batch 1190

Typical Calibration Curve

Dose, Gy	ΔA at 600 nm	ΔA at 656 nm
100.0	0.086	
1002	0.832	0.043
3010	2.087	0.119
10000		0.318
13010		0.383
16020		0.421

NOTE: This typical calibration curve is provided as a guide to the response of FWT-70 Opti-Chromic Detectors to ionizing radiation. Actual response also depends on the instrumentation used to measure optical densities.

This characterization is for this batch of FWT-70-83M formulation only and should be used in within the range covered.

Scott A. Larson
Authorization for Release

2025-11-05
Date