**DENTEX: 3,653 PNG files, 2,716 distinct radiographs**

labeled: quadrant 693, enumeration 634, enumeration-disease 705

excluded: official test set (Turkish taxonomy); 50 unlabeled demo images

*1,133 distinct labeled radiographs*

**Group by pixel SHA-256, then split 0.70 / 0.10 / 0.10 / 0.10**

identical films form one group; seed 42; assignments are append-only

| Train | Val. | Cal. | Test |
|---|---|---|---|

| Split | Films | Role |
|---|---|---|
| Train | 794 | fit network weights |
| Validation | 113 | epoch selection; validation metrics |
| Calibration | 113 | uncertainty thresholds; temperature |
| Test | 113 | final metrics; every use in the ledger |

**LEAKAGE CHECKS (2026-10-15)**

| Check | Result |
|---|---|
| file-hash overlap, all split pairs | 0 |
| training / epoch-selection films vs. test | 0 |
| near duplicates, pixel correlation ≥ 0.97 | 0 confirmed |
| two films of one patient | not checkable |

*No patient IDs exist in DENTEX. The test split was used once per model (2026-10-14).*