

EU-Declaration of Conformity (MDR)

Manufacturer: Xinhua Healthcare Industry Co., Ltd.



Address: **Head office:** A2 Building, Xinhua Health Industrial Park, No.2999 Zunxian Road, High-tech Zone, Zibo City, Shandong Province, 255086, P.R.China **Phone:** (+86) 0533-2320708

Product: Veterinary Electrosurgical unit **XGP-D-100**
Veterinary Ultrasonic Surgical System **XGP-C-100**

We declare under our sole responsibility that the above-mentioned products are in conformity with all applicable provisions of:

Directive 2014/35/EU (Low-Voltage Directive, LVD)

Directive 2014/30/EU (Electromagnetic Compatibility Directive, EMC)

Directive 2011/65/EU (RoHS 2.0), as amended.

Applicable harmonized standards(dated references)

EN 62368-1:2014+A11:2017

EN 55032:2015/A11:2020 Class A

EN 55035:2017/A11:2020



Additional information:

Technical documentation includes risk assessment for veterinary surgical application.

Reference test reports according to **EN 60601-1**, **EN 60601-2-2**, **EN IEC 61847** are kept within technical documentation for risk-evaluation purpose only, not used as harmonised standards for this declaration.

Intended use: For surgical cutting, coagulation and vessel- sealing procedures in veterinary/animal surgery.

Place: Zibo City, Shandong Province, China

Signed for and on behalf of: Shandong Xinhua Healthcare Industry Co., Ltd.

Name: Zhang Jian

Signature:

Title:

General Manager

Issue Date:

14.08.2025

Shandong Xinhua Healthcare Industry Co., Ltd.

A2 Building, Xinhua Health Industrial Park, No.2999 Zunxian Road,
High-tech Zone, Zibo City, Shandong Province, 255086, P.R.China