

EU-Declaration of Conformity (MDR)

Manufacturer: Xinhua Healthcare Industry Co., Ltd.



Address: **Head office:** A2 Building, Xinhua Health Industrial Park, No.2999 Zunxian Road, High-tech Zone, Zibo City, Shandong Province, 255086, P.R.China **Phone:** (+86) 0533-2320708

Product: Electrosurgical unit **XGP-D-I, XGP-D-II, XGP-Y-I**
Surgical energy platform **XGP-L-300/300A, XGP-L-400/400A**

We declare under our sole responsibility that the above-mentioned medical devices are in conformity with all applicable provisions of:

Regulation (EU) 2017/745 of the European Parliament and of the Council of 5 April 2017 on medical devices (MDR), as amended.

Applicable harmonized standards:

EN 60601-1: 2006 +A1 +A12 +A2: 2021

EN 60601-1-2: 2015 +A1: 2021

EN IEC 60601-2-2: 2018

EN ISO 13485: 2016 +A11: 2021

EN ISO 14971: 2019 +A11: 2021



Additional information:

The above-mentioned Electrosurgical unit XGP-D-I, XGP-D-II, XGP-Y-I; Surgical energy platform XGP-L-300, XGP-L-300A are classified as **Class IIb**, according to **Rule 9 of Annex VIII of Regulation (EU) 2017/745 (MDR)**.

Conformity assessment procedure: **Annex IX of Regulation (EU) 2017/745**.

Intended use: For surgical cutting, coagulation and vessel-sealing procedures in human surgery.

Place: Zibo City, Shandong Province, China

Signed for and on behalf of: Xinhua Healthcare Industry Co., Ltd.

Name: Zhang Jian

Signature:

Title:

General Manager

Issue Date:

08.08.2025

Xinhua Healthcare Industry Co., Ltd.

A2 Building, Xinhua Health Industrial Park, No.2999 Zunxian Road,
High-tech Zone, Zibo City, Shandong Province, 255086, P.R.China